

CATEGORY 6 ER AIR SAMPLING AND METEROLOGICAL
CATEGORY 6.1 SPECIFICATIONS
ER AIR SAMPLING AND METEROLOGICAL
Personal Sampling Pumps SKC PCXR 8 or Equivalent

Minimum Specifications:

1. 1cc to 5 liters per minute.
2. Programable sampling time.
3. Intermittent Sampling.
4. Flow rate + or - 5%
5. Must be able to sample
6. Must have digital readout, battery jumper
ability attach another battery externally