

## Questions / Comments on PHILIS RFP PR-HQ-08-13082

- 13.) Attachment 1, PHILIS Performance Work Statement: Paragraph 6.0, Administrative and Technical Personnel Categories (Page 1-14 of 22). This section states that, "The PHILIS requires staffing sufficient for continuous operation, 24 hrs/day, for an indefinite period."

Q13: Given that only the 20 Key Personnel Categories are being priced for labor, the total contract value will be determined based upon this pricing, along with Other Direct Costs. How will the cost for the surge capacity be addressed?

*Revised A13: See RFP Amendment #3.*

### *QUESTION FROM SITE VISIT*

- 23.) During the site visit it was mentioned that the methods being developed to detect WMD's would be much lower than those issued by the DoD. Is there a document that contains these target levels? If not, which agencies will be used as references?

*A23. The EPA will be using the SAM methods for the detection of the CWAs. The detection limits are still under development but would approximate those listed in EPA methods 8260 and 8270D. The draft SAM document web link is listed below:*

<http://www.epa.gov/nhsrc/pubs/600r04126d.pdf>.