NELAC PT
Fields of Testing
CWA-Water
Feb-00

Program	Matrix	Analyte	PTOB/PTPA	Conc Range	Acceptance Criteria			
					a	b	c	d
		Microbiology		CFU/100 ml				
Microbiology-CWA	Water	Total Coliform, MF	Per Sec 2.0	20 to 2400	Mean ±3 SD			
Microbiology-CWA	Water	Total Coliform, MPN	Per Sec 2.0	20 to 2400	Log transform; ±3 SD			
Microbiology-CWA	Water	Fecal Coliform, MF	Per Sec 2.0	20 to 2400	Mean ±3 SD			
Microbiology-CWA	Water	Fecal Coliform, MPN	Per Sec 2.0	20 to 2400	Log transform; ±3 SD			
		Trace Metals		µg/L				
Chemistry-CWA	Water	Aluminum	NIST NVLAP	130 to 4000	0.9950	2.5987	0.0427	6.0933
Chemistry-CWA	Water	Antimony	NIST NVLAP	95 to 900	0.9590	-3.6479	0.0779	3.2351
Chemistry-CWA	Water	Arsenic	NIST NVLAP	70 to 900	1.0062	-0.7508	0.0529	1.4080
Chemistry-CWA	Water	Barium	Sec. 2.0	100 to 2500	0.9881	6.8960	0.0473	-0.3500
Chemistry-CWA	Water	Beryllium	NIST NVLAP	8 to 900	0.9910	-0.6177	0.0460	0.2780
Chemistry-CWA	Water	Cadmium	NIST NVLAP	8 to 750	0.9940	0.2323	0.0463	0.3919
Chemistry-CWA	Water	Chromium, total	NIST NVLAP	17 to 1000	1.0015	-0.2586	0.0420	0.7988
Chemistry-CWA	Water	Chromium VI	Per Sec 2.0	15 to 1000	0.9820	0.0013	0.0560	0.0027
Chemistry-CWA	Water	Cobalt	NIST NVLAP	28 to 1000	1.0002	-0.2810	0.0395	0.4922
Chemistry-CWA	Water	Copper	NIST NVLAP	18 to 900	1.0031	-0.0890	0.0296	1.2415
Chemistry-CWA	Water	Iron	NIST NVLAP	30 to 4000	1.0056	1.1497	0.0390	2.0258
Chemistry-CWA	Water	Lead	NIST NVLAP	70 to 3000	0.9974	0.2778	0.0377	2.5294
Chemistry-CWA	Water	Manganese	NIST NVLAP	70 to 4000	1.0059	-1.1375	0.0351	0.3422
Chemistry-CWA	Water	Mercury	NIST NVLAP	0.5 to 30	0.9976	0.0054	0.0799	0.0444
Chemistry-CWA	Water	Molybdenum	NIST NVLAP	24 to 600	1.0030	-0.7711	0.0460	0.7257
Chemistry-CWA	Water	Nickel	NIST NVLAP	80 to 3000	1.0125	-1.6585	0.0333	2.0479
Chemistry-CWA	Water	Selenium	NIST NVLAP	90 to 2000	0.9774	-1.2658	0.0594	1.0204
Chemistry-CWA	Water	Silver	NIST NVLAP	26 to 600	1.0024	-0.2284	0.0475	0.1752
Chemistry-CWA	Water	Strontium	NIST NVLAP	3.5 to 300	0.9987	0.0099	0.0487	0.1633
Chemistry-CWA	Water	Thallium	NIST NVLAP	60 to 900	0.9828	0.2870	0.0573	1.0927
Chemistry-CWA	Water	Tin	Per Sec 2.0	1000 to 5000	1.0050	-6.8244	0.0730	-4.2660
Chemistry-CWA	Water	Titanium	NIST NVLAP	35 to 300	0.9927	0.0750	0.0420	0.5770
Chemistry-CWA	Water	Vanadium	NIST NVLAP	200 to 10000	1.0016	-2.0962	0.0325	0.3299
Chemistry-CWA	Water	Zinc	NIST NVLAP	30 to 2000	1.0045	0.3790	0.0378	1.6644
		Demands		mg/L				
Chemistry-CWA	Water	5-day BOD	NIST NVLAP	15 to 250	0.6312	0.1919	0.1032	0.1670
Chemistry-CWA	Water	Carbonaceous BOD	NIST NVLAP	15 to 250	0.5423	0.2956	0.0996	0.0697
Chemistry-CWA	Water	COD	NIST NVLAP	15 to 250	0.9517	0.4748	0.0471	2.4507
Chemistry-CWA	Water	TOC	NIST NVLAP	6 to 100	0.9904	0.1647	0.0508	0.1115
		Minerals		mg/L				
Chemistry-CWA	Water	Alkalinity, total ($CaCO_3$)	NIST NVLAP	10 to 120	0.9887	0.7718	0.0186	1.0581
Chemistry-CWA	Water	Calcium	NIST NVLAP	3.5 to 110	1.0135	0.0036	0.0377	0.1333
Chemistry-CWA	Water	Chloride	NIST NVLAP	35 to 250	1.0042	-0.4465	0.0262	0.4403
Chemistry-CWA	Water	Fluoride	NIST NVLAP	0.2 to 4	0.9934	0.0025	0.0362	0.0126
Chemistry-CWA	Water	Hardness, total ($CaCO_3$)	NIST NVLAP	45 to 350	1.0008	0.2222	0.0256	0.9247
Chemistry-CWA	Water	Magnesium	NIST NVLAP	0.9 to 40	0.9968	0.0113	0.0398	0.0310
Chemistry-CWA	Water	Potassium	NIST NVLAP	2.6 to 40	1.0017	0.0057	0.0426	0.1072
Chemistry-CWA	Water	Sodium	NIST NVLAP	7 to 100	0.9949	0.3338	0.0290	0.2329
Chemistry-CWA	Water	Spec. Cond. (25^oC)	NIST NVLAP	50 to 1050 µmhos/cm	study mean		0.0290	-1.2008
Chemistry-CWA	Water	Sulfate	NIST NVLAP	5 to 125	0.9854	0.0483	0.0471	0.4629
Chemistry-CWA	Water	Total Dissolved Solids at 180°	NIST NVLAP	30 to 650	study mean		0.0686	4.3676
Chemistry-CWA	Water	Total Residue	Sec. 2.0	140 to 675	0.9875	1.7890	0.0107	9.5940
		Nutrients		mg/L				
Chemistry-CWA	Water	Ammonia as N	NIST NVLAP	0.25 to 19	0.9940	0.0131	0.0700	0.0356
Chemistry-CWA	Water	Nitrate as N	NIST NVLAP	0.25 to 40	0.9896	0.0017	0.0649	0.0152
Chemistry-CWA	Water	Orthophosphate as P	NIST NVLAP	0.05 to 5.5	1.0047	-0.0009	0.0488	0.0059
Chemistry-CWA	Water	Total Kjeldahl-Nitrogen	NIST NVLAP	0.35 to 35	0.9796	0.0786	0.0773	0.1436
Chemistry-CWA	Water	Total Phosphorus	NIST NVLAP	0.15 to 10	0.9618	0.0215	0.0669	0.0088

NELAC PT
Fields of Testing
CWA-Water
Feb-00

Program	Matrix	Analyte	PTOB/PTPA	Conc Range	Acceptance Criteria			
					a	b	c	d
		Misc. Analytes		mg/L				
Chemistry-CWA	Water	Non-Filterable Residue	NIST NVLAP	23 to 100	0.9472	-1.6873	0.0477	0.2676
Chemistry-CWA	Water	Oil & Grease	NIST NVLAP	8 to 50	0.8926	0.3248	0.0503	1.2068
Chemistry-CWA	Water	pH	NIST NVLAP	5 to 10 units	0.9987	0.0092	0.0145	-0.0406
Chemistry-CWA	Water	Total Cyanide	NIST NVLAP	0.03 to 1	0.9903	-0.0004	0.0949	0.0027
Chemistry-CWA	Water	Total Phenolics (4AAP)	NIST NVLAP	0.04 to 5	0.6618	0.0001	0.0975	0.0030
Chemistry-CWA	Water	Total Residual Chlorine	NIST NVLAP	0.15 to 5	study mean		0.0520	0.0239
		Pesticides		µg/L				
Chemistry-CWA	Water	Aldrin	NIST NVLAP	0.1 to 3	0.8037	-0.0004	0.1821	0.0012
Chemistry-CWA	Water	alpha-BHC	Per Sec 2.0	2 to 15	0.8680	0.0608	0.1810	-0.0780
Chemistry-CWA	Water	beta-BHC	Per Sec 2.0	0.5 to 25	0.9030	0.0299	0.1700	0.0767
Chemistry-CWA	Water	delta-BHC	Per Sec 2.0	2 to 75	0.8495	-0.5260	0.2205	-0.5370
Chemistry-CWA	Water	gamma-BHC (Lindane)	Per Sec 2.0	0.5 to 25	0.8810	0.0532	0.1930	-0.0610
Chemistry-CWA	Water	Chlordane (total)	NIST NVLAP	1 to 13	0.9305	-0.0025	0.1649	-0.0035
Chemistry-CWA	Water	DDD (4,4)	NIST NVLAP	0.1 to 7	0.9440	-0.0013	0.1493	0.0053
Chemistry-CWA	Water	DDE (4,4)	NIST NVLAP	0.1 to 4	0.9215	-0.0069	0.1480	0.0015
Chemistry-CWA	Water	DDT (4,4)	NIST NVLAP	0.1 to 8	0.9039	0.0060	0.1474	0.0088
Chemistry-CWA	Water	Dieldrin	NIST NVLAP	0.1 to 5	0.9479	0.0008	0.1436	0.0014
Chemistry-CWA	Water	Endosulfan I	Per Sec 2.0	1 to 25	0.9460	-0.1522	0.1770	-0.0430
Chemistry-CWA	Water	Endosulfan II	Per Sec 2.0	2 to 75	0.9013	-0.0170	0.2512	-0.4510
Chemistry-CWA	Water	Endosulfan sulfate	Per Sec 2.0	0.5 to 25	0.8960	0.1841	0.2630	-0.3350
Chemistry-CWA	Water	Endrin	Per Sec 2.0	2 to 75	0.9402	-0.2420	0.2040	-0.2160
Chemistry-CWA	Water	Endrin aldehyde	Per Sec 2.0	0.5 to 15	0.8920	0.0411	0.1890	0.0052
Chemistry-CWA	Water	Heptachlor	NIST NVLAP	0.15 to 4	0.8278	0.0017	0.1891	-0.0023
Chemistry-CWA	Water	Heptachlor Epoxide (beta)	NIST NVLAP	0.1 to 3	0.8963	0.0013	0.1290	0.0012
Chemistry-CWA	Water	Methoxychlor	Per Sec 2.0	0.5 to 25	0.8960	0.3047	0.1500	0.2460
Chemistry-CWA	Water	Toxaphene	Per Sec 2.0	1 to 10	1.0120	-0.1622	0.3270	-0.2880
		Volatile Aromatics		µg/L				
Chemistry-CWA	Water	Benzene	NIST NVLAP	8 to 75	1.0026	0.0663	0.0908	0.1798
Chemistry-CWA	Water	1,2 Dichlorobenzene	NIST NVLAP	8 to 65	0.9812	0.1849	0.0886	0.2152
Chemistry-CWA	Water	1,3 Dichlorobenzene	NIST NVLAP	8 to 60	0.9741	0.1037	0.0863	0.1708
Chemistry-CWA	Water	1,4 Dichlorobenzene	NIST NVLAP	7 to 65	0.9838	0.2265	0.0971	0.1507
Chemistry-CWA	Water	Ethylbenzene	NIST NVLAP	8 to 75	0.9867	0.0216	0.1091	-0.1180
Chemistry-CWA	Water	Toluene	NIST NVLAP	7 to 65	0.9804	0.1515	0.0843	0.1981
Chemistry-CWA	Water	Xylenes, total	Sec. 2.0	10 to 500	0.9498	1.1598	0.1232	0.7309
		Volatile Halocarbons		µg/L				
Chemistry-CWA	Water	Bromodichloromethane	NIST NVLAP	7 to 75	1.0056	-0.0205	0.0967	0.1039
Chemistry-CWA	Water	Bromoform	NIST NVLAP	11 to 75	1.0191	-0.4676	0.1147	0.2535
Chemistry-CWA	Water	Bromomethane	Sec. 2.0	10 to 200	1.0200	-0.5330	0.2502	-0.0970
Chemistry-CWA	Water	Carbontetrachloride	NIST NVLAP	9 to 75	1.0097	0.2449	0.1329	-0.0432
Chemistry-CWA	Water	Chlorobenzene	NIST NVLAP	11 to 75	0.9849	0.2402	0.0982	-0.1784
Chemistry-CWA	Water	Chloroethane	Sec. 2.0	10 to 200	0.7522	9.8210	0.1623	1.0520
Chemistry-CWA	Water	Chloroform	NIST NVLAP	10 to 75	0.9802	0.4148	0.0982	-0.0085
Chemistry-CWA	Water	Chloromethane	Sec. 2.0	10 to 200	0.9315	3.3290	0.2069	1.3060
Chemistry-CWA	Water	Dibromochloromethane	NIST NVLAP	9 to 75	0.9968	-0.1129	0.1113	-0.0043
Chemistry-CWA	Water	1,2 Dichloroethane	NIST NVLAP	10 to 75	1.0002	0.5510	0.1049	0.0917
Chemistry-CWA	Water	1,1-Dichloroethene	Sec. 2.0	10 to 200	1.0300	0.5257	0.1500	0.8827
Chemistry-CWA	Water	trans-1,2-Dichloroethene	Sec. 2.0	10 to 200	1.0020	-0.1039	0.1150	0.7301
Chemistry-CWA	Water	1,2-Dichloropropane	Sec. 2.0	10 to 200	0.9690	0.3978	0.1030	-0.3240
Chemistry-CWA	Water	trans-1,2,-Dichloropropene	Sec. 2.0	10 to 200	0.9197	0.0000	0.2091	-0.9360
Chemistry-CWA	Water	Methylene Chloride	NIST NVLAP	8 to 75	0.9945	0.5176	0.1213	0.1537
Chemistry-CWA	Water	1,1,2,2-Tetrachloroethane	Sec. 2.0	10 to 200	0.9940	-0.2281	0.1330	0.1771
Chemistry-CWA	Water	Tetrachloroethene	NIST NVLAP	9 to 75	0.9663	0.2554	0.1010	0.1153
Chemistry-CWA	Water	1,1,1-Trichloroethane	NIST NVLAP	9 to 75	0.9797	0.1711	0.1140	-0.0877
Chemistry-CWA	Water	1,1,2-Trichloroethane	Sec. 2.0	10 to 200	0.9870	0.2575	0.1060	-0.2230
Chemistry-CWA	Water	Trichloroethene	NIST NVLAP	7 to 75	0.9661	0.2374	0.1070	0.0134
Chemistry-CWA	Water	Trichlorofluoromethane	Sec. 2.0	10 to 200	0.9522	1.6640	0.2807	-1.7000
Chemistry-CWA	Water	Vinyl chloride	Sec. 2.0	10 to 200	1.1660	-3.8700	0.2222	0.3442

NELAC PT
Fields of Testing
CWA-Water
Feb-00

Program	Matrix	Analyte	PTOB/PTPA	Conc Range	Acceptance Criteria			
					a	b	c	d
		B/N		µg/L				
Chemistry-CWA	Water	Acenaphthene	Per Sec 2.0	10 to 200	0.8170	0.8563	0.1530	-0.2540
Chemistry-CWA	Water	Acenaphthylene	Per Sec 2.0	10 to 200	0.7990	0.6883	0.1300	0.6054
Chemistry-CWA	Water	Anthracene	Per Sec 2.0	10 to 200	0.7900	3.4560	0.1664	-1.0700
Chemistry-CWA	Water	Benzidine	Per Sec 2.0	10 to 200	1.1670	-12.2680	0.5790	-0.3010
Chemistry-CWA	Water	Benzo(a)anthracene	Per Sec 2.0	10 to 200	0.8730	-0.1342	0.1340	-1.2970
Chemistry-CWA	Water	Benzo butyl phthalate	Per Sec 2.0	10 to 200	0.6323	9.5980	0.3117	-2.5700
Chemistry-CWA	Water	Benzo(b)fluoranthene	Per Sec 2.0	10 to 200	0.8099	2.0050	0.2210	-2.0200
Chemistry-CWA	Water	Benzo(k)fluoranthene	Per Sec 2.0	10 to 200	0.8223	1.9960	0.1862	1.1260
Chemistry-CWA	Water	Benzo(g,h,i)perylene	Per Sec 2.0	10 to 200	0.9120	-3.2300	0.1914	1.2510
Chemistry-CWA	Water	4-Bromophenyl-phenylether	Per Sec 2.0	10 to 200	0.8400	1.5522	0.1460	-0.4440
Chemistry-CWA	Water	bis(2-Chloroethoxy)methane	Per Sec 2.0	10 to 200	0.7828	0.8980	0.1280	0.4366
Chemistry-CWA	Water	bis(2-Chloroethyl)ether	Per Sec 2.0	10 to 200	0.7120	3.7209	0.1540	0.4800
Chemistry-CWA	Water	bis(2-Chloroiospropyl) ether	Per Sec 2.0	10 to 200	0.9293	-6.4969	0.2293	-2.4073
Chemistry-CWA	Water	Bis(2-ethylhexyl) phthlate	Per Sec 2.0	10 to 200	0.7600	6.3794	0.2390	-2.6700
Chemistry-CWA	Water	4-Chlorophenyl-phenylether	Per Sec 2.0	10 to 200	0.7890	3.5242	0.1630	-1.3880
Chemistry-CWA	Water	2-Chloronaphthalene	Per Sec 2.0	10 to 200	0.7460	2.5724	0.1120	1.1912
Chemistry-CWA	Water	Chrysene	Per Sec 2.0	10 to 200	0.8340	1.3192	0.1670	-1.0290
Chemistry-CWA	Water	Dibenzo(a,h)anthracene	Per Sec 2.0	10 to 200	0.8726	-1.2390	0.2112	-0.4430
Chemistry-CWA	Water	3.3'-dichlorobenzidine	Per Sec 2.0	10 to 200	0.9010	-0.5596	0.1990	2.5071
Chemistry-CWA	Water	Diethyl phthalate	Per Sec 2.0	10 to 200	0.6641	8.8170	0.2919	-3.3500
Chemistry-CWA	Water	Dimethyl phthalate	Per Sec 2.0	10 to 200	0.5420	13.2000	0.3312	-4.5700
Chemistry-CWA	Water	Di-n-butylphthalate	Per Sec 2.0	10 to 200	0.7000	7.9114	0.2260	-1.7350
Chemistry-CWA	Water	2,4-Dinitrotoluene	Per Sec 2.0	10 to 200	0.9031	-3.4569	0.1466	0.7149
Chemistry-CWA	Water	2,6-Dinitrotoluene	Per Sec 2.0	10 to 200	0.8671	-2.0045	0.1616	-0.6668
Chemistry-CWA	Water	Di-n-octylphthalate	Per Sec 2.0	10 to 200	0.7590	4.6078	0.2220	-0.4400
Chemistry-CWA	Water	Fluoranthene	Per Sec 2.0	10 to 200	0.8480	1.3758	0.1500	-0.6930
Chemistry-CWA	Water	Fluorene	Per Sec 2.0	10 to 200	0.8420	0.8642	0.1420	0.1801
Chemistry-CWA	Water	Hexachlorobenzene	Per Sec 2.0	10 to 200	0.8620	0.1945	0.1410	0.1137
Chemistry-CWA	Water	Hexachlorobutadiene	Per Sec 2.0	10 to 200	0.6840	0.1935	0.1640	0.8656
Chemistry-CWA	Water	Hexachlorocyclopentadiene	Per Sec 2.0	10 to 200	0.5780	0.4585	0.2200	1.4156
Chemistry-CWA	Water	Hexachloroethane	Per Sec 2.0	10 to 200	0.6730	0.5246	0.2060	-1.0630
Chemistry-CWA	Water	Indeno(1,2,3, cd)pyrene	Per Sec 2.0	10 to 200	0.9670	-7.8000	0.3449	-5.7640
Chemistry-CWA	Water	Isophorone	Per Sec 2.0	10 to 200	0.8280	1.0402	0.1730	-0.9640
Chemistry-CWA	Water	Naphthalene	Per Sec 2.0	10 to 200	0.7520	2.7662	0.1810	-0.6120
Chemistry-CWA	Water	Nitrobenzene	Per Sec 2.0	10 to 200	0.8080	0.8319	0.1680	-0.5000
Chemistry-CWA	Water	N-Nitrosodimethylamine	Per Sec 2.0	10 to 200	0.5320	0.7787	0.2020	1.4455
Chemistry-CWA	Water	N-Nitroso-di-n-propylamine	Per Sec 2.0	10 to 200	0.8070	-0.0632	0.1610	0.5176
Chemistry-CWA	Water	N-Nitrosodiphenylamine	Per Sec 2.0	10 to 200	0.7760	1.9604	0.1780	0.9231
Chemistry-CWA	Water	Phenanthrene	Per Sec 2.0	10 to 200	0.8024	3.7280	0.1828	-2.7400
Chemistry-CWA	Water	Pyrene	Per Sec 2.0	10 to 200	0.8196	2.6820	0.1610	1.0620
Chemistry-CWA	Water	1,2,4-Trichlorobenzene	Per Sec 2.0	10 to 200	0.7230	1.9368	0.1420	0.8157
		Acids		µg/L				
Chemistry-CWA	Water	4-Chloro-3-methylphenol	Per Sec 2.0	10 to 200	0.8450	-0.8910	0.1460	0.3823
Chemistry-CWA	Water	2-Chlorophenol	Per Sec 2.0	10 to 200	0.7540	2.2054	0.1630	-0.1850
Chemistry-CWA	Water	2,4-Dichlorophenol	Per Sec 2.0	10 to 200	0.7670	1.8422	0.1360	0.5551
Chemistry-CWA	Water	2,4-Dimethylphenol	Per Sec 2.0	10 to 200	0.7700	-0.7906	0.1740	1.0376
Chemistry-CWA	Water	2,4-Dinitrophenol	Per Sec 2.0	10 to 200	0.5336	12.4300	0.2011	3.2470
Chemistry-CWA	Water	2-Methyl-4,6-Dinitrophenol	Per Sec 2.0	10 to 200	0.7940	-1.7131	0.2630	-5.9330
Chemistry-CWA	Water	2-Nitrophenol	Per Sec 2.0	10 to 200	0.8070	-0.0321	0.1730	-0.4150
Chemistry-CWA	Water	4-Nitrophenol	Per Sec 2.0	10 to 200	0.5403	1.5470	0.2449	2.9140
Chemistry-CWA	Water	Phenol	Per Sec 2.0	10 to 200	0.5570	0.5929	0.2530	1.0269
Chemistry-CWA	Water	Pentachlorophenol	Per Sec 2.0	10 to 200	0.8490	-3.1159	0.1780	1.0189
Chemistry-CWA	Water	2,4,6,-Trichlorophenol	Per Sec 2.0	10 to 200	0.8230	0.3273	0.1710	-0.5830

NELAC PT
Fields of Testing
CWA-Water
Feb-00

Program	Matrix	Analyte	PTOB/PTPA	Conc Range	Acceptance Criteria			
					a	b	c	d
		PCBs in Water		µg/L				
Chemistry-CWA	Water	Aroclor 1016/1242	NIST NVLAP	3.8 to 13	0.8344	0.0810	0.2101	-0. 1922
Chemistry-CWA	Water	Arochlor 1221	Per Sec. 2.0	1 to 15	0.7867	0.2517	0.2005	0.1023
Chemistry-CWA	Water	Aroclor 1232	NIST NVLAP	1.4 to 4	0.9463	-0.0779	0.3325	-0.2539
Chemistry-CWA	Water	Aroclor 1248	NIST NVLAP	1.5 to 5.5	0.9327	-0.0919	0.1699	-0.0187
Chemistry-CWA	Water	Aroclor 1254	NIST NVLAP	1.7 to 5.5	0.8622	0.1140	0.1129	0.1214
Chemistry-CWA	Water	Aroclor 1260	NIST NVLAP	1.6 to 5	0.9507	-0.1281	0.1087	0.0850
		PCBs in Oil		mg/kg				
Chemistry-CWA	Oil	Aroclor 1016/1242	NIST NVLAP	17 to 50	0.7208	1.6866	0.1569	1.4646
Chemistry-CWA	Oil	Aroclor 1254	NIST NVLAP	16 to 50	0.7936	0.5516	0.1759	1.6115
Chemistry-CWA	Oil	Aroclor 1260	NIST NVLAP	12 to 50	0.7803	0.5911	0.2019	0.1025
		Herbicides		µg/L				
Chemistry-CWA	Water	2,4-D	Per Sec 2.0	1 to 10	0.6980	0.3133	0.2290	0.2377
Chemistry-CWA	Water	Dicamba	Per Sec 2.0	1 to 10	0.7590	0.0590	0.2140	0.0954
Chemistry-CWA	Water	2,4,5-T	Per Sec 2.0	1 to 10	0.7830	-0.0043	0.2050	0.1616
Chemistry-CWA	Water	Silvex	Per Sec 2.0	1 to 10	0.6608	0.2491	0.1956	0.1332