

METHOD 7000B

FLAME ATOMIC ABSORPTION SPECTROPHOTOMETRY

SW-846 is not intended to be an analytical training manual. Therefore, method procedures are written based on the assumption that they will be performed by analysts who are formally trained in at least the basic principles of chemical analysis and in the use of the subject technology.

In addition, SW-846 methods, with the exception of required method use for the analysis of method-defined parameters, are intended to be guidance methods which contain general information on how to perform an analytical procedure or technique which a laboratory can use as a basic starting point for generating its own detailed Standard Operating Procedure (SOP), either for its own general use or for a specific project application. The performance data included in this method are for guidance purposes only, and are not intended to be and must not be used as absolute QC acceptance criteria for purposes of laboratory accreditation.

1.0 SCOPE AND APPLICATION

1.1 Metals in solution may be readily determined by flame (direct aspiration) atomic absorption spectrophotometry. The method is simple, rapid, and applicable to a large number of environmental samples including, but not limited to, ground water, aqueous samples, extracts, industrial wastes, soils, sludges, sediments, and similar wastes. With the exception of the analyses for dissolved constituents, all samples require digestion prior to analysis (see Chapter Three). Analysis for dissolved elements does not require digestion if the sample has been filtered and then acidified.

NOTE: Organo-metallic species may not be detected if the sample is not digested.

The following elements have been determined by this method:

<u>ELEMENT</u>	<u>CASRN^a</u>
Aluminum (Al)	7429-90-5
Antimony (Sb)	7440-36-0
Barium (Ba)	7440-39-3
Beryllium (Be)	7440-41-7
Cadmium (Cd)	7440-43-9
Calcium (Ca)	7440-70-2
Chromium (Cr)	7440-47-3
Cobalt (Co)	7440-48-4
Copper (Cu)	7440-50-8
Iron (Fe)	7439-89-6
Lead (Pb)	7439-92-1
Lithium (Li)	7439-93-2
Magnesium (Mg)	7439-95-4
Manganese (Mn)	7439-96-5
Molybdenum (Mo)	7439-98-7
Nickel (Ni)	7440-02-0
Osmium (Os)	7440-04-2
Potassium (K)	7440-09-7
Silver (Ag)	7440-22-4

Sodium	(Na)	7440-23-5
Strontium	(Sr)	7440-24-6
Thallium	(Tl)	7440-28-0
Tin	(Sn)	7440-31-5
Vanadium	(V)	7440-62-2
<u>Zinc</u>	<u>(Zn)</u>	<u>7440-66-6</u>

^a Chemical Abstract Service Registry Number

1.2 Lower limits of quantitation and optimum ranges of the metals will vary with the matrices and models of atomic absorption spectrophotometers. The data shown in Table 1 provide some indication of the lower limits of quantitation obtainable by the direct aspiration technique. For clean aqueous samples, the quantitation limits shown in the table by direct aspiration may be extended downward with scale expansion and upward by using a less sensitive wavelength or by rotating the burner head. Quantitation limits by direct aspiration may also be extended through concentration of the sample and/or through solvent extraction techniques.

1.3 Users of this method should state the data quality objectives prior to analysis and must document and have on file the required initial demonstration performance data described in the following sections prior to using this method for analysis.

1.4 Where direct-aspiration atomic absorption techniques do not provide adequate sensitivity, refer to specialized procedures such as graphite furnace atomic absorption (Method 7010) or the gaseous-hydride methods.

1.5 Other elements and matrices may be analyzed by this method as long as the method performance is demonstrated for these additional elements of interest, in the additional matrices of interest, at the concentration levels of interest in the same manner as the listed elements and matrices (see Sec. 9.0).

1.6 Prior to employing this method, analysts are advised to consult each type of procedure (e.g., sample preparation methods) that may be employed in the overall analysis for additional information on quality control procedures, development of QA acceptance criteria, calculations, and general guidance. Analysts also should consult the disclaimer statement at the front of the manual and the information in Chapter Two for guidance on the intended flexibility in the choice of methods, apparatus, materials, reagents, and supplies, and on the responsibilities of the analyst for demonstrating that the techniques employed are appropriate for the analytes of interest, in the matrix of interest, and at the levels of concern.

In addition, analysts and data users are advised that, except where explicitly specified in a regulation, the use of SW-846 methods is *not* mandatory in response to Federal testing requirements. The information contained in this method is provided by EPA as guidance to be used by the analyst and the regulated community in making judgments necessary to generate results that meet the data quality objectives for the intended application.

1.7 Use of this method is restricted to use by, or under supervision of, properly experienced and trained personnel, including analysts who are knowledgeable in the chemical and physical interferences described in this method. Each analyst must demonstrate the ability to generate acceptable results with this method.

2.0 SUMMARY OF METHOD

2.1 Although methods have been reported for the analysis of solids by atomic absorption spectrophotometry, the technique generally is limited to metals in solution or dissolved through some form of sample processing (see Chapter Three). Preliminary treatment of waste water, ground water, extracts, and industrial waste is always necessary because of the complexity and variability of sample matrix. Solids, slurries, and suspended material must be subjected to a solubilization process before analysis. This process may vary because of the metals to be determined and the nature of the sample being analyzed. Solubilization and digestion procedures are presented in Chapter Three.

2.2 In direct-aspiration atomic absorption spectrophotometry, a sample is aspirated and atomized in a flame. A light beam from a hollow cathode lamp or an electrodeless discharge lamp is directed through the flame into a monochromator, and onto a detector that measures the amount of absorbed light. Absorption depends upon the presence of free unexcited ground-state atoms in the flame. Because the wavelength of the light beam is characteristic of only the metal being determined, the light energy absorbed by the flame is a measure of the concentration of that metal in the sample. This principle is the basis of atomic absorption spectrophotometry.

3.0 DEFINITIONS

Refer to Chapter One, Chapter Three, and the manufacturer's instructions for a definitions that may be relevant to this procedure.

4.0 INTERFERENCES

4.1 Solvents, reagents, glassware, and other sample processing hardware may yield artifacts and/or interferences to sample analysis. All of these materials must be demonstrated to be free from interferences under the conditions of the analysis by analyzing method blanks. Specific selection of reagents and purification of solvents by distillation in all-glass systems may be necessary. Refer to each method to be used for specific guidance on quality control procedures and to Chapter Three for general guidance on the cleaning of glassware.

4.2 The most troublesome type of interference in atomic absorption spectrophotometry is usually termed "chemical" and is caused by lack of absorption of atoms bound in molecular combination in the flame. This phenomenon can occur when the flame is not sufficiently hot to dissociate the molecule, as in the case of phosphate interference with magnesium, or when the dissociated atom is immediately oxidized to a compound that will not dissociate further at the temperature of the flame. The addition of lanthanum will overcome phosphate interference in magnesium, calcium, and barium determinations. Similarly, silica interference in the determination of manganese can be eliminated by the addition of calcium. A nitrous oxide/acetylene gas mixture may be used to help prevent interferences from refractory compounds.

4.3 Chemical interferences may also be eliminated by separating the metal from the interfering material. Although complexing agents are employed primarily to increase the sensitivity of the analysis, they may also be used to eliminate or reduce interferences.

4.4 The presence of high dissolved solids in the sample may result in an interference from non-atomic absorbance such as light scattering. In the absence of background correction, this can result in false positives and/or falsely elevated values. If background correction is not

available, a non-absorbing wavelength should be checked. Signal contribution from uncorrected background can not be diagnosed through the analysis of spike recovery, nor is it compensated for by the application of the method of standard additions (MSA). If background correction is not available and the non-absorbing wavelength test indicates the presence of background interference, the sample digestates must be extracted (liquid-liquid or solid phase) prior to analysis, or another analytical method must be selected.

4.5 Ionization interferences occur when the flame temperature is sufficiently high to generate the removal of an electron from a neutral atom, giving a positively charged ion. This type of interference can generally be controlled by the addition, to both standard and sample solutions, of a large excess (1,000 mg/L) of an easily ionized element such as K, Na, Li or Cs. Each sample and standard should contain 2 mL KCl/100 mL of solution. Use 95 g of potassium chloride in 1 L of reagent water for the KCl solution.

4.6 Spectral interference can occur when an absorbing wavelength of an element present in the sample, but not being determined, falls within the width of the absorption line of the element of interest. The results of the determination will then be erroneously high, due to the contribution of the interfering element to the atomic absorption signal. Interference can also occur when resonant energy from another element in a multielement lamp, or from a metal impurity in the lamp cathode, falls within the bandpass of the slit setting when that other metal is present in the sample. This type of interference may sometimes be reduced by narrowing the slit width.

4.7 The analyst should be aware that viscosity differences and/or high dissolved or suspended solids may alter the aspiration rate.

4.8 All metals are not equally stable in the digestate, especially if it only contains nitric acid and not a combination of acids including hydrochloric acid. The addition of HCl helps stabilize Sn, Sb, Mo, Ba, and Ag in the digestate. The digestate should be analyzed as soon as possible, with preference given to these analytes. Refer to Chapter Three for suggested decomposition methods.

4.9 Specific interference problems related to individual analytes

4.9.1 Aluminum -- Aluminum may be as much as 15% ionized in a nitrous-oxide/acetylene flame. Use of an ionization suppressor (1,000 µg/mL of K as KCl) as described in Sec. 4.5 will eliminate this interference.

4.9.2 Antimony -- In the presence of lead (1,000 mg/L), a spectral interference may occur at the 217.6-nm resonance line. In this case, the 231.1-nm resonance line should be used. Excess concentrations of copper and nickel (and potentially other elements), as well as acids, can interfere with antimony analyses. If the sample contains these matrix types, either matrices of the standards should be matched to those of the sample or the sample should be analyzed using a nitrous oxide/acetylene flame.

4.9.3 Barium -- Barium undergoes significant ionization in the nitrous oxide/acetylene flame, resulting in a significant decrease in sensitivity. All samples and standards must contain 2 mL of the KCl ionization suppressant per 100 mL of solution (refer to Sec. 4.5). In addition, high hollow cathode current settings and a narrow spectral band pass must be used because both barium and calcium emit strongly at barium's analytical wavelength.

4.9.4 Beryllium -- Concentrations of Al greater than 500 ppm may suppress beryllium absorbance. The addition of 0.1% fluoride has been found effective in

eliminating this interference. High concentrations of magnesium and silicon cause similar problems and require the use of the method of standard additions.

4.9.5 Calcium -- All elements forming stable oxyanions will complex calcium and interfere unless lanthanum is added. Addition of lanthanum to prepared samples rarely presents a problem because virtually all environmental samples contain sufficient calcium to require dilution to be within the linear range of the method.

4.9.6 Chromium -- An ionization interference may occur if the samples have a significantly higher alkali metal content than the standards. If this interference is encountered, an ionization suppressant (KCl) should be added to both samples and standards (refer to Sec. 4.5).

4.9.7 Magnesium -- All elements forming stable oxyanions (P, B, Si, Cr, S, V, Ti, Al, etc.) will complex magnesium and interfere unless lanthanum is added. Addition of lanthanum to prepared samples rarely presents a problem because virtually all environmental samples contain sufficient magnesium to require dilution.

4.9.8 Molybdenum -- Interferences in an air/acetylene flame from Ca, Sr, SO₄, and Fe are severe. These interferences are greatly reduced in the nitrous oxide flame and by the addition of 1,000 mg/L of aluminum to samples and standards (refer to Sec. 7.7).

4.9.9 Nickel -- High concentrations of iron, cobalt, or chromium may interfere, requiring either matrix matching or use of a nitrous-oxide/acetylene flame. A non-response line of Ni at 232.14 nm causes non-linear calibration curves at moderate to high nickel concentrations, requiring sample dilution or use of the 352.4 nm line.

4.9.10 Osmium -- Due to the volatility of osmium, standards must be made on a daily basis, and the applicability of sample preparation techniques must be verified for the sample matrices of interest.

4.9.11 Potassium -- In air/acetylene or other high temperature flames (>2800 EC), potassium can experience partial ionization, which indirectly affects absorption sensitivity. The presence of other alkali salts in the sample can reduce ionization and thereby enhance analytical results. The ionization-suppressive effect of sodium is small if the ratio of Na to K is under 10. Any enhancement due to sodium can be stabilized by adding excess sodium (1,000 µg/mL) to both sample and standard solutions. If more stringent control of ionization is needed, the addition of cesium should be considered.

4.9.12 Silver -- Since silver nitrate solutions are light sensitive and have the tendency to plate silver out on the container walls, they should be stored in dark-colored bottles. In addition, it is recommended that the stock standard concentrations be kept below 2 ppm and the chloride content increased to prevent precipitation. If precipitation is occurring, a 5%:2% HCl:HNO₃ stock solution may prevent precipitation. Daily standard preparation may also be needed to prevent precipitation of silver.

4.9.13 Strontium -- Chemical interference caused by silicon, aluminum, and phosphate are controlled by adding lanthanum chloride. Potassium chloride is added to suppress the ionization of strontium. All samples and standards should contain 1 mL of lanthanum chloride/potassium chloride solution per 10 mL of solution (refer to Sec. 7.8).

4.9.14 Vanadium -- High concentrations of aluminum or titanium, or the presence of Bi, Cr, Fe, acetic acid, phosphoric acid, surfactants, detergents, or alkali

metals, may interfere. The interference can be controlled by adding 1,000 mg/L of aluminum to samples and standards (refer to Sec. 7.7).

4.9.15 Zinc -- High levels of silicon, copper, or phosphate may interfere. Addition of strontium (1,500 mg/L) removes the copper and phosphate interference.

5.0 SAFETY

5.1 This method does not address all safety issues associated with its use. The laboratory is responsible for maintaining a safe work environment and a current awareness file of OSHA regulations regarding the safe handling of the chemicals listed in this method. A reference file of material safety data sheets (MSDSs) should be available to all personnel involved in these analyses.

5.2 Concentrated nitric and hydrochloric acids are moderately toxic and extremely irritating to skin and mucus membranes. Use these reagents in a hood whenever possible and, if eye or skin contact occurs, flush with large volumes of water. Always wear safety glasses or a shield for eye protection when working with these reagents.

5.3 Hydrofluoric acid is a very toxic acid and penetrates the skin and tissues deeply if not treated immediately. Injury occurs in two stages; first, by hydration that induces tissue necrosis and then by penetration of fluoride ions deep into the tissue and by reaction with calcium. Boric acid and other complexing reagents and appropriate treatment agents should be administered immediately. Consult appropriate safety literature and have the appropriate treatment materials readily available prior to working with this acid. See Method 3052 for specific suggestions for handling hydrofluoric acid from a safety and an instrument standpoint.

5.4 Many metal salts are extremely toxic if inhaled or swallowed. Extreme care must be taken to ensure that samples and standards are handled properly and that all exhaust gases are properly vented. Wash hands thoroughly after handling.

5.5 Protective eyewear and/or flame shields should be used when conducting analyses by acetylene-nitrous oxide flame due to the emission of UV light.

5.6 The acidification of samples containing reactive materials may result in the release of toxic gases, such as cyanides or sulfides. For this reason, the acidification and digestion of samples should be performed in an approved fume hood.

6.0 EQUIPMENT AND SUPPLIES

The mention of trade names or commercial products in this manual is for illustrative purposes only, and does not constitute an EPA endorsement or exclusive recommendation for use. The products and instrument settings cited in SW-846 methods represent those products and settings used during method development or subsequently evaluated by the Agency. Glassware, reagents, supplies, equipment, and settings other than those listed in this manual may be employed provided that method performance appropriate for the intended application has been demonstrated and documented.

This section does not list common laboratory glassware (e.g., beakers and flasks).

6.1 Atomic absorption spectrophotometer -- Single- or dual-channel, single- or double-beam instrument having a grating monochromator, photomultiplier detector, adjustable slits, a wavelength range of 190 to 800 nm, and provisions for a computer or graphical interface.

6.2 Burner -- The burner recommended by the particular instrument manufacturer should be used. For certain elements the nitrous oxide burner is needed. Under no circumstance should an acetylene-air burner head be used with an acetylene-nitrous oxide flame.

6.3 Hollow cathode lamps -- Single-element lamps are preferred, but multielement lamps may be used. Electrodeless discharge lamps may also be used when available. Other types of lamps meeting the performance criteria of this method may be used.

6.4 Graphical display and recorder -- A recorder is recommended for flame work so that there will be a permanent record and that any problems with the analysis such as drift, incomplete atomization, losses during charring, changes in sensitivity, peak signal, etc., can be easily recognized.

6.5 Pipets -- Class A or microliter, with disposable tips. Sizes can range from 5 to 100 μ L as needed. Pipet tips should be checked as a possible source of contamination when contamination is suspected or when a new source or batch of pipet tips is received by the laboratory. The accuracy of variable pipets must be verified daily. Class A pipets can be used for the measurement of volumes equal to or larger than 1 mL.

6.6 Pressure-reducing valves -- The supplies of fuel and oxidant should be maintained at pressures somewhat higher than the controlled operating pressure of the instrument by suitable valves.

6.7 Glassware -- All glassware, polypropylene, or fluorocarbon (PFA or TFM) containers, including sample bottles, flasks and pipets, should be washed in the following sequence -- 1:1 hydrochloric acid, tap water, 1:1 nitric acid, tap water, detergent, tap water, and reagent water. (Chromic acid should not be used as a cleaning agent for glassware if chromium is to be included in the analytical scheme.) If it can be documented through an active analytical quality control program using spiked samples and method blanks that certain steps in the cleaning procedure are not needed for routine samples, those steps may be eliminated from the procedure. Alternative cleaning procedures must also be documented.

6.8 Volumetric flasks of suitable precision and accuracy.

7.0 REAGENTS AND STANDARDS

7.1 Reagent grade- or trace metals-grade chemicals must be used in all tests. Unless otherwise indicated, it is intended that all reagents conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society, where such specifications are available. Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination. All reagents should be analyzed to demonstrate that the reagents do not contain target analytes at or above the lowest limit of quantitation.

7.2 Reagent water -- All references to water in the method refer to reagent water, unless otherwise specified. Reagent water must be free of interferences.

7.3 Nitric acid, HNO_3 -- Use a spectrograde acid certified for AA use. Prepare a 1:1 dilution with water by adding the concentrated acid to an equal volume of water. If the method blank does not contain target analytes at or above the lowest limit of quantitation, then the acid may be used.

7.4 Hydrochloric acid (1:1), HCl -- Use a spectrograde acid certified for AA use. Prepare a 1:1 dilution with water by adding the concentrated acid to an equal volume of water. If the method blank does not contain target analytes at or above the lowest limit of quantitation, then the acid may be used.

7.5 Fuel and oxidant -- High purity acetylene is generally acceptable. Air may be supplied from a compressed air line, a laboratory compressor, or a cylinder of compressed air and should be clean and dry. Nitrous oxide is also required for certain determinations. A centrifuge filter on the compressed air lines is also recommended to remove particulates.

7.6 Stock standard metal solutions -- Stock standard solutions are prepared from analytical reagent grade high purity metals, oxides, or nonhygroscopic salts using reagent water and redistilled nitric or hydrochloric acids. Sulfuric or phosphoric acids should be avoided as they produce an adverse effect on many elements. The stock solutions are prepared at concentrations of 1,000 mg of the metal per liter. Commercially available standard solutions may also be used. When using pure metals (especially wire) for standards preparation, cleaning procedures, as detailed in Chapter Three, should be used to ensure that the solutions are not compromised. Stability of standards will be verified through the use of the QC protocols as specified in this method. Comparison of the daily ICVs and CCVs with the calibration curve enables the standards to be prepared as needed.

7.6.1 Aluminum -- Dissolve 1.000 g of aluminum metal in dilute HCl with gentle warming and dilute to 1 L with reagent water.

7.6.2 Antimony -- Carefully weigh 2.743 g of antimony potassium tartrate, $\text{K}(\text{SbO})\text{C}_4\text{H}_4\text{O}_6 \cdot 1/2\text{H}_2\text{O}$, and dissolve in reagent water. Dilute to 1 L with reagent water.

7.6.3 Barium -- Dissolve 1.779 g of barium chloride, $\text{BaCl}_2 \cdot 2\text{H}_2\text{O}$, analytical grade and dilute to 1 L with reagent water.

7.6.4 Beryllium -- Dissolve 11.659 g of beryllium sulfate, BeSO_4 , in reagent water containing 2 mL of nitric acid (conc.) and dilute to 1 L with reagent water.

7.6.5 Cadmium -- Dissolve 1.000 g of cadmium metal in 20 mL of 1:1 HNO_3 and dilute to 1 L with reagent water.

7.6.6 Calcium -- Suspend 2.500 g of calcium carbonate, CaCO_3 , dried for 1 hr at 180 °C in reagent water and dissolve by adding a minimum of dilute HCl . Dilute to 1 L with reagent water.

7.6.7 Chromium -- Dissolve 1.923 g of chromium trioxide, CrO_3 , in reagent water, acidify (to pH # 2) with redistilled HNO_3 (conc.), and dilute to 1 L with reagent water.

7.6.8 Cobalt -- Dissolve 1.000 g of cobalt metal in 20 mL of 1:1 HNO_3 and dilute to 1 L with reagent water. Chloride or nitrate salts of cobalt(II) may be used. Although numerous hydrated forms exist, they are not recommended unless the exact composition of the compound is known.

7.6.9 Copper -- Dissolve 1.000 g of electrolytic copper in 5 mL of redistilled HNO₃ (conc.) and dilute to 1 L with reagent water.

7.6.10 Iron -- Dissolve 1.000 g of iron wire in 10 mL redistilled HNO₃ (conc.) and reagent water and dilute to 1 L with reagent water. Note that iron passivates in conc. HNO₃, and therefore some water should be present.

7.6.11 Lead -- Dissolve 1.599 g of lead nitrate, Pb(NO₃)₂, in reagent water, acidify with 10 mL of redistilled HNO₃ (conc.), and dilute to 1 L with reagent water.

7.6.12 Lithium -- Dissolve 5.324 g of lithium carbonate, Li₂CO₃, in a minimum volume of 1:1 HCl and dilute to 1 L with reagent water.

7.6.13 Magnesium -- Dissolve 1.000 g of magnesium metal in 20 mL 1:1 HNO₃ and dilute to 1 L with reagent water.

7.6.14 Manganese -- Dissolve 1.000 g of manganese metal in 10 mL of redistilled HNO₃ (conc.) and dilute to 1 L with reagent water.

7.6.15 Molybdenum -- Dissolve 1.840 g of ammonium molybdate, (NH₄)₆Mo₇O₂₄·4H₂O, and dilute to 1 L with reagent water.

7.6.16 Nickel -- Dissolve 1.000 g of nickel metal or 4.953 g of nickel nitrate, Ni(NO₃)₂·6H₂O, in 10 mL of HNO₃ (conc.) and dilute to 1 L with reagent water.

7.6.17 Osmium -- Procure a certified aqueous standard from a supplier and verify by comparison with a second standard. If necessary, standards can be made from osmium compounds. However, due to the toxicity of these compounds, this approach is not advised.

7.6.18 Potassium -- Dissolve 1.907 g of potassium chloride, KCl, dried at 110 EC, in reagent water and dilute to 1 L with reagent water.

7.6.19 Silver -- Dissolve 1.575 g of anhydrous silver nitrate, AgNO₃, in reagent water. Add 10 mL of HNO₃ (conc.) and dilute to 1 L with reagent water. Store in a dark-colored glass bottle in a refrigerator.

7.6.20 Sodium -- Dissolve 2.542 g of sodium chloride, NaCl, in reagent water, acidify with 10 mL of redistilled HNO₃ (conc.), and dilute to 1 L with reagent water.

7.6.21 Strontium -- Dissolve 2.415 g of strontium nitrate, Sr(NO₃)₂, in 10 mL of conc. HCl and 700 mL of reagent water. Dilute to 1 L with reagent water.

7.6.22 Thallium -- Dissolve 1.303 g of thallium nitrate, TlNO₃, in reagent water, acidify (to pH # 2) with 10 mL of conc. HNO₃, and dilute to 1 L with reagent water.

7.6.23 Tin -- Dissolve 1.000 g of tin metal in 100 mL conc. HCl and dilute to 1 L with reagent water.

7.6.24 Vanadium -- Dissolve 1.785 g of vanadium pentoxide, V₂O₅, in 10 mL of conc. HNO₃ and dilute to 1 L with reagent water.

7.6.25 Zinc -- Dissolve 1.000 g of zinc metal in 10 mL of conc. HNO₃ and dilute to 1 L with reagent water.

7.7 Aluminum nitrate solution -- Dissolve 139 g of aluminum nitrate, $\text{Al}(\text{NO}_3)_3 \cdot 9\text{H}_2\text{O}$, in 150 mL reagent water and heat to effect solution. Allow to cool and make to 200 mL. Add 2 mL of this solution to a 100 mL volume of standards and samples.

7.8 Lanthanum chloride/potassium chloride solution -- Dissolve 11.73 g of lanthanum oxide, La_2O_3 , in a minimum amount (approximately 50 mL) of conc. HCl. Add 1.91 g of potassium chloride, KCl. Allow solution to cool to room temperature and dilute to 100 mL with reagent water.

WARNING: REACTION IS VIOLENT!

Add acid slowly and in small portions to control the reaction rate upon mixing.

7.9 Blanks

Two types of blanks are required for the analysis of samples prepared by any method other than Method 3040. The calibration blank is used in establishing the analytical curve and the method blank is used to identify possible contamination resulting from either the reagents (acids) or the equipment used during sample processing including filtration.

7.9.1 The calibration blank is prepared by acidifying reagent water to the same concentrations of the acids found in the standards and samples. Prepare a sufficient quantity to flush the system between standards and samples. The calibration blank will also be used for all initial (ICB) and continuing calibration blank (CCB) determinations.

7.9.2 The method blank must contain all of the reagents in the same volumes as used in the processing of the samples. The method blank must be carried through the complete procedure and contain the same acid concentration in the final solution as the sample solution used for analysis (refer to Sec. 9.5).

7.10 The initial calibration verification (ICV) standard is prepared by the analyst (or a purchased second source reference material) by combining compatible elements from a standard source different from that of the calibration standard, and at concentration near the midpoint of the calibration curve (see Sec. 10.2.1 for use). This standard may also be purchased.

7.11 The continuing calibration verification (CCV) standard should be prepared in the same acid matrix using the same standards used for calibration, at a concentration near the mid-point of the calibration curve (see Sec. 10.2.2 for use).

8.0 SAMPLE COLLECTION, PRESERVATION, AND STORAGE

See the introductory material in Chapter Three, "Inorganic Analytes."

9.0 QUALITY CONTROL

9.1 Refer to Chapter One for additional guidance on quality assurance (QA) and quality control (QC) protocols. When inconsistencies exist between QC guidelines, method-specific QC criteria take precedence over both technique-specific criteria and those criteria given in Chapter One, and technique-specific QC criteria take precedence over the criteria in Chapter One. Any effort involving the collection of analytical data should include development of a structured and systematic planning document, such as a Quality Assurance Project Plan

(QAPP) or a Sampling and Analysis Plan (SAP), which translates project objectives and specifications into directions for those that will implement the project and assess the results. Each laboratory should maintain a formal quality assurance program. The laboratory should also maintain records to document the quality of the data generated. All data sheets and quality control data should be maintained for reference or inspection.

9.2 Refer to a 3000 series method (Method 3005, 3010, 3015, 3031, 3040, 3050, 3051, or 3052) for appropriate QC procedures to ensure the proper operation of the various sample preparation techniques.

9.3 Instrument detection limits (IDLs) are a useful tool to evaluate the instrument noise level and response changes over time for each analyte from a series of reagent blank analyses to obtain a calculated concentration. They are not to be confused with the lower limit of quantitation, nor should they be used in establishing this limit. It may be helpful to compare the calculated IDLs to the established lower limit of quantitation, however, it should be understood that the lower limit of quantitation needs to be verified according to the guidance in Sec. 10.2.3.

IDLs in $\mu\text{g/L}$ can be estimated by calculating the average of the standard deviations of three runs on three non-consecutive days from the analysis of a reagent blank solution with seven consecutive measurements per day. Each measurement should be performed as though it were a separate analytical sample (i.e., each measurement must be followed by a rinse and/or any other procedure normally performed between the analysis of separate samples). IDLs should be determined at least every three months or at a project-specific designated frequency and kept with the instrument log book.

9.4 Initial demonstration of proficiency

Each laboratory must demonstrate initial proficiency with each sample preparation (a 3000 series method) and determinative method combination it utilizes by generating data of acceptable accuracy and precision for target analytes in a clean matrix. If an autosampler is used to perform sample dilutions, before using the autosampler to dilute samples, the laboratory should satisfy itself that those dilutions are of equivalent or better accuracy than is achieved by an experienced analyst performing manual dilutions. The laboratory must also repeat the demonstration of proficiency whenever new staff members are trained or significant changes in instrumentation are made.

9.5 For each batch of samples processed, at least one method blank must be carried throughout the entire sample preparation and analytical process, as described in Chapter One. A method blank is prepared by using a volume or weight of reagent water at the volume or weight specified in the preparation method, and then carried through the appropriate steps of the analytical process. These steps may include, but are not limited to, prefiltering, digestion, dilution, filtering, and analysis. If the method blank does not contain target analytes at a level that interferes with the project-specific DQOs, then the method blank would be considered acceptable.

In the absence of project-specific DQOs, if the blank is less than 10% of the lower limit of quantitation check sample concentration, less than 10% of the regulatory limit, or less than 10% of the lowest sample concentration for each analyte in a given preparation batch, whichever is greater, then the method blank is considered acceptable. If the method blank cannot be considered acceptable, the method blank should be re-run once, and if still unacceptable, then all samples after the last acceptable method blank should be reprepared and reanalyzed along with the other appropriate batch QC samples. These blanks will be useful in determining if samples are being contaminated. If the method blank exceeds the criteria, but the samples are all either below the reporting level or below the applicable action level or other DQOs, then the

sample data may be used despite the contamination of the method blank. Refer to Chapter One for the proper protocol when analyzing blanks.

9.6 Laboratory control sample (LCS)

For each batch of samples processed, at least one LCS must be carried throughout the entire sample preparation and analytical process as described in Chapter One. The laboratory control samples should be spiked with each analyte of interest at the project-specific action level or, when lacking project-specific action levels, at approximately mid-point of the linear dynamic range. Acceptance criteria should either be defined in the project-specific planning documents or set at a laboratory derived limit developed through the use of historical analyses. In the absence of project-specific or historical data generated criteria, this limit should be set at $\pm 20\%$ of the spiked value. Acceptance limits derived from historical data should be no wider than $\pm 20\%$. If the laboratory control sample is not acceptable, then the laboratory control sample should be re-run once and, if still unacceptable, all samples after the last acceptable laboratory control sample should be reprepared and reanalyzed.

Concurrent analyses of reference materials (SRMs) containing known amounts of analytes in the media of interest are recommended and may be used as an LCS. For solid SRMs, 80 - 120% accuracy may not be achievable and the manufacturer's established acceptance criterion should be used for soil SRMs.

9.7 Matrix spike, unspiked duplicate, or matrix spike duplicate (MS/Dup or MS/MSD)

Documenting the effect of the matrix, for a given preparation batch consisting of similar sample characteristics, should include the analysis of at least one matrix spike and one duplicate unspiked sample or one matrix spike/matrix spike duplicate pair. The decision on whether to prepare and analyze duplicate samples or a matrix spike/matrix spike duplicate must be based on a knowledge of the samples in the sample batch or as noted in the project-specific planning documents. If samples are expected to contain target analytes, then laboratories may use one matrix spike and a duplicate analysis of an unspiked field sample. If samples are not expected to contain target analytes, laboratories should use a matrix spike and matrix spike duplicate pair.

For each batch of samples processed, at least one MS/Dup or MS/MSD sample set should be carried throughout the entire sample preparation and analytical process as described in Chapter One. MS/MSDs are intralaboratory split samples spiked with identical concentrations of each analyte of interest. The spiking occurs prior to sample preparation and analysis. An MS/Dup or MS/MSD is used to document the bias and precision of a method in a given sample matrix.

Refer to Chapter One for definitions of bias and precision, and for the proper data reduction protocols. MS/MSD samples should be spiked at the same level, and with the same spiking material, as the corresponding laboratory control sample that is at the project-specific action level or, when lacking project-specific action levels, at approximately mid-point of the linear dynamic range. Acceptance criteria should either be defined in the project-specific planning documents or set at a laboratory-derived limit developed through the use of historical analyses per matrix type analyzed. In the absence of project-specific or historical data generated criteria, these limits should be set at $\pm 25\%$ of the spiked value for accuracy and 20 relative percent difference (RPD) for precision. Acceptance limits derived from historical data should be no wider than $\pm 25\%$ for accuracy and 20% for precision. Refer to Chapter One for additional guidance. If the bias and precision indicators are outside the laboratory control limits, if the percent recovery is less than 75% or greater than 125%, or if the relative percent

difference is greater than 20%, then the interference test discussed in Sec. 9.8 should be conducted.

9.7.1 The relative percent difference between spiked matrix duplicate or unspiked duplicate determinations is to be calculated as follows:

$$\text{RPD} = \frac{D_1 \text{ \& } D_2}{\left(\frac{D_1 + D_2}{2} \right)} \times 100$$

where:

RPD = relative percent difference.

D_1 = first sample value.

D_2 = second sample value (spiked or unspiked duplicate).

9.7.2 The spiked sample or spiked duplicate sample recovery should be within $\pm 25\%$ of the actual value, or within the documented historical acceptance limits for each matrix.

9.8 If less than acceptable accuracy and precision data are generated, the following additional quality control tests are recommended prior to reporting concentration data for the elements in this method. At a minimum these tests, outlined in Secs. 9.8.1 and 9.8.2, should be performed with each batch of samples prepared/analyzed with corresponding unacceptable data quality results. These tests will then serve to ensure that neither positive nor negative interferences are affecting the measurement of any of the elements or distorting the accuracy of the reported values. If matrix effects are confirmed, the laboratory should consult with the data user when feasible for possible corrective actions which may include the use of alternative or modified test procedures or possibly the method of standard additions so that the analysis is not impacted by the same interference.

9.8.1 Post digestion spike addition

The same sample from which the MS/MSD aliquots were prepared (assuming the MS/MSD recoveries are unacceptable) should also be spiked with a post digestion spike. Otherwise another sample from the same preparation should be used as an alternative. An analyte spike is added to a portion of a prepared sample, or its dilution, and should be recovered to within 80% to 120% of the known value. The spike addition should produce a minimum level of 10 times and a maximum of 100 times the lower limit of quantitation. If this spike fails, then the dilution test (Sec. 9.8.2) should be run on this sample. If both the MS/MSD and the post digestion spike fail, then matrix effects are confirmed.

9.8.2 Dilution test

If the analyte concentration is sufficiently high (minimally, a factor of 10 above the lower limit of quantitation after dilution), an analysis of a 1:5 dilution should agree within $\pm 10\%$ of the original determination. If not, then a chemical or physical interference effect should be suspected. For both a failed post digestion spike or an unacceptable dilution test agreement result, the method of standard additions should be used as the primary means to quantitate all samples in the associated preparation batch.

9.9 Where the sample matrix is so complex that viscosity, surface tension, and components cannot be accurately matched with standards, the method of standard additions (MSA) is recommended (see Sec. 9.10 below). Other options including the use of different matrix modifiers, different furnace conditions, different preparatory methods or different analytical methods may also be attempted to properly characterize a sample. Sec. 9.8 provides tests to determine the potential of an interference and evaluates the need for using the MSA.

9.10 Method of standard additions -- The standard addition technique involves adding known amounts of standard to one or more aliquots of the processed sample solution. This technique attempts to compensate for a sample constituent that enhances or depresses the analyte signal, thus producing a different slope from that of the calibration standards. It will not correct for additive interferences which cause a baseline shift. The method of standard additions may be appropriate for analysis of extracts, on analyses submitted as part of a delisting petition, whenever a new sample matrix is being analyzed and on every batch that fails the recovery test.

9.10.1 The simplest version of this technique is the single-addition method, in which two identical aliquots of the sample solution, each of volume V_x , are taken. To the first (labeled A) is added a known volume V_s of a standard analyte solution of concentration C_s . To the second aliquot (labeled B) is added the same volume V_s of reagent water. The analytical signals of A and B are measured and corrected for non-analyte signals. The unknown sample concentration C_x is calculated:

$$C_x = \frac{S_B V_s C_s}{(S_A - S_B) V_x}$$

where S_A and S_B are the analytical signals (corrected for the blank) of solutions A and B, respectively. V_s and C_s should be chosen so that S_A is roughly twice S_B on the average, avoiding excess dilution of the sample. If a separation or concentration step is used, the additions are best made first and carried through the entire procedure.

9.10.2 Improved results can be obtained by employing a series of standard additions. To equal volumes of the sample are added a series of standard solutions containing different known quantities of the analyte, and all solutions are diluted to the same final volume. For example, addition 1 should be prepared so that the resulting concentration is approximately 50 percent of the expected absorbance from the indigenous analyte in the sample. Additions 2 and 3 should be prepared so that the concentrations are approximately 100 and 150 percent of the expected endogenous sample absorbance. The absorbance of each solution is determined and then plotted on the vertical axis of a graph, with the concentrations of the known standards plotted on the horizontal axis. When the resulting line is extrapolated to zero absorbance, the point of interception of the abscissa is the endogenous concentration of the analyte in the sample. The abscissa on the left of the ordinate is scaled the same as on the right side, but in the opposite direction from the ordinate. An example of a plot so obtained is shown in Figure 1. A linear regression program may be used to obtain the intercept concentration.

9.10.3 For the results of this MSA technique to be valid, the following limitations must be taken into consideration:

1. The apparent concentrations from the calibration curve must be linear (0.995 or greater) over the concentration range of concern. For the best results, the slope of the MSA plot should be nearly the same as the slope of the standard curve.
2. The effect of the interference should not vary as the ratio of analyte concentration to sample matrix changes, and the standard addition should respond in a similar manner as the analyte.
3. The determination must be free of spectral interference and corrected for nonspecific background interference.

9.11 Ultra-trace analysis requires the use of clean chemistry preparation and analysis techniques. Several suggestions for minimizing analytical blank contamination are provided in Chapter Three.

10.0 CALIBRATION AND STANDARDIZATION

10.1 Calibration standards -- For those instruments which do not read out directly in concentration, a calibration curve is prepared to cover the appropriate concentration range. Usually, this means the preparation of a blank and standards which produce an absorbance of 0.0 to 0.7. Calibration standards can be prepared by diluting the stock metal solutions in the same acids and acid concentrations as the samples.

10.1.1 Calibration standards can be prepared fresh each time a batch of samples is analyzed. If the ICV solution is prepared daily and the ICV is analyzed within the acceptance criteria, calibration standards do not need to be prepared daily and may be prepared and stored for as long as the calibration standard viability can be verified through the use of the ICV. If the ICV is outside of the acceptance criteria, the calibration standards must be prepared fresh and the instrument recalibrated. Prepare a blank and at least three calibration standards in graduated amounts in the appropriate range of the linear part of the curve.

10.1.2 The calibration standards should be prepared using the same type of acid or combination of acids and at the same concentration as will result in the samples following processing.

10.1.3 Beginning with the calibration blank and working toward the highest standard, aspirate the solutions and record the readings. Repeat the operation with both the calibration standards and the samples a sufficient number of times to secure an average reading for each solution. Calibration curves are always required.

10.2 A calibration curve must be prepared each day with a minimum of a calibration blank and three standards. The curve must be linear and have a correlation coefficient of at least 0.995.

10.2.1 After initial calibration, the calibration curve must be verified by use of an initial calibration blank (ICB) and an initial calibration verification (ICV) standard. The ICV standard must be made from an independent (second source) material at or near mid-range. The acceptance criteria for the ICV standard must be $\pm 10\%$ of its true value and the ICB must not contain target analytes at or above the lowest limit of quantitation for the curve to be considered valid. If the calibration curve cannot be verified within the specified limits, the cause must be determined and the instrument recalibrated before samples are

analyzed. The analysis data for the ICV must be kept on file with the sample analysis data.

10.2.2 The calibration curve must also be verified at the end of each analysis batch and/or after every 10 samples by use of a continuing calibration blank (CCB) and a continuing calibration verification (CCV) standard. The CCV standard should be made from the same material as the initial calibration standards at or near midrange. The acceptance criteria for the CCV standard must be $\pm 10\%$ of its true value and the CCB must not contain target analytes at or above the lowest limit of quantitation for the curve to be considered valid. If the calibration cannot be verified within the specified limits, the sample analysis must be discontinued, the cause determined and the instrument recalibrated. All samples following the last acceptable CCV/CCB must be reanalyzed. The analysis data for the CCV/CCB must be kept on file with the sample analysis data.

10.2.3 The lower limits of quantitation should be established for all analytes for each type of matrix analyzed and for each preparation method used and for each instrument. These limits are considered the lowest reliable laboratory reporting concentrations and should be established from the lower limit of quantitation check sample and then confirmed using either the lowest calibration point or from a low-level calibration check standard.

10.2.3.1 Lower limit of quantitation check sample

The lower limit of quantitation check (LLQC) sample should be analyzed after establishing the lower laboratory reporting limits and on an as needed basis to demonstrate the desired detection capability. Ideally, this check sample and the low-level calibration verification standard will be prepared at the same concentrations with the only difference being the LLQC sample is carried through the entire preparation and analytical procedure. Lower limits of quantitation are verified when all analytes in the LLQC sample are detected within $\pm 30\%$ of their true value. This check should be used to both establish and confirm the lowest quantitation limit.

10.2.3.2 The lower limits of quantitation determination using reagent water represents a best case situation and does not represent possible matrix effects of real-world samples. For the application of lower limits of quantitation on a project-specific basis with established data quality objectives, low-level matrix-specific spike studies may provide data users with a more reliable indication of the actual method sensitivity and minimum detection capabilities.

10.3 It is recommended that each standard should be analyzed (injected) twice and an average value determined. Replicate standard values should be within $\pm 10\%$ RPD.

10.4 If conducting trace analysis, it is recommended that the lowest calibration standard be set at the laboratory's lower limit of quantitation. The laboratory can use a reporting limit that is below the lower limit of quantitation but all values reported below the low standard should be reported as estimated values.

11.0 PROCEDURE

11.1 Preliminary treatment of aqueous and solid wastes is always necessary because of the complexity and variability of sample matrices. Solids, slurries, and suspended material must be subjected to a solubilization process before analysis. This process may vary because of the

metals to be determined and the nature of the sample being analyzed. Solubilization and digestion procedures are presented in Chapter Three. Samples which are to be analyzed for dissolved constituents need not be digested if they have been filtered and then acidified. Also see the note in Sec. 1.1.

11.2 All atomic absorption analyses must be performed using a suitable form of background correction. Refer to Chapter Three for a information regarding background correction.

11.3 Differences between the various makes and models of satisfactory atomic absorption spectrophotometers prevent the formulation of detailed instructions applicable to every instrument. The analyst should follow the manufacturer's operating instructions for a particular instrument.

11.3.1 In general, after choosing the proper lamp for the analysis, allow the lamp to warm up for a minimum of 15 minutes.

11.3.2 During this period, align the instrument, position the monochromator at the correct wavelength, select the proper monochromator slit width, and adjust the current according to the manufacturer's recommendation.

11.3.3 Light the flame and regulate the flow of fuel and oxidant. Adjust the burner and nebulizer flow rate for maximum percent absorption and stability. Balance the photometer.

11.3.4 Run a series of standards of the element under analysis. Construct a calibration curve by plotting the concentrations of the standards against absorbances. Set the curve corrector of a direct reading instrument to read out the proper concentration.

11.3.5 Aspirate the samples and determine the concentrations either directly or from the calibration curve. Standards must be run each time a sample or series of samples is run.

12.0 DATA ANALYSIS AND CALCULATIONS

12.1 For determination of metal concentration, read the concentration from the calibration curve or directly from the read-out system of the instrument.

12.1.1 If dilution of the sample was required:

$$\mu\text{g/L metal in sample} = \frac{A (C\%B)}{C}$$

where:

A = $\mu\text{g/L}$ of metal in diluted aliquot from calibration curve.
B = Starting sample volume, mL.
C = Final volume of sample, mL.

12.1.2 For solid samples, report all concentrations in consistent units based on weight. Ensure that, if the dry weight was used for the analysis, percent solids are reported to the client.

$$\text{mg metal/kg sample} = \frac{A \times V}{W}$$

where:

A = mg/L of metal in processed sample from calibration curve.
V = Final volume of the processed sample, L.
W = Weight of sample, Kg.

12.1.3 Different integration times must not be used for samples and standards. Instead, the sample should be diluted and the same integration time should be used for both samples and standards. If dilution of the sample was required:

$$\mu\text{/L of metal sample} = \frac{Z (C \% B)}{C}$$

where:

Z = $\mu\text{g/L}$ of metal read from calibration curve or read-out system.
B = Starting sample volume, mL.
C = Final volume of sample, mL.

12.2 Results need to be reported in units commensurate with their intended use and all dilutions need to be taken into account when computing final results.

13.0 METHOD PERFORMANCE

13.1 Performance data and related information are provided in SW-846 methods only as examples and guidance. The data do not represent required performance criteria for users of the methods. Instead, performance criteria should be developed on a project-specific basis, and the laboratory should establish in-house QC performance criteria for the application of this method. These performance data are not intended to be and must not be used as absolute QC acceptance criteria for purposes of laboratory accreditation.

13.2 For relevant performance data, see the methods of Ref. 1.

14.0 POLLUTION PREVENTION

14.1 Pollution prevention encompasses any technique that reduces or eliminates the quantity and/or toxicity of waste at the point of generation. Numerous opportunities for pollution prevention exist in laboratory operation. The EPA has established a preferred hierarchy of environmental management techniques that places pollution prevention as the management option of first choice. Whenever feasible, laboratory personnel should use pollution prevention

techniques to address their waste generation. When wastes cannot be feasibly reduced at the source, the Agency recommends recycling as the next best option.

14.2 For information about pollution prevention that may be applicable to laboratories and research institutions consult *Less is Better: Laboratory Chemical management for Waste Reduction* available from the American Chemical Society's Department of Government Relations and Science Policy, 1155 16th St., N.W. Washington, D.C. 20036, <http://www.acs.org>.

15.0 WASTE MANAGEMENT

The Environmental Protection Agency requires that laboratory waste management practices be conducted consistent with all applicable rules and regulations. The Agency urges laboratories to protect the air, water, and land by minimizing and controlling all releases from hoods and bench operations, complying with the letter and spirit of any sewer discharge permits and regulations, and by complying with all solid and hazardous waste regulations, particularly the hazardous waste identification rules and land disposal restrictions. For further information on waste management, consult *The Waste Management Manual for Laboratory Personnel* available from the American Chemical Society at the address listed in Sec. 14.2.

16.0 REFERENCES

1. Methods for Chemical Analysis of Water and Wastes; U.S. Environmental Protection Agency. Office of Research and Development. Environmental Monitoring and Support Laboratory. ORD Publication Offices of Center for Environmental Research Information: Cincinnati, OH, 1983; EPA-600/4-79-020.
2. W. G. Rohrbough, et al., Reagent Chemicals, American Chemical Society Specifications, 7th ed.; American Chemical Society: Washington, DC, 1986.
3. 1985 Annual Book of ASTM Standards, Vol. 11.01; "Standard Specification for Reagent Water"; ASTM: Philadelphia, PA, 1985; D1193-77.

17.0 TABLES, DIAGRAMS, FLOWCHARTS, AND VALIDATION DATA

The following pages contain the tables and figure referenced by this method. A flow diagram of the procedure follows the tables.

TABLE 1

EXAMPLE ATOMIC ABSORPTION LOWER LIMITS OF QUANTITATION AND SENSITIVITY FOR ANALYTES IN REAGENT WATER

Metal	Direct Aspiration	
	Lower Limit of Quantitation (mg/L)	Sensitivity (mg/L)
Aluminum	0.1	1
Antimony	0.2	0.5
Barium	0.1	0.4
Beryllium	0.005	0.025
Cadmium	0.005	0.025
Calcium	0.01	0.08
Chromium	0.05	0.25
Cobalt	0.05	0.2
Copper	0.02	0.1
Iron	0.03	0.12
Lead	0.1	0.5
Lithium	0.002	0.04
Magnesium	0.001	0.007
Manganese	0.01	0.05
Molybdenum	0.1	0.4
Nickel	0.04	0.15
Osmium	0.03	1
Potassium	0.01	0.04
Silver	0.01	0.06
Sodium	0.002	0.015
Strontium	0.03	0.15
Thallium	0.1	0.5
Tin	0.8	4
Vanadium	0.2	0.8
Zinc	0.005	0.02

These data are provided for guidance purposes only.


TABLE 2

INSTRUMENT PARAMETERS (Ref. 1)

ELEMENT	WAVELENGTH (nm)	FUEL	OXIDANT	TYPE OF FLAME
Al	324.7	acetylene	nitrous oxide	fuel rich
Sb	<u>217.6</u> , 231.1	acetylene	air	fuel lean
Ba	553.6	acetylene	nitrous oxide	fuel rich
Be	234.9	acetylene	nitrous oxide	fuel rich
Cd	228.8	acetylene	air	fuel lean
Ca	422.7	acetylene	nitrous oxide	stoichiometric
Cr	357.9	acetylene	nitrous oxide	fuel rich
Co	240.7	acetylene	air	fuel lean
Cu	324.7	acetylene	air	fuel lean
Fe	<u>248.3</u> , 248.8, 271.8, 302.1, 252.7	acetylene	air	fuel lean
Pb	<u>283.3</u> , 217.0	acetylene	air	fuel lean
Li	670.8	acetylene	air	fuel lean
Mg	285.2	acetylene	air	fuel lean
Mn	<u>279.5</u> , 403.1	acetylene	air	fuel lean to stoichiometric
Mo	313.3	acetylene	nitrous oxide	fuel rich
Ni	<u>232.0</u> , 352.4	acetylene	air	fuel lean
Os	290.0	acetylene	nitrous oxide	fuel rich
K	766.5	acetylene	air	fuel lean
Ag	328.1	acetylene	air	fuel lean
Na	589.6	acetylene	air	fuel lean
Sr	460.7	acetylene	air	fuel lean
Tl	276.8	acetylene	air	fuel lean
Sn	286.3	acetylene	nitrous oxide	fuel rich
V	318.4	acetylene	nitrous oxide	fuel rich
Zn	213.9	acetylene	air	fuel lean

Note: If more than one wavelength is listed, the primary line is underlined.

FIGURE 1
STANDARD ADDITION PLOT



METHOD 7000B

FLAME ATOMIC ABSORPTION SPECTROPHOTOMETRY

