

Method 1624, Revision C: Volatile Organic Compounds by Isotope Dilution GCMS

Method 1624
Revision C
Volatile Organic Compounds by Isotope Dilution
GCMS

Method 1624

Volatile Organic Compounds by Isotope Dilution GCMS

1. SCOPE AND APPLICATION

- 1.1 This method is designed to meet the survey requirements of the USEPA ITD. The method is used to determine the volatile toxic organic pollutants associated with the Clean Water Act (as amended 1987); the Resource Conservation and Recovery Act (as amended in 1986); the Comprehensive Environmental Response, Compensation, and Liability Act (as amended in 1986); and other compounds amenable to purge and trap gas chromatography/mass spectrometry (GCMS).
- 1.2 The chemical compounds listed in Tables 1 and 2 may be determined in waters, soils, and municipal sludges by the method.
- 1.3 The detection limits of the method are usually dependent on the level of interferences rather than instrumental limitations. The levels in Table 3 typify the minimum quantities that can be detected with no interferences present.
- 1.4 The GCMS portions of the method are for use only by analysts experienced with GCMS or under the close supervision of such qualified persons. Laboratories unfamiliar with analysis of environmental samples by GCMS should run the performance tests in Reference 1 before beginning.

2. SUMMARY OF METHOD

- 2.1 The percent solids content of the sample is determined. If the solids content is known or determined to be less than 1%, stable isotopically labeled analogs of the compounds of interest are added to a 5-mL sample and the sample is purged with an inert gas at 20 to 25°C in a chamber designed for soil or water samples. If the solids content is greater than one, mL of reagent water and the labeled compounds are added to a 5-*aliquot* of sample and the mixture is purged at 40°C. Compounds that will not purge at 20 to 25°C or at 40°C are purged at 75 to 85°C (see Table 2). In the purging process, the volatile compounds are transferred from the aqueous phase into the gaseous phase where they are passed into a sorbent column and trapped. After purging is completed, the trap is backflushed and heated rapidly to desorb the compounds into a gas chromatograph (GC). The compounds are separated by the GC and detected by a mass spectrometer (MS) (References 2 and 3). The labeled compounds serve to correct the variability of the analytical technique.
- 2.2 Identification of a pollutant (qualitative analysis) is performed in one of three ways: (1) For compounds listed in Table 1 and other compounds for which authentic standards are available, the GCMS system is calibrated and the mass spectrum and retention time for each standard are stored in a user created library. A compound is identified when its retention time and mass spectrum agree with the library retention time and spectrum. (2) For compounds listed in Table 2 and other compounds for which standards are not available, a compound is identified when the retention time and mass spectrum agree with those specified in this method. (3) For chromatographic peaks which are not identified by (1) and (2) above, the background corrected spectrum at the peak maximum

is compared with spectra in the EPA/NIH mass spectral file (Reference 4). Tentative identification is established when the spectrum agrees (see Section 12).

- 2.3** Quantitative analysis is performed in one of four ways by GCMS using extracted ion current profile (EICP) areas: (1) For compounds listed in Table 1 and other compounds for which standards and labeled analogs are available, the GCMS system is calibrated and the compound concentration is determined using an isotope dilution technique. (2) For compounds listed in Table 1 and for other compounds for which authentic standards but no labeled compounds are available, the GCMS system is calibrated and the compound concentration is determined using an internal standard technique. (3) For compounds listed in Table 2 and other compounds for which standards are not available, compound concentrations are determined using known response factors. (4) For compounds for which neither standards nor known response factors are available, compound concentration is determined using the sum of the EICP areas relative to the sum of the EICP areas of the nearest eluted internal standard.
- 2.4** The quality of the analysis is assured through reproducible calibration and testing of the purge and trap and GCMS systems.

Table 1. Volatile Organic Compounds Determined by GCMS Using Isotope Dilution and Internal Standard Techniques

Compound	Pollutant				Labeled Compound		
	STORET	CAS Registry	EPA EGD	NPDES	Analog	CAS Registry	EPA EGD
Acetone	81552	67-64-1	516 V		d ₆	666-52-4	616 V
Acrolein	34210	107-02-8	002 V	001 V	d ₄	33984-05-3	202 V
Acrylonitrile	34215	107-13-1	003 V	002 V	d ₃	53807-26-4	203 V
Benzene	34030	71-43-2	004 V	003 V	d ₆	1076-43-3	204 V
Bromodichloromethane	32101	75-27-4	048 V	012 V	¹³ C	93952-10-4	248 V
Bromoform	32104	75-25-2	047 V	005 V	¹³ C	72802-81-4	247 V
Bromomethane	34413	74-83-9	046 V	020 V	d ₃	1111-88-2	246 V
Carbon tetrachloride	32102	56-23-5	006 V	006 V	¹³ C	32488-50-9	206 V
Chlorobenzene	34301	108-90-7	007 V	007 V	d ₅	3114-55-4	207 V
Chloroethane	34311	75-00-3	016 V	009 V	d ₅	19199-91-8	216 V
2-Chloroethylvinyl ether	34576	110-75-8	019 V	010 V			
Chloroform	32106	67-66-3	023 V	011 V	¹³ C	31717-44-9	223 V
Chloromethane	34418	74-87-3	045 V	021 V	d ₃	1111-89-3	245 V
Dibromochloromethane	32105	124-48-1	051 V	008 V	¹³ C	93951-99-6	251 V
1,1-Dichloroethane	34496	75-34-3	013 V	014 V	d ₃	56912-77-7	213 V
1,2-Dichloroethane	32103	107-06-2	010 V	015 V	d ₄	17070-07-0	210 V
1,1-Dichloroethene	34501	75-35-4	029 V	016 V	d ₂	22280-73-5	229 V
trans-1,2-Dichloroethene	34546	156-60-5	030 V	026 V	d ₃	42366-47-2	230 V
1,2-Dichloropropane	34541	78-87-5	032 V	017 V	d ₆	93952-08-0	232 V
trans-1,3-Dichloropropene	34699	10061-02-6	033 V		d ₄	93951-86-1	233 V
Diethyl ether	81576	60-29-7	515 V		d ₁₀	2679-89-2	615 V
<i>p</i> -Dioxane	81582	123-91-1	527 V		d ₈	17647-74-4	627 V
Ethylbenzene	34371	100-41-4	038 V	019 V	d ₁₀	25837-05-2	238 V
Methylene chloride	34423	75-09-2	044 V	022 V	d ₂	1665-00-5	244 V
Methyl ethyl ketone	81595	78-93-3	514 V		d ₃	53389-26-7	614 V
1,1,2,2-Tetrachloroethane	34516	79-34-5	015 V	023 V	d ₂	33685-54-0	215 V
Tetrachloroethene	34475	127-18-4	085 V	024 V	¹³ C ₂	32488-49-6	285 V
Toluene	34010	108-88-3	086 V	025 V	d ₈	2037-26-5	286 V
1,1,1-Trichloroethane	34506	71-55-6	011 V	027 V	d ₃	2747-58-2	211 V
1,1,2-Trichloroethane	34511	79-00-5	014 V	028 V	¹³ C ₂	93952-09-1	214 V
Trichloroethene	39180	79-01-6	087 V	029 V	¹³ C ₂	93952-00-2	287 V
Vinyl chloride	39175	75-01-4	088 V	031 V	d ₃	6745-35-3	288 V

Table 2. Volatile Organic Compounds to be Determined by Reverse Search and Quantitation Using Known Retention Times, Response Factors, Reference Compounds, and Mass Spectra

EGD No.	Compound	CAS Registry
532	Allyl alcohol ¹	107-18-6
533	Carbon disulfide	75-15-0
534	2-Chloro-1,3-butadiene (Chloroprene)	126-99-8
535	Chloroacetonitrile ¹	107-14-2
536	3-Chloropropene	107-05-1
537	Crotonaldehyde ¹	123-73-9
538	1,2-Dibromoethane (EDB)	106-93-3
539	Dibromomethane	74-95-3
540	trans-1,4-Dichloro-2-butene	110-57-6
541	1,3-Dichloropropane	142-28-9
542	cis-1,3-Dichloropropene	10061-01-5
543	Ethyl cyanide ¹	107-12-0
544	Ethyl methacrylate	97-63-2
545	2-Hexanone	591-78-6
546	Iodomethane	74-88-4
547	Isobutyl alcohol ¹	78-83-1
548	Methacrylonitrile	126-98-7
549	Methyl methacrylate	78-83-1
550	4-Methyl-2-pentanone	108-10-1
551	1,1,1,2-Tetrachloroethane	630-20-6
552	Trichlorofluoromethane	75-69-4
553	1,2,3-Trichloropropane	96-18-4
554	Vinyl acetate	108-05-4
951	<i>m</i> -Xylene	108-38-3
952	<i>o</i> - and <i>p</i> -Xylene	

¹ Determined at a purge temperature of 75–85°C.

3. CONTAMINATION AND INTERFERENCES

- 3.1** Impurities in the purge gas, organic compounds out-gassing from the plumbing upstream of the trap, and solvent vapors in the laboratory account for the majority of contamination problems. The analytical system is demonstrated to be free from interferences under conditions of the analysis by analyzing reagent water blanks initially and with each sample batch (samples analyzed on the same 8-hour shift), as described in Section 8.5.
- 3.2** Samples can be contaminated by diffusion of volatile organic compounds (particularly methylene chloride) through the bottle seal during shipment and storage. A field blank prepared from reagent water and carried through the sampling and handling protocol may serve as a check on such contamination.
- 3.3** Contamination by carry-over can occur when high level and low level samples are analyzed sequentially. To reduce carry-over, the purging device (Figure 1 for samples containing less than one percent solids; Figure 2 for samples containing one percent solids or greater) is cleaned or replaced with a clean purging device after each sample is analyzed. When an unusually concentrated sample is encountered, it is followed by analysis of a reagent water blank to check for carry-over. Purging devices are cleaned by washing with soap solution, rinsing with tap and distilled water, and drying in an oven at 100 to 125°C. The trap and other parts of the system are also subject to contamination; therefore, frequent bakeout and purging of the entire system may be required.
- 3.4** Interferences resulting from samples will vary considerably from source to source, depending on the diversity of the site being sampled.

Table 3. Gas Chromatography of Purgeable Organic Compounds

EGD No. ¹	Compound	Retention Time			Method Detection Limit ⁴		
		Mean (sec)	EGD Ref	Relative ²	Minimum Level ³ (µg/L)	Low Solids (µg/kg)	High Solids (µg/kg)
245	Chloromethane-d ₃	147	181	0.141-0.270	50		
345	Chloromethane	148	245	0.922-1.210	50	207 ⁷	13
246	Bromomenthane-d ₃	243	181	0.233-0.423	50		
346	Bromomethane	246	246	0.898-1.195	50	148 ⁷	11
288	Vinyl chloride-d ₃	301	181	0.286-0.501	50		
388	Vinyl chloride	304	288	0.946-1.023	10	190 ⁷	11
216	Chloroethane-d ₅	378	181	0.373-0.620	50		
316	Chloroethane	386	216	0.999-1.060	50	789 ⁷	24
244	Methylene chloride-d ₂	512	181	0.582-0.813	10		
344	Methylene chloride	517	244	0.999-1.017	10	566 ⁷	280 ⁷
546	Iodomethane	498	181	0.68			
616	Acetone-d ₆	554	181	0.628-0.889	50		
716	Acetone	565	616	0.984-1.019 ⁷	50	3561	322 ⁷
202	Acrolein-d ₄	564	181	0.641-0.903	-- ⁵	50	
302	Acrolein	566	202	0.984-1.018 ⁵	50	377 ⁷	18
203	Acrylonitrile-d ₃	606	181	0.735-0.926	50		
303	Acrylonitrile	612	203	0.985-1.030	50	360 ⁷	9
533	Carbon disulfide	631	181	0.86			
552	Trichlorofluoromethane	663	181	0.91			
543	Ethyl cyanide	672	181	0.92			
229	1,1-Dichloroethene-d ₂	696	181	0.903-0.976	10		
329	1,1-Dichloroethene	696	229	0.999-1.011	10	31	5
536	3-Chloropropene	696	181	0.95			
532	Allyl alcohol	703	181	0.96			
181	Bromochloromethane (I.S.)	730	181	1.000-1.000	10		
213	1,1-Dichloroethane-d ₃	778	181	1.031-1.119	10		
313	1,1-Dichloroethane	786	213	0.999-1.014	10	16	1
615	Diethyl ether-d ₁₀	804	181	1.067-1.254	50		
715	Diethyl ether	820	615	1.010-1.048	50	63	12
230	trans-1,2-Dichloroethene-d ₂	821	181	1.056-1.228	10		
330	trans-1,2-Dichloroethene	821	230	0.996-1.011	10	41	3
614	Methyl ethyl ketone-d ₃	840	181	0.646-1.202	50		
714	Methyl ethyl ketone	848	614	0.992-1.055	50	241 ⁷	80 ⁷
223	Chloroform- ¹³ C ₁	861	181	1.092-1.322	10		
323	Chloroform	861	223	0.961-1.009	10	21	2

EGD No. ¹	Compound	Retention Time			Method Detection Limit ⁴		
		Mean (sec)	EGD Ref	Relative ²	Minimum Level ³ (µg/L)	Low Solids (µg/kg)	High Solids (µg/kg)
535	Chloroacetonitrile	884	181	1.21			
210	1,2-Dichloroethane-d ₄	901	181	1.187–1.416	10		
310	1,2-Dichloroethane	910	210	0.973–1.032	10	23	3
539	Dibromomethane	910	181	1.25			
548	Methacrylonitrile	921	181	1.26			
547	Isobutyl alcohol	962	181	1.32			
211	1,1,1-Trichloroethane- ¹³ C ₂	989	181	1.293–1.598	10		
311	1,1,1-Trichloroethane	999	211	0.989–1.044	10	16	4
627	<i>p</i> -Dioxane-d ₈	982	181	1.262–1.448 ⁵	50		
727	<i>p</i> -Dioxane	1001	627	1.008–1.040 ⁵	50	--	140 ⁷
206	Carbon tetrachloride- ¹³ C ₂	1018	182	0.754–0.805	10		
306	Carbon tetrachloride	1018	206	0.938–1.005	10	87	9
554	Vinyl acetate	1031	182	0.79			
248	Bromodichloromethane- ¹³ C ₁	1045	182	0.766–0.825	10		
348	Bromodichloromethane	1045	248	0.978–1.013	10	28	3
534	2-Chloro-1,3-butadiene	1084	182	0.83			
537	Crotonaldehyde	1098	182	0.84			
232	1,2-Dichloropropane-d ₆	1123	182	0.830–0.880	10		
332	1,2-Dichloropropane	1134	232	0.984–1.018	10	29	5
542	cis-1,3-Dichloropropene	1138	182	0.87			
287	Trichloroethene- ¹³ C ₂	1172	182	0.897–0.917	10		
387	Trichloroethene	1187	287	0.991–1.037	10	41	2
541	1,3-Dichloropropane	1196	182	0.92			
204	Benezene-d ₆	1200	182	0.888–0.952	10		
304	Benezene	1212	204	1.002–1.026	10	23	8
251	Chlorodibromomethane- ¹³ C ₁	1222	182	0.915–0.949	10		
351	Chlorodibromomethane	1222	231	0.989–1.030	10	15	2
214	1,1,2-Trichloroethane- ¹³ C ₂	1224	182	0.922–0.953	10		
314	1,1,2-Trichloroethane	1224	214	0.975–1.027	10	26	1
233	trans-1,3-Dichloropropene- d ₄	1226	182	0.922–0.959	10		
333	trans-1,3-Dichloropropene	1226	233	0.993–1.016	10	-- ^{6,7}	-- ^{6,7}
019	2-Chloroethyvinyl ether	1278	182	0.983–1.026	10	122	21
538	1,2-Dibromoethane	1279	182	0.98			
182	2-bromo-1-chloropropane (I.S.)	1306	182	1.000–1.000	10		

EGD No. ¹	Compound	Retention Time			Method Detection Limit ⁴		
		Mean (sec)	EGD Ref	Relative ²	Minimum Level ³ (µg/L)	Low Solids (µg/kg)	High Solids (µg/kg)
549	Methyl methacrylate	1379	182	1.06			
247	Bromoform- ¹³ C ₁	1386	182	1.048–1.087	10		
347	Bromoform	1386	247	0.992–1.003	10	91	7
551	1,1,1,2-Tetrachloroethane	1408	182	1.08			
550	4-Methyl-2-pentanone	1435	183	0.92			
553	1,2,3-Trichloropropane	1520	183	0.98			
215	1,1,2,2-Tetrachloroethane-d ₂	1525	183	0.969–0.996	10		
315	1,1,2,2-Tetrachloroethane	1525	215	0.890–1.016	10	20	6
545	2-Hexanone	1525	183	0.98			
285	Tetrachloroethene- ¹³ C ₂	1528	183	0.966–0.996	10		
385	Tetrachloroethene	1528	285	0.997–1.003	10	106	10
540	trans-1,4-Dichloro-2-butene	1551	183	1.00			
183	1,4-Dichlorobutane (int std)	1555	183	1.000–1.000	10		
544	Ethyl methacrylate	1594	183	1.03			
286	Toluene-d ₈	1603	183	1.016–1.054	10		
386	Toluene	1619	286	1.001–1.019	10	27	4
207	Chlorobenzene-d ₅	1679	183	1.066–1.135	10		
307	Chlorobenzene	1679	207	0.914–1.019	10	21	58 ⁷
238	Ethylbenzene-d ₁₀	1802	183	1.144–1.293	10		
338	Ethylbenzene	1820	238	0.981–1.018	10	28	4
185	Bromofluorobenzene	1985	183	1.255–1.290	10		
951	<i>m</i> -Xylene	2348	183	1.51	10		
952	<i>o</i> - and <i>p</i> -Xylene	2446	183	1.57	10		

¹ Reference numbers beginning with 0, 1, 5, or 9 indicate a pollutant quantified by the internal standard method; reference numbers beginning with 2 or 6 indicate a labeled compound quantified by the internal standard method; reference numbers beginning with 3 or 7 indicate a pollutant quantified by isotope dilution.

² The retention time limits in this column are based on data from four wastewater laboratories. The single values for retention times in this column are based on data from one wastewater laboratory.

³ This is a minimum level at which the analytical system shall give recognizable mass spectra (background corrected) and acceptable calibration points when calibrated using reagent water. The concentration in the aqueous or solid phase is determined using the equations in Section 13.

⁴ Method detection limits determined in digested sludge (low solids) and in filter cake or compost (high solids).

⁵ Specification derived from related compound.

- ⁶ An unknown interference in the particular sludge studied precluded measurement of the method detection limit (MDL) for this compound.
- ⁷ Background levels of these compounds were present in the sludge resulting in higher than expected MDLs. The MDL for these compounds is expected to be approximately 20 µg/kg (100 to 200 µg/kg for the gases and water soluble compounds) for the low solids method and 5 to 10 µg/kg (25 to 50 µg/kg for the gases and water soluble compounds) for the high solids methods, with no interferences present.

Column: 2.4 m (8 ft) × 2 mm I.D. glass, packed with 1% SP-1000 coated on 60/80 Carbopak B.
Carrier gas: Helium at 40 mL/min.

Temperature program: 3 min at 45°C, 8°C/min to 240°C, hold at 240°C for 15 minutes.

4. SAFETY

- 4.1** The toxicity or carcinogenicity of each compound or reagent used in this method has not been precisely determined; however, each chemical compound should be treated as a potential health hazard.

Exposure to these compounds should be reduced to the lowest possible level. The laboratory is responsible for maintaining a current awareness file of OSHA regulations regarding the safe handling of the chemicals specified in this method. A reference file of data handling sheets should also be made available to all personnel involved in these analyses. Additional information on laboratory safety can be found in References 5 through 7.

- 4.2** The following compounds covered by this method have been tentatively classified as known or suspected human or mammalian carcinogens: benzene, carbon tetrachloride, chloroform, and vinyl chloride. Primary standards of these toxic compounds should be prepared in a hood, and a NIOSH/MESA approved toxic gas respirator should be worn when high concentrations are handled.

5. APPARATUS AND MATERIALS

- 5.1** Sample bottles for discrete sampling.

5.1.1 Bottle: 25- to 40-mL with screw—cap (Pierce 13075, or equivalent). Detergent—wash, rinse with tap and distilled water, and dry at >105°C for a minimum of 1 hour before use.

5.1.2 Septum: Teflon-faced silicone (Pierce 12722, or equivalent), cleaned as above and baked at 100 to 200°C for 1 hour minimum.

- 5.2** Purge and trap device: Consists of purging device, trap, and desorber.

5.2.1 Purging devices for water and soil samples.

5.2.1.1 Purging device for water samples Designed to accept 5-mL samples with water column at least 3 cm deep. The volume of the gaseous head space between the water and trap shall be less than 15 mL. The purge gas shall be introduced less than 5 mm from the base of the water column and shall pass through the water as bubbles with a

diameter less than 3 mm. The purging device shown in Figure 1 meets these criteria.

5.2.1.2 Purging device for solid samples: Designed to accept 5 g of solids plus 5 mL of water. The volume of the gaseous head space between the water and trap shall be less than 25 mL. The purge gas shall be introduced less than 5 mm from the base of the sample and shall pass through the water as bubbles with a diameter less than 3 mm. The purging device shall be capable of operating at ambient temperature (20 to 25°C) and of being controlled at temperatures of 40°C ($\pm 2^\circ\text{C}$) and 80°C ($\pm 5^\circ\text{C}$) while the sample is being purged. The purging device shown in Figure 2 meets these criteria.

5.2.2 Trap: 25 to 30 cm long \times 2.5 mm I.D. minimum, containing the following:

5.2.2.1 Methyl silicone packing: 1cm ($\pm 0.2\text{cm}$), 3% OV-1 on 60/80 mesh Chromosorb W, or equivalent.

5.2.2.2 Porous polymer: 15cm ($\pm 1.0\text{ cm}$), Tenax GC (2,6-diphenylene oxide polymer), 60/80 mesh, chromatographic grade, or equivalent.

5.2.2.3 Silica gel: 8cm ($\pm 1.0\text{ cm}$), Davison Chemical, 35/60 mesh, grade 15, or equivalent. The trap shown in Figure 3 meets these specifications.

5.2.3 Desorber: Shall heat the trap to 175°C ($\pm 5^\circ\text{C}$) in 45 seconds or less. The polymer section of the trap shall not exceed a temperature of 180°C and the remaining sections shall not exceed 220°C during desorb, and no portion of the trap shall exceed 225°C during bakeout. The desorber shown in Figure 3 meets these specifications.

5.2.4 The purge and trap device may be a separate unit, or coupled to a GC as shown in Figures 4 and 5.

5.3 Gas chromatograph: Shall be linearly temperature programmable with initial and final holds, shall contain a glass jet separator as the MS interface, and shall produce results which meet the calibration (Section 7), quality assurance (Section 8), and performance tests (Section 11) of this method.

5.3.1 Column: 2.8 \cdot 0.4 m \times 2 \cdot 0.5 mm I.D. glass, packed with 1% SP-1000 on Carbowax B, 60/80 mesh, or equivalent.

5.4 Mass spectrometer: 70 eV electron impact ionization; shall repetitively scan from 20 to 250 amu every 2 to 3 seconds, and produce a unit resolution (valleys between m/z 174 to 176 less than 10% of the height of the m/z 175 peak), background corrected mass spectrum from 50 ng 4-bromofluorobenzene (BFB) injected into the GC. The BFB spectrum shall meet the mass-intensity criteria in Table 4. All portions of the GC column, transfer lines, and separator which connect the GC column to the ion source shall remain at or above the column temperature during analysis to preclude condensation of less volatile compounds.

5.5 Data system: Shall collect and record MS data, store mass-intensity data in spectral libraries, process GCMS data and generate reports, and shall calculate and record response factors.

Table 4
BFB Mass-Intensity Specifications

m/z	Intensity Required
50	15 to 40% of m/z 95
75	30 to 60% of m/z 95
95	base peak, 100%
96	5 to 9% of m/z 95
173	less than 2% of m/z 174
174	greater than 50% of m/z 95
175	5 to 9% of m/z 174
176	95 to 101% of m/z 174
177	5 to 9% of m/z 176

- 5.5.1** Data acquisition: Mass spectra shall be collected continuously throughout the analysis and stored on a mass storage device.
- 5.5.2** Mass spectral libraries: User-created libraries containing mass spectra obtained from analysis of authentic standards shall be employed to reverse search GCMS runs for the compounds of interest (Section 7.2).
- 5.5.3** Data processing: The data system shall be used to search, locate, identify, and quantify the compounds of interest in each GCMS analysis. Software routines shall be employed to compute retention times and EICP areas. Displays of spectra, mass chromatograms, and library comparisons are required to verify results.
- 5.5.4** Response factors and multipoint calibrations: The data system shall be used to record and maintain lists of response factors (response ratios for isotope dilution) and generate multi-point calibration curves (Section 7). Computations of relative standard deviation (coefficient of variation) are useful for testing calibration linearity. Statistics on initial and ongoing performance shall be maintained (Sections 8 and 11).
- 5.6** Syringes: 5-mL glass hypodermic, with Luer-lok tips.
- 5.7** Micro syringes: 10-, 25-, and 100 μ L.
- 5.8** Syringe valves: 2-way, with Luer ends (Teflon or Kel-F).
- 5.9** Syringe: 5-mL, gas-tight, with shut-off valve.
- 5.10** Bottles: 15-mL, screw-cap with Teflon liner.
- 5.11** Balances.
- 5.11.1** Analytical, capable of weighing 0.1 mg.
- 5.11.2** Top-loading, capable of weighing 10 mg.
- 5.12** Equipment for determining percent moisture.

5.12.1 Oven, capable of being temperature-controlled at 110°C ($\pm 5^\circ\text{C}$).

5.12.2 Dessicator.

5.12.3 Beakers: 50 to 100-mL.

6. REAGENTS AND STANDARDS

6.1 Reagent water: Water in which the compounds of interest and interfering compounds are not detected by this method (Section 11.7). It may be generated by any of the following methods:

6.1.1 Activated carbon: pass tap water through a carbon bed (Calgon Filtrasorb-300, or equivalent).

6.1.2 Water purifier: Pass tap water through a purifier (Millipore Super Q, or equivalent).

6.1.3 Boil and purge: Heat tap water to between 90 and 100°C and bubble contaminant free inert gas through it for approximately 1 hour. While still hot, transfer the water to screw-cap bottles and seal with a Teflon-lined cap.

6.2 Sodium thiosulfate: ACS granular.

6.3 Methanol: Pesticide-quality or equivalent.

6.4 Standard solutions: Purchased as solutions or mixtures with certification to their purity, concentration, and authenticity, or prepared from materials of known purity and composition. If compound purity is 96% or greater, the weight may be used without correction to calculate the concentration of the standard.

6.5 Preparation of stock solutions: Prepare in methanol using liquid or gaseous standards per the steps below. Observe the safety precautions given in Section 4.

6.5.1 Place approximately 9.8 mL of methanol in a 10-mL ground-glass-stoppered volumetric flask. Allow the flask to stand unstoppered for approximately 10 minutes or until all methanol wetted surfaces have dried. In each case, weigh the flask, immediately add the compound, then immediately reweigh to prevent evaporation losses from affecting the measurement.

6.5.1.1 Liquids: Using a 100 μL syringe, permit 2 drops of liquid to fall into the methanol without contacting the neck of the flask. Alternatively, inject a known volume of the compound into the methanol in the flask using a micro-syringe.

6.5.1.2 Gases (chloromethane, bromomethane, chloroethane, vinyl chloride): Fill a valved 5-mL gas-tight syringe with the compound. Lower the needle to approximately 5 mm above the methanol meniscus. Slowly introduce the compound above the surface of the meniscus. The gas will dissolve rapidly in the methanol.

6.5.2 Fill the flask to volume, stopper, then mix by inverting several times. Calculate the concentration in mg/mL ($\mu\text{g}/\mu\text{L}$) from the weight gain (or density if a known volume was injected).

- 6.5.3** Transfer the stock solution to a Teflon-sealed screw-cap bottle. Store, with minimal headspace, in the dark at -10 to -20°C .
- 6.5.4** Prepare fresh standards weekly for the gases and 2-chloroethylvinyl ether. All other standards are replaced after one month, or sooner if comparison with check standards indicate a change in concentration. Quality control check standards that can be used to determine the accuracy of calibration standards are available from the US Environmental Protection Agency, Environmental Monitoring and Support Laboratory, Cincinnati, Ohio.
- 6.6** Labeled compound spiking solution: From stock standard solutions prepared as above, or from mixtures, prepare the spiking solution to contain a concentration such that a 5- to 10- μL spike into each 5-mL sample, blank, or aqueous standard analyzed will result in a concentration of 20 $\mu\text{g}/\text{L}$ of each labeled compound. For the gases and for the water soluble compounds (acrolein, acrylonitrile, acetone, diethyl ether, p-dioxane, and MEK), a concentration of 100 $\mu\text{g}/\text{L}$ may be used. Include the internal standards (Section 7.5) in this solution so that a concentration of 20 $\mu\text{g}/\text{L}$ in each sample, blank, or aqueous standard will be produced.
- 6.7** Secondary standards: Using stock solutions, prepare a secondary standard in methanol to contain each pollutant at a concentration of 500 $\mu\text{g}/\text{mL}$. For the gases and water soluble compounds (Section 6.6), a concentration of 2.5 mg/mL may be used.
- 6.7.1** Aqueous calibration standards: Using a 25- μL syringe, add 20 μL of the secondary standard (Section 6.7) to 50, 100, 200, 500, and 1000 mL of reagent water to produce concentrations of 200, 100, 50, 20, and 10 $\mu\text{g}/\text{L}$, respectively. If the higher concentration standard for the gases and water soluble compounds was chosen (Section 6.6), these compounds will be at concentrations of 1000, 500, 250, 100, and 50 $\mu\text{g}/\text{L}$ in the aqueous calibration standards.
- 6.7.2** Aqueous performance standard: An aqueous standard containing all pollutants, internal standards, labeled compounds, and BFB is prepared daily, and analyzed each shift to demonstrate performance (Section 11). This standard shall contain either 20 or 100 $\mu\text{g}/\text{L}$ of the labeled and pollutant gases and water soluble compounds, 10 $\mu\text{g}/\text{L}$ BFB, and 20 $\mu\text{g}/\text{L}$ of all other pollutants, labeled compounds, and internal standards. It may be the nominal 20 $\mu\text{g}/\text{L}$ aqueous calibration standard (Section 6.7.1).
- 6.7.3** A methanolic standard containing all pollutants and internal standards is prepared to demonstrate recovery of these compounds when syringe injection and purge-and-trap analyses are compared. This standard shall contain either 100 $\mu\text{g}/\text{mL}$ or 500 $\mu\text{g}/\text{mL}$ of the gases and water soluble compounds, and 100 $\mu\text{g}/\text{mL}$ of the remaining pollutants and internal standards (consistent with the amounts in the aqueous performance standard in 6.7.2).
- 6.7.4** Other standards which may be needed are those for test of BFB performance (Section 7.1) and for collection of mass spectra for storage in spectral libraries (Section 7.2).

7. CALIBRATION

Calibration of the GCMS system is performed by purging the compounds of interest and their labeled analogs from reagent water at the temperature to be used for analysis of samples.

- 7.1** Assemble the gas chromatographic apparatus and establish operating conditions given in Table 3. By injecting standards into the GC, demonstrate that the analytical system meets the minimum levels in Table 3 for the compounds for which calibration is to be performed, and the mass-intensity criteria in Table 4 for 50 ng BFB.
- 7.2** Mass spectral libraries: Detection and identification of the compounds of interest are dependent upon the spectra stored in user created libraries.
 - 7.2.1** For the compounds in Table 1 and other compounds for which the GCMS is to be calibrated, obtain a mass spectrum of each pollutant and labeled compound and each internal standard by analyzing an authentic standard either singly or as part of a mixture in which there is no interference between closely eluted components. Examine the spectrum to determine that only a single compound is present. Fragments not attributable to the compound under study indicate the presence of an interfering compound. Adjust the analytical conditions and scan rate (for this test only) to produce an undistorted spectrum at the GC peak maximum. An undistorted spectrum will usually be obtained if five complete spectra are collected across the upper half of the GC peak. Software algorithms designed to "enhance" the spectrum may eliminate distortion, but may also eliminate authentic m/z 's or introduce other distortion.
 - 7.2.2** The authentic reference spectrum is obtained under BFB tuning conditions (Section 7.1 and Table 4) to normalize it to spectra from other instruments.
 - 7.2.3** The spectrum is edited by saving the five most intense mass spectral peaks and all other mass spectral peaks greater than 10% of the base peak. The spectrum may be further edited to remove common interfering masses. If five mass spectral peaks cannot be obtained under the scan conditions given in Section 5.4, the mass spectrometer may be scanned to an m/z lower than 20 to gain additional spectral information. The spectrum obtained is stored for reverse search and for compound confirmation.
 - 7.2.4** For the compounds in Table 2 and other compounds for which the mass spectra, quantitation m/z 's, and retention times are known but the instrument is not to be calibrated, add the retention time and reference compound (Table 3); the response factor and the quantitation m/z (Table 5); and spectrum (Appendix A) to the reverse search library. Edit the spectrum per Section 7.2.3, if necessary.
- 7.3** Assemble the purge-and-trap device. Pack the trap as shown in Figure 3 and condition overnight at 170 to 180°C by backflushing with an inert gas at a flow rate of 20 to 30 mL/min. Condition traps daily for a minimum of 10 minutes prior to use.
 - 7.3.1** Analyze the aqueous performance standard (Section 6.7.2) according to the purge-and-trap procedure in Section 10. Compute the area at the primary m/z (Table 5) for each compound. Compare these areas to those obtained by injecting 1 μ L of the methanolic standard (Section 6.7.3) to determine compound recovery. The recovery shall be greater than 20% for the water soluble compounds (Section 6.6), and 60 to 110% for all other compounds. This recovery is demonstrated initially for each purge-and-trap GCMS system. The test is repeated only if the

purge-and-trap or GCMS systems are modified in any way that might result in a change in recovery.

- 7.3.2** Demonstrate that 100 ng toluene (or toluene-d₈) produces an area at m/z 91 (or 99) approximately one-tenth that required to exceed the linear range of the system. The exact value must be determined by experience for each instrument. It is used to match the calibration range of the instrument to the analytical range and detection limits required.

Table 5. Volatile Organic Compound Characteristic M/Z'S

Compound	labeled Analog	Primary m/z	Reference Compound ²	Response purge temp. Of:	
				20 °C	80 °C
Acetone	d ₆	58/64			
Acrolein	d ₄	56/60			
Acrylonitrile	d ₃	53/56			
Allyl alcohol		57	181	-- ³	0.20
Benzene	d ₆	78/84			
2-Bromo-1-chloropropane 4		77			
Bromochloromethane ⁴		128			
Bromodichloromethane	13 _c	83/86			
Bromoform	13 _c	173/176			
Bromomethane	d ₃	96/99			
Carbon disulfide		76	181	1.93	2.02
Carbon tetrachloride	13 _c	47/48			
2-Chloro-1,3-butadiene		53	182	0.29	0.50
Chloroacetonitrile		75	181	-- ³	1.12
Chlorobenzene	d ₅	112/117			
Chloroethane	d ₅	64/71			
2-Chloroethylvinyl ether	d ₇	106/113			
Chloroform	13 _c	85/86			
Chloromethane	d ₃	50/53			
3-Chloropropene		76	181	0.43	0.63
Crotonaldehyde		70	182	-- ³	0.090
Dibromochloromethane	13 _c	129/130			
1,2-Dibromoethane		107	182	0.86	0.68
Dibromomethane		93	181	1.35	1.91
1,4-Dichlorobutane		55			
trans-1,4-Dichloro-2-butene		75	183	0.093	0.014
1,1-Dichloroethane	d ₃	63/66			

Compound	labeled Analog	Primary m/z ¹	Reference Compound ²	Response purge temp. Of:	
				20 °C	80 °C
1,2-Dichloroethane	d ₄	62/67			
1,1-Dichloroethene	d ₂	61/65			
trans-1,2-Dichloroethene	d ₂	61/65			
1,2-Dichloropropane	d ₆	63/67			
1,3-Dichloropropane		76	182	0.89	0.88
cis-1,3-Dichloropropene		75	182	0.29	0.41
trans-1,3-Dichloropropene	d ₄	75/79			
Diethyl ether	d ₁₀	74/84			
<i>p</i> -Dioxane	d ₈	88/96			
Ethyl cyanide		54	181	(3)	1.26
Ethyl methacrylate		69	183	0.69	0.52
Ethylbenzene	d ₁₀	106/116			
2-Hexanone		58	183	0.076	0.33
Iodomethane		142	181	4.55	2.55
Isobutyl alcohol		74	181	(3)	0.22
Methylene chloride	d ₂	84/88			
Methyl ethyl ketone	d ₈	72/80			
Methyl methacrylate		69	182	0.23	0.79
4-Methyl-2-pentanone		58	183	0.15	0.29
Methacrylonitrile		67	181	0.25	0.79
1,1,1,2-Tetrachloroethane		131	182	0.20	0.25
1,1,2,2-Tetrachloroethane	d ₂	83/84			
Tetrachloroethene	13 _C ²	164/172			
Toluene	d ₈	92/100			
1,1,1-Trichloroethane	d ₃	97/102			
1,1,2-Trichloroethane	13 _C ²	83/84			
Trichloroethene	13 _C ²	95/136			
Trichlorofluoromethane		101	181	2.31	2.19
1,2,3-Trichloropropane		75	183	0.89	0.72
Vinyl acetate		86	182	0.054	0.19
Vinyl chloride	d ₃	62/65			
<i>m</i> -Xylene		106	183	1.69	-
<i>o</i> - and <i>p</i> -Xylene		106	183	3.33	-

¹ Native/labeled

² 181 = bromochloromethane
182 = 2-bromo-1-chloropropane
183 = 1,4-dichlorobutane

³ Not detected at a purge temperature of 20°C

⁴ Internal standard

Note: Because the composition and purity of commercially-supplied isotopically labeled standard's may vary, the primary m/z of the labeled analogs given in this table should be used as guidance. The appropriate m/z of the labeled analogs should be determined prior to use for sample analysis. Deviations from the m/z's listed here must be documented by the laboratory and submitted with the data.

7.4 Calibration by isotope dilution: The isotope dilution approach is used for the purgeable organic compounds when appropriate labeled compounds are available and when interferences do not preclude the analysis. If labeled compounds are not available, or interferences are present, the internal standard method (Section 7.5) is used. A calibration curve encompassing the concentration range of interest is prepared for each compound determined. The relative response (RR) vs. concentration ($\mu\text{g/L}$) is plotted or computed using a linear regression. An example of a calibration curve for toluene using toluene- d_8 is given in Figure 6. Also shown are the $\pm 10\%$ error limits (dotted lines). Relative response is determined according to the procedures described below. A minimum of five data points are required for calibration (Section 7.4.4).

7.4.1 The relative response (RR) of pollutant to labeled compound is determined from isotope ratio values calculated from acquired data. Three isotope ratios are used in this process:

- R_x = the isotope ratio measured in the pure pollutant (Figure 7A).
- R_y = the isotope ratio of pure labeled compound (Figure 7B).
- R_m = the isotope ratio measured in the analytical mixture of the pollutant and labeled compounds (Figure 7C.)

The correct way to calculate RR is:

$$RR = \frac{(R_y - R_m) (R_x + 1)}{(R_m - R_x) (R_y + 1)}$$

If R_m is not between $2R_y$ and $0.5R_x$, the method does not apply and the sample is analyzed by the internal standard method (Section 7.5).

7.4.2 In most cases, the retention times of the pollutant and labeled compound are the same, and isotope ratios (R's) can be calculated from the EICP areas, where:

$$R = \frac{(\text{area at } m_1/z)}{(\text{area at } m_2/z)}$$

If either of the areas is zero, it is assigned a value of one in the calculations; that is, if:
 area of $m_1/z = 50721$,
 area of $m_2/z = 0$,
 then $R = 50721/1 = 50720$

The data from these analyses are reported to three significant figures (see Section 13.6). In order to prevent rounding errors from affecting the values to be

reported, all calculations performed prior to the final determination of concentrations should be carried out using at least four significant figures. Therefore, the calculation of R above is rounded to four significant figures. The m/z's are always selected such that $R_x > R_y$. When there is a difference in retention times (RT) between the pollutant and labeled compounds, special precautions are required to determine the isotope ratios.

R_x , R_y , and R_m are defined as follows:

$$R_x = \frac{[\text{area } m_1/z \text{ (at } RT_1)]}{1}$$

$$R_y = \frac{1}{[\text{area } m_2/z \text{ (at } RT_2)]}$$

$$R_m = \frac{[\text{area } m_1/z \text{ (at } RT_1)]}{[\text{area } m_2/z \text{ (at } RT_2)]}$$

7.4.3 An example of the above calculations can be taken from the data plotted in Figure 7 for toluene and toluene-d₈. For these data:

$$R_x = \frac{168920}{1} = 168900$$

$$R_y = \frac{1}{60960} = 0.00001640$$

$$R_m = \frac{96868}{82508} = 1.174$$

The RR for the above data is then calculated using the equation given in Section 7.4.1. For the example, rounded to four significant figures, RR = 1.174. Not all labeled compounds elute before their pollutant analogs.

7.4.4 To calibrate the analytical system by isotope dilution, analyze a 5-mL aliquot of each of the aqueous calibration standards (Section 6.7.1) spiked with an appropriate constant amount of the labeled compound spiking solution (Section 6.6), using the purge-and-trap procedure in Section 10. Compute the RR at each concentration.

7.4.5 Linearity: If the ratio of relative response to concentration for any compound is constant (less than 20% coefficient of variation) over the five point calibration range, an averaged relative response/concentration ratio may be used for that compound; otherwise, the complete calibration curve for that compound shall be used over the five point calibration range.

7.5 Calibration by internal standard: Used when criteria for isotope dilution (Section 7.4) cannot be met. The method is applied to pollutants having no labeled analog and to the labeled compounds. The internal standards used for volatiles analyses are bromochloromethane, 2-bromo-1-chloropropane, and 1,4-dichlorobutane. Concentrations

of the labeled compounds and pollutants without labeled analogs are computed relative to the nearest eluting internal standard, as shown in Tables 3 and 5.

7.5.1 Response factors: Calibration requires the determination of response factors (RF) which are defined by the following equation:

$$R = \frac{(A_s \times C_{is})}{(a_{is} \times C_s)}$$

Where:

A = is the EICP area at the characteristic m/z for the compound in the daily standard.

A_{is} = is the EICP area at the characteristic m/z for the internal standard.

C_{is} = is the concentration (µg/L) of the internal standard.

C_s = is the concentration of the pollutant in the daily standard.

7.5.2 The response factor is determined at 10, 20, 50, 100, and 200 µg/L for the pollutants (optionally at five times these concentrations for gases and water soluble pollutants; see Section 6.7), in a way analogous to that for calibration by isotope dilution (Section 7.4.4). The RF is plotted against concentration for each compound in the standard (*C_s*) to produce a calibration curve.

7.5.3 Linearity: If the response factor (RF) for any compound is constant (less than 35% coefficient of variation) over the five-point calibration range, an averaged response factor may be used for that compound; otherwise, the complete calibration curve for that compound shall be used over the five-point range.

7.6 Combined calibration: By adding the isotopically labeled compounds and internal standards (Section 6.6) to the aqueous calibration standards (Section 6.7.1), a single set of analyses can be used to produce calibration curves for the isotope dilution and internal standard methods. These curves are verified each shift (Section 11.5) by purging the aqueous performance standard (Section 6.7.2). Recalibration is required only if calibration and ongoing performance (Section 11.5) criteria cannot be met.

7.7 Elevated purge temperature calibration: Samples containing greater than 1% solids are analyzed at a temperature of 40°C (±2°C) (Section 10). For these samples, the analytical system may be calibrated using a purge temperature of 40°C(±2°C) in order to more closely approximate the behavior of the compounds of interest in high solids samples.

8. QUALITY ASSURANCE/QUALITY CONTROL

8.1 Each laboratory that uses this method is required to operate a formal quality assurance program (Reference 8). The minimum requirements of this program consist of an initial demonstration of laboratory capability, analysis of samples spiked with labeled compounds to evaluate and document data quality, and analysis of standards and blanks as tests of continued performance. Laboratory performance is compared to established performance criteria to determine if the results of analyses meet the performance characteristics of the method.

8.1.1 The analyst shall make an initial demonstration of the ability to generate acceptable accuracy and precision with this method. This ability is established as described in Section 8.2.

- 8.1.2** The analyst is permitted to modify this method to improve separations or lower the costs of measurements, provided all performance specifications are met. Each time a modification is made to the method, the analyst is required to repeat the procedure in Section 8.2 to demonstrate method performance.
- 8.1.3** Analyses of blanks are required to demonstrate freedom from contamination and that the compounds of interest and interfering compounds have not been carried over from a previous analysis (Section 3). The procedures and criteria for analysis of a blank are described in Section 8.5.
- 8.1.4** The laboratory shall spike all samples with labeled compounds to monitor method performance. This test is described in Section 8.3. When results of these spikes indicate atypical method performance for samples, the samples are diluted to bring method performance within acceptable limits (Section 14.2).
- 8.1.5** The laboratory shall, on an ongoing basis, demonstrate through the analysis of the aqueous performance standard (Section 6.7.2) that the analysis system is in control. This procedure is described in Sections 11.1 and 11.5.
- 8.1.6** The laboratory shall maintain records to define the quality of data that is generated. Development of accuracy statements is described in Sections 8.4 and 11.5.2.

Table 6. Acceptance Criteria for Performance Tests

		Acceptance criteria at 20 µg/L or as noted				
EGD	No. ¹	Compound	Labeled and native compound initial precision and accuracy (Sect. 8.2.3)		Labeled compound recovery (Sect. 8.3 and 14.2)	Labeled and native compound ongoing accuracy (Sect. 11.5)
			s (µg/L)	X (µg/L)	P (%)	R (µg/L)
	516	acetone*	51.0	77 - 153	35 - 165	55 - 145
	002	acrolein*	72.0	32 - 168	37 - 163	7 - 190
	003	acrylonitrile*	16.0	70 - 132	ns - 204	58 - 144
	004	benzene	9.0	13 - 28	ns - 196	4 - 33
	048	bromodichloromethane	8.2	7 - 32	ns - 199	4 - 34
	047	bromoform	7.0	7 - 35	ns - 214	6 - 36
	046	bromomethane	25.0	d - 54	ns - 414	d - 61
	006	carbon tetrachloride	6.9	16 - 25	42 - 165	12 - 30
	007	chlorobenzene	8.2	14 - 30	ns - 205	4 - 35
	016	chloroethane	15.0	d - 47	ns - 308	d - 51
	019	2-chloroethylvinyl ether	36.0	d - 70	ns - 554	d - 79
	023	chloroform	7.9	12 - 26	18 - 172	8 - 30
	045	chloromethane	26.0	d - 56	ns - 410	d - 64
	051	dibromochloromethane	7.9	11 - 29	16 - 185	8 - 32
	013	1,1-dichloroethane	6.7	11 - 31	23 - 191	9 - 33
	010	1,2-dichloroethane	7.7	12 - 30	12 - 192	8 - 33
	029	1,1-dichloroethene	12.0	d - 50	ns - 315	d - 52
	030	trans-1,2-dichloroethene	7.4	11 - 32	15 - 195	8 - 34
	032	1,2-dichloropropane	19.0	d - 47	ns - 343	d - 51
	033	trans-1,3-dichloropropene	15.0	d - 40	ns - 284	d - 44
	515	diethyl ether*	44.0	75 - 146	44 - 156	55 - 145

		Acceptance criteria at 20 µg/L or as noted			
EGD		Labeled and native compound initial precision and accuracy (Sect. 8.2.3)	Labeled compound recovery (Sect. 8.3 and 14.2)	Labeled and native compound ongoing accuracy (Sect. 11.5)	
No. ¹	Compound	s (µg/L)	X (µg/L)	P (%)	R (µg/L)
527	p-dioxane*	7.2	13 - 27	ns - 239	11 - 29
038	ethylbenzene	9.6	16 - 29	ns - 203	5 - 35
044	methylene chloride	9.7	d - 50	ns - 316	d - 50
514	methyl ethyl ketone*	57.0	66 - 159	36 - 164	42 - 158
015	1,1,2,2-tetrachloroethane	9.6	11 - 30	5 - 199	7 - 34
085	tetrachloroethane	6.6	15 - 29	31 - 181	11 - 32
086	toluene	6.3	15 - 29	4 - 193	6 - 33
011	1,1,1-trichloroethane	5.9	11 - 33	12 - 200	8 - 35
014	1,1,2-trichloroethane	7.1	12 - 30	21 - 184	9 - 32
087	trichloroethene	8.9	17 - 30	35 - 196	12 - 34
088	vinyl chloride	28.0	d - 59	ns - 452	d - 65

* acceptance criteria at 100 µg/L

d = detected; result must be greater than zero.

ns = no specification; limit would be below detection limit.

¹ Reference numbers beginning with 0, 1, or 5 indicate a pollutant quantified by the internal standard method; reference numbers beginning with 2 or 6 indicate a labeled compound quantified by the internal Standard method; reference numbers beginning with 3 or 7 indicate a pollutant quantified by isotope dilution.

- 8.2** Initial precision and accuracy: To establish the ability to generate acceptable precision and accuracy, the analyst shall perform the following operations for compounds to be calibrated:
- 8.2.1** Analyze two sets of four 5-mL aliquots (8 aliquots total) of the aqueous performance standard (Section 6.7.2) according to the method beginning in Section 10.
 - 8.2.2** Using results of the first set of four analyses in Section 8.2.1, compute the average recovery (X) in $\mu\text{g/L}$ and the standard deviation of the recovery (s) in $\mu\text{g/L}$ for each compound, by isotope dilution for pollutants with a labeled analog, and by internal standard for labeled compounds and pollutants with no labeled analog.
 - 8.2.3** For each compound, compare s and X with the corresponding limits for initial precision and accuracy found in Table 6. If s and X for all compounds meet the acceptance criteria, system performance is acceptable and analysis of blanks and samples may begin. If, however, any individual s exceeds the precision limit or any individual X falls outside the range for accuracy, system performance is unacceptable for that compound.

NOTE: *The large number of compounds in Table 6 present a substantial probability that one or more will fail one of the acceptance criteria when all compounds are analyzed. To determine if the analytical system is out of control, or if the failure can be attributed to probability, proceed as follows:*

- 8.2.4** Using the results of the second set of four analyses, compute s and X for only those compounds which failed the test of the first set of four analyses (Section 8.2.3). If these compounds now pass, system performance is acceptable for all compounds and analysis of blanks and samples may begin. If, however, any of the same compounds fail again, the analysis system is not performing properly for the compound (s) in question. In this event, correct the problem and repeat the entire test (Section 8.2.1).
- 8.3** The laboratory shall spike all samples with labeled compounds to assess method performance on the sample matrix.
- 8.3.1** Spike and analyze each sample according to the method beginning in Section 10.
 - 8.3.2** Compute the percent recovery (P) of the labeled compounds using the internal standard method (Section 7.5).
 - 8.3.3** Compare the percent recovery for each compound with the corresponding labeled compound recovery limit in Table 6. If the recovery of any compound falls outside its warning limit, method performance is unacceptable for that compound in that sample. Therefore, the sample matrix is complex and the sample is to be diluted and reanalyzed, per Section 14.2.
- 8.4** As part of the QA program for the laboratory, method accuracy for wastewater samples shall be assessed and records shall be maintained. After the analysis of five wastewater samples for which the labeled compounds pass the tests in Section 8.3.3, compute the

average percent recovery (P) and the standard deviation of the percent recovery (sp) for the labeled compounds only. Express the accuracy assessment as a percent recovery interval from $P - 2sp$ to $P + 2sp$. For example, if $P = 90\%$ and $sp = 10\%$, the accuracy interval is expressed as 70 to 110%. Update the accuracy assessment for each compound on a regular basis (e.g., after each 5 to 10 new accuracy measurements).

8.5 Blanks: Reagent water blanks are analyzed to demonstrate freedom from carry-over (Section 3) and contamination.

8.5.1 The level at which the purge and trap system will carry greater than $5 \mu\text{g/L}$ of a pollutant of interest (Tables 1 and 2) into a succeeding blank shall be determined by analyzing successively larger concentrations of these compounds. When a sample contains this concentration or more, a blank shall be analyzed immediately following this sample to demonstrate no carry-over at the $5 \mu\text{g/L}$ level.

8.5.2 With each sample lot (samples analyzed on the same 8-hour shift), a blank shall be analyzed immediately after analysis of the aqueous performance standard (Section 11.1) to demonstrate freedom from contamination. If any of the compounds of interest (Tables 1 and 2) or any potentially interfering compound is found in a blank at greater than $10 \mu\text{g/L}$ (assuming a response factor of 1 relative to the nearest eluted internal standard for compounds not listed in Tables 1 and 2), analysis of samples is halted until the source of contamination is eliminated and a blank shows no evidence of contamination at this level.

8.6 The specifications contained in this method can be met if the apparatus used is calibrated properly, then maintained in a calibrated state. The standards used for calibration (Section 7), calibration verification (Section 11.5) and for initial (Section 8.2) and ongoing (Section 11.5) precision and accuracy should be identical, so that the most precise results will be obtained. The GCMS instrument in particular will provide the most reproducible results if dedicated to the settings and conditions required for the analyses of volatiles by this method.

8.7 Depending on specific program requirements, field replicates may be collected to determine the precision of the sampling technique, and spiked samples may be required to determine the accuracy of the analysis when the internal method is used.

9. SAMPLE COLLECTION, PRESERVATION, AND HANDLING

9.1 Grab samples are collected in glass containers having a total volume greater than 20 mL. For aqueous samples which pour freely, fill sample bottles so that no air bubbles pass through the sample as the bottle is filled and seal each bottle so that no air bubbles are entrapped. Maintain the hermetic seal on the sample bottle until time of analysis.

9.2 Samples are maintained at 0 to 4°C from the time of collection until analysis. If an aqueous sample contains residual chlorine, add sodium thiosulfate preservative (10 mg/40 mL) to the empty sample bottles just prior to shipment to the sample site. EPA Methods 330.4 and 330.5 may be used for measurement of residual chlorine (Reference 9). If preservative has been added, shake the bottle vigorously for one minute immediately after filling.

9.3 For aqueous samples, experimental evidence indicates that some aromatic compounds, notably benzene, toluene, and ethyl benzene are susceptible to rapid biological

degradation under certain environmental conditions. Refrigeration alone may not be adequate to preserve these compounds in wastewaters for more than seven days. For this reason, a separate sample should be collected, acidified, and analyzed when these aromatics are to be determined. Collect about 500 mL of sample in a clean container. Adjust the pH of the sample to about 2 by adding HCl (1+1) while stirring. Check pH with narrow range (1.4 to 2.8) pH paper. Fill a sample container as described in Section 9.1. If residual chlorine is present, add sodium thiosulfate to a separate sample container and fill as in Section 9.1.

9.4 All samples shall be analyzed within 14 days of collection.

10. PURGE, TRAP, AND GCMS ANALYSIS

Samples containing less than one percent solids are analyzed directly as aqueous samples (Section 10.4). Samples containing one percent solids or greater are analyzed as solid samples utilizing one of two methods, depending on the levels of pollutants in the sample. Samples containing one percent solids or greater and low to moderate levels of pollutants are analyzed by purging a known weight of sample added to 5 mL of reagent water (Section 10.5). Samples containing 1% solids or greater and high levels of pollutants are extracted with methanol, and an aliquot of the methanol extract is added to reagent water and purged (Section 10.6).

10.1 Determination of percent solids.

10.1.1 Weigh 5 to 10 g of sample into a tared beaker.

10.1.2 Dry overnight (12 hours minimum) at 110°C (±5°C), and cool in a dessicator.

10.1.3 Determine percent solids as follows:

$$\% \text{ solids} = \frac{\text{weight of sample dry}}{\text{weight of sample wet}} \times 100$$

10.2 Remove standards and samples from cold storage and bring to 20 to 25°C.

10.3 Adjust the purge gas flow rate to 40 (±4mL/min).

10.4 Samples containing less than 1% solids.

10.4.1 Mix the sample by shaking vigorously. Remove the plunger from a 5-mL syringe and attach a closed syringe valve. Open the sample bottle and carefully pour the sample into the syringe barrel until it overflows. Replace the plunger and compress the sample. Open the syringe valve and vent any residual air while adjusting the sample volume to 5 mL (±0.1 mL). Because this process of taking an aliquot destroys the validity of the sample for future analysis, fill a second syringe at this time to protect against possible loss of data.

10.4.2 Add an appropriate amount of the labeled compound spiking solution (Section 6.6) through the valve bore, then close the valve.

10.4.3 Attach the syringe valve assembly to the syringe valve on the purging device. Open both syringe valves and inject the sample into the purging chamber. Purge the sample per Section 10.7.

- 10.5** Samples containing 1% solids or greater and low to moderate levels of pollutants.
- 10.5.1** Mix the sample thoroughly using a clean spatula.
 - 10.5.2** Weigh 5 g (± 1 g) of sample into a purging vessel (Figure 2). Record the weight to three significant figures.
 - 10.5.3** Add 5 mL (± 0.1 mL) of reagent water to the vessel.
 - 10.5.4** Using a metal spatula, break up any lumps of sample to disperse the sample in the water.
 - 10.5.5** Add an appropriate amount of the labeled compound spiking solution (Section 6.6) to the sample in the purge vessel. Place a cap on the purging vessel and shake vigorously to further disperse the sample. Attach the purge vessel to the purging device, and purge the sample per Section 10.7.
- 10.6** Samples containing 1% solids or greater and high levels of pollutants, or samples requiring dilution by a factor of more than 100 (see Section 13.4).
- 10.6.1** Mix the sample thoroughly using a clean spatula.
 - 10.6.2** Weigh 5g (± 1 g) of sample into a calibrated 15- to 25-mL centrifuge tube. Record the weight of the sample to three significant figures.
 - 10.6.3** Add 10 mL of methanol to the centrifuge tube. Cap the tube and shake it vigorously for 15 to 20 seconds to disperse the sample in the methanol. Allow the sample to settle in the tube. If necessary, centrifuge the sample to settle suspended particles.
 - 10.6.4** Remove approximately 0.1% of the volume of the supernatant methanol using a 15- to 25- μ L syringe. This volume will be in the range of 10 to 15 μ L.
 - 10.6.5** Add this volume of the methanol extract to 5 mL reagent water in a 5 mL syringe, and analyze per Section 10.4.1.
 - 10.6.6** For further dilutions, dilute 1 mL of the supernatant methanol (Section 10.6.4) to 10 mL, 100 mL, 1000 mL, etc., in reagent water. Remove a volume of this methanol extract/reagent water mixture equivalent to the volume in Section 10.6.4, add it to 5 mL reagent water in a 5 mL syringe, and analyze per Section 10.4.1.
- 10.7** Purge the sample for 11 minutes (± 0.1 minute) at 20 to 25°C for samples containing less than 1% solids. Purge samples containing one percent solids or greater at 40° ($\pm 2^\circ$). If the compounds in Table 2 that do not purge at 20 to 40°C are to be determined, a purge temperature of 80°C ($\pm 5^\circ$ C) is used.
- 10.8** After the 11 minute purge time, attach the trap to the chromatograph and set the purge- and- trap apparatus to the desorb mode (Figure 5). Desorb the trapped compounds into the GC column by heating the trap to between 170 and 180°C while backflushing with carrier gas at 20 to 60 mL/min for 4 minutes. Start MS data acquisition upon start of the desorb cycle, and start the GC column temperature program 3 minutes later. Table 3 summarizes the recommended operating conditions for the gas chromatograph. Included in this table are retention times and minimum levels that can be achieved under these conditions. An example of the separations achieved by the column listed is shown in Figure 9. Other columns may be used provided the requirements in Section 8 are met.

If the priority pollutant gases produce GC peaks so broad that the precision and recovery specifications (Section 8.2) cannot be met, the column may be cooled to ambient or subambient temperatures to sharpen these peaks.

- 10.9** After desorbing the sample for four minutes, recondition the trap by purging with purge gas while maintaining the trap temperature at between 170 and 180°C. After approximately 7 minutes, turn off the trap heater to stop the gas flow through the trap. When cool, the trap is ready for the next sample.
- 10.10** While analysis of the desorbed compounds proceeds, remove and clean the purge device. Rinse with tap water, clean with detergent and water, rinse with tap and distilled water, and dry for a minimum of 1 hour in an oven at a temperature greater than 150°C.

11. SYSTEM PERFORMANCE

- 11.1** At the beginning of each 8 hour shift during which analyses are performed, system calibration and performance shall be verified for the pollutants and labeled compounds (Table 1). For these tests, analysis of the aqueous performance standard (Section 6.7.2) shall be used to verify all performance criteria. Adjustment and/or recalibration (per Section 7) shall be performed until all performance criteria are met. Only after all performance criteria are met may blanks and samples be analyzed.
- 11.2** BFB spectrum validity: The criteria in Table 4 shall be met.
- 11.3** Retention times: The absolute retention times of the internal standards shall be as follows: bromochloromethane: 653 to 782 seconds; 2-bromo-1-chloropropane: 1270 to 1369 seconds; 1,4-dichlorobutane: 1510 to 1605 seconds. The relative retention times of all pollutants and labeled compounds shall fall within the limits given in Table 3.
- 11.4** GC resolution: The valley height between toluene and toluene-d₈ (at m/z 91 and 99 plotted on the same graph) shall be less than 10% of the taller of the two peaks.
- 11.5** Calibration verification and ongoing precision and accuracy: Compute the concentration of each pollutant (Table 1) by isotope dilution (Section 7.4) for those compounds which have labeled analogs. Compute the concentration of each pollutant which has no labeled analog by the internal standard method (Section 7.5). Compute the concentrations of the labeled compounds themselves by the internal standard method. These concentrations are computed based on the calibration data determined in Section 7.
- 11.5.1** For each pollutant and labeled compound, compare the concentration with the corresponding limit for ongoing accuracy in Table 6. If all compounds meet the acceptance criteria, system performance is acceptable and analysis of blanks and samples may continue. If any individual value falls outside the range given, system performance is unacceptable for that compound.

NOTE: *The large number of compounds in Table 6 present a substantial probability that one or more will fail the acceptance criteria when all compounds are analyzed. To determine if the analytical system is out of control, or if the failure may be attributed to probability, proceed as follows:*

- 11.5.1.1 Analyze a second aliquot of the aqueous performance standard (Section 6.7.2).

11.5.1.2 Compute the concentration for only those compounds which failed the first test (Section 11.5.1). If these compounds now pass, system performance is acceptable for all compounds, and analyses of blanks and samples may proceed. If, however, any of the compounds fail again, the measurement system is not performing properly for these compounds. In this event, locate and correct the problem or recalibrate the system (Section 7), and repeat the entire test (Section 11.1) for all compounds.

11.5.2 Add results which pass the specification in Section 11.5.1.2 to initial (Section 8.2) and previous on-going data. Update QC charts to form a graphic representation of laboratory performance (Figure 8). Develop a statement of accuracy for each pollutant and labeled compound by calculating the average percent recovery (R) and the standard deviation of percent recovery (sr). Express the accuracy as a recovery interval from $R - 2sr$ to $R + 2sr$. For example, if $R = 95\%$ and $sr = 5\%$, the accuracy is 85 to 105%.

12. QUALITATIVE DETERMINATION

Identification is accomplished by comparison of data from analysis of a sample or blank with data stored in the mass-spectral libraries. For compounds for which the relative retention times and mass spectra are known, identification is confirmed per Sections 12.1 and 12.2. For unidentified GC peaks, the spectrum is compared to spectra in the EPA/NIH mass spectral file per Section 12.3.

12.1 Labeled compounds and pollutants having no labeled analog (Tables 1 and 2):

12.1.1 The signals for all characteristic m/z 's stored in the spectral library (Section 7.2.3) shall be present and shall maximize within the same two consecutive scans.

12.1.2 Either (1) the background corrected EICP areas or (2) the corrected relative intensities of the mass spectral peaks at the GC peak maximum shall agree within a factor of 2 (0.5 to 2 times) for all masses stored in the library.

12.1.3 In order for the compounds for which the system has been calibrated (Table 1) to be identified, their relative retention times shall be within the retention-time windows specified in Table 3.

12.1.4 The system has not been calibrated for the compounds listed in Table 2; however, the relative retention times and mass spectra of these compounds are known. Therefore, for a compound in Table 2 to be identified, its relative retention time must fall within a retention-time window of ± 60 seconds or ± 20 scans (whichever is greater) of the nominal retention time of the compound specified in Table 3.

12.2 Pollutants having a labeled analog (Table 1):

12.2.1 The signals for all characteristic m/z 's stored in the spectral library (Section 7.2.3) shall be present and shall maximize within the same two consecutive scans.

12.2.2 Either (1) the background corrected EICP areas or (2) the corrected relative intensities of the mass spectral peaks at the GC peak maximum shall agree within a factor of two for all masses stored in the spectral library.

- 12.2.3** The relative retention time between the pollutant and its labeled analog shall be within the windows specified in Table 3.
- 12.3** Unidentified GC peaks.
- 12.3.1** The signals for m/z's specific to a GC peak shall all maximize within the same two consecutive scans.
- 12.3.2** Either (1) the background corrected EICP areas or (2) the corrected relative intensities of the mass spectral peaks at the GC peak maximum shall agree within a factor of 2 with the masses stored in the EPA/NIH mass-spectral file.
- 12.4** The m/z's present in the sample mass spectrum that are not present in the reference mass spectrum shall be accounted for by contaminant or background ions. If the sample mass spectrum is contaminated, or if identification is ambiguous, an experienced spectrometrist (Section 1.4) is to determine the presence or absence of the compound.

13. QUANTITATIVE DETERMINATION

- 13.1** Isotope dilution: Because the pollutant and its labeled analog exhibit the same effects upon purging, desorption, and gas chromatography, correction for recovery of the pollutant can be made by adding a known amount of a labeled compound to every sample prior to purging. Relative response (RR) values for sample mixtures are used in conjunction with the calibration curves described in Section 7.4 to determine concentrations directly, so long as labeled compound spiking levels are constant. For the toluene example given in Figure 7 (Section 7.4.3), RR would be equal to 1.174. For this RR value, the toluene calibration curve given in Figure 6 indicates a concentration of 31.8 µg/L.
- 13.2** Internal standard: For the compounds for which the system was calibrated (Table 1) according to Section 7.5, use the response factor determined during the calibration to calculate the concentration from the following equation.

$$\text{Concentration} = \frac{(A_s \times C_{is})}{(A_{is} \times RF)}$$

where the terms are as defined in Section 7.5.1. For the compounds for which the system was not calibrated (Table 2), use the response factors in Table 5 to calculate the concentration.

- 13.3** The concentration of the pollutant in the solid phase of the sample is computed using the concentration of the pollutant detected in the aqueous solution, as follows:

$$\text{Concentration in solid } (\mu\text{g/kg}) = \frac{0.005 \text{ L} \times \text{aqueous conc } (\mu\text{g/L})}{0.01 \times \text{percent solids}(g)}$$

where

"percent solids" is from Section 10.1.3

- 13.4** Dilution of samples: If the EICP area at the quantitation m/z exceeds the calibration range of the system, samples are diluted by successive factors of 10 until the area is within the calibration range.

- 13.4.1** For aqueous samples, bring 0.50 mL, 0.050 mL, 0.0050 mL, etc., to 5-mL volume with reagent water and analyze per Section 10.4.
- 13.4.2** For samples containing high solids, substitute 0.50 or 0.050 g in Section 10.5.2 to achieve a factor of 10 or 100 dilution, respectively.
- 13.4.3** If dilution of high solids samples by greater than a factor of 100 is required, then extract the sample with methanol, as described in Section 10.6.
- 13.5** Dilution of samples containing high concentrations of compounds not in Table 1: When the EICP area of the quantitation m/z of a compound to be identified per Section 12.3 exceeds the linear range of the GCMS system, or when any peak in the mass spectrum is saturated, dilute the sample per Sections 13.4.1 through 13.4.3.
- 13.6** Report results for all pollutants, labeled compounds, and tentatively identified compounds found in all standards, blanks, and samples to three significant figures. For samples containing less than 1% solids, the units are $\mu\text{g/L}$; and for undiluted samples containing 1% solids or greater, units are $\mu\text{g/kg}$.
- 13.6.1** Results for samples which have been diluted are reported at the least dilute level at which the area at the quantitation m/z is within the calibration range (Section 13.4), or at which no m/z in the spectrum is saturated (Section 13.5). For compounds having a labeled analog, results are reported at the least dilute level at which the area at the quantitation m/z is within the calibration range (Section 13.4) and the labeled compound recovery is within the normal range for the method (Section 14.2).

14. ANALYSIS OF COMPLEX SAMPLES

- 14.1** Some samples may contain high levels ($>1000 \mu\text{g/kg}$) of the compounds of interest and of interfering compounds. Some samples will foam excessively when purged. Others will overload the trap or the GC column.
- 14.2** When the recovery of any labeled compound is outside the range given in Table 6, dilute 0.5 mL of samples containing less than 1% solids, or 0.5 g of samples containing 1% solids or greater, with 4.5 mL of reagent water and analyze this diluted sample. If the recovery remains outside of the range for this diluted sample, the aqueous performance standard shall be analyzed (Section 11) and calibration verified (Section 11.5). If the recovery for the labeled compound in the aqueous performance standard is outside the range given in Table 6, the analytical system is out of control. In this case, the instrument shall be repaired, the performance specifications in Section 11 shall be met, and the analysis of the undiluted sample shall be repeated. If the recovery for the aqueous performance standard is within the range given in Table 6, then the method does not apply to the sample being analyzed, and the result may not be reported for regulatory compliance purposes.
- 14.3** When a high level of the pollutant is present, reverse search computer programs may misinterpret the spectrum of chromatographically unresolved pollutant and labeled compound pairs with overlapping spectra. Examine each chromatogram for peaks greater than the height of the internal standard peaks. These peaks can obscure the compounds of interest.

15. METHOD PERFORMANCE

- 15.1 The specifications for this method were taken from the interlaboratory validation of EPA Method 624 (Reference 10). Method 1624 has been shown to yield slightly better performance on treated effluents than method 624. Results of initial tests of this method at a purge temperature of 80°C can be found in Reference 11 and results of initial tests of this method on municipal sludge can be found in Reference 12.
- 15.2 A chromatogram of the 20 µg/L aqueous performance standards (Sections 6.7.2 and 11.1) is shown in Figure 9.

Reference

1. "Performance Tests for the Evaluation of Computerized Gas Chromatography/Mass Spectrometry Equipment and Laboratories," USEPA, EMSL Cincinnati, OH 45268, EPA-600/4-80-025 (April 1980).
2. Bellar, T. A. and Lichtenberg, J. J., "Journal American Water Works Association," 66, 739 (1974).
3. Bellar, T. A. and Lichtenberg, J. J., "Semi-Automated Headspace Analysis of Drinking Waters and Industrial Waters for Purgeable Volatile Organic Compounds," in *Measurement of Organic Pollutants in Water and Wastewater*, C. E. VanHall, ed., American Society for Testing Materials, Philadelphia, PA, Special Technical Publication 686, (1978).
4. National Standard Reference Data System, "Mass Spectral Tape Format," U.S. National Bureau of Standards (1979 and later attachments).
5. "Working with Carcinogens," DHEW, PHS, NIOSH, Publication 77-206 (1977).
6. "OSHA Safety and Health Standards, General Industry," 29 CFR 1910, OSHA 2206, (1976).
7. "Safety in Academic Chemistry Laboratories," American Chemical Society Publication, Committee on Chemical Safety (1979).
8. "Handbook of Analytical Quality Control in Water and Wastewater Laboratories," USEPA, EMSL Cincinnati, OH 45268, EPA-4-79-019 (March 1979).
9. "Methods 330.4 and 330.5 for Total Residual Chlorine," USEPA, EMSL Cincinnati, OH 45268, EPA-4-79-020 (March 1979).
10. "Method 624--Purgeables", 40 CFR Part 136 (49 FR 43234), 26 October 1984.
11. "Narrative for SAS 106: Development of an Isotope Dilution GC/MS Method for Hot Purge and-Trap Volatiles Analysis," S-CUBED Division of Maxwell Laboratories, Inc., Prepared for W. A. Telliard, Industrial Technology Division (WH-552), USEPA, 401 M St. SW, Washington DC 20460 (July 1986).
12. Colby, Bruce N. and Ryan, Philip W., "Initial Evaluation of Methods 1634 and 1635 for the Analysis of Municipal Wastewater Treatment Sludges by Isotope Dilution GCMS," Pacific Analytical Inc., Prepared for W. A. Telliard, Industrial Technology Division (WH-552), USEPA, 401 M St. SW, Washington DC 20460 (July 1986).

Appendix A Mass Spectra in the Form of Mass/Intensity Lists

532 allyl alcohol											
<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>
42	30	43	39	44	232	45	12	53	13	55	59
56	58	57	1000	58	300	61	15				
533 carbon disulfide											
<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>
44	282	46	10	64	14	76	1000	77	27	78	82
534 2-chloro-1,3-butadiene (chloroprene)											
<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>
48	21	49	91	50	223	51	246	52	241	53	1000
54	41	61	30	62	54	63	11	64	16	73	21
87	12	88	452	89	22	90	137				
535 chloroacetonitrile											
<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>
47	135	48	1000	49	88	50	294	51	12	73	22
74	43	75	884	76	39	77	278				
536 3-chloropropene											
<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>
35	39	36	40	40	44	42	206	47	40	58	35
49	176	51	64	52	31	61	29	73	22	75	138
76	1000	77	74	78	324						
537 crotonaldehyde											
<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>
35	26	40	28	42	339	43	48	44	335	49	27
50	40	51	20	52	21	53	31	55	55	68	24
69	511	70	1000	71	43						

Appendix A Mass Spectra in the Form of Mass/Intensity Lists (continued)

538 1,2-dibromoethane (EDB)											
<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>
79	50	80	13	31	51	82	15	93	54	95	42
105	32	106	29	107	1000	108	38	109	922	110	19
186	13	188	27	190	13						
539 dibromomethane											
<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>
43	99	44	101	45	30	79	184	80	35	81	175
91	142	92	61	93	1000	94	64	95	875	160	18
172	375	173	14	174	719	175	12	176	342		
540 trans-1,4-dichloro-2-butene											
<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>
49	166	50	171	51	289	52	85	53	878	54	273
62	286	64	91	75	1000	77	323	88	246	89	415
90	93	91	129	124	138	126	86	128	12		
541 1,3-dichloropropane											
<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>
40	15	42	44	47	19	48	20	49	193	51	55
61	18	62	22	63	131	65	38	75	47	76	1000
77	46	78	310	79	12						
542 cis-1,3-dichloropropene											
<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>
37	262	38	269	39	998	49	596	51	189	75	1000
77	328	110	254	112	161						
543 ethyl cyanide											
<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>
44	115	50	34	51	166	52	190	53	127	54	1000
55	193										
544 ethyl methacrylate											
<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>	<u>m/z</u>	<u>int.</u>
42	127	43	48	45	155	55	32	58	39	68	60
69	1000	70	83	71	25	85	14	86	169	87	21
96	17	99	93	113	11	114	119				

Appendix A Mass Spectra in the Form of Mass/Intensity Lists (continued)

545 2-hexanone (methyl butyl ketone)											
m/z	int.	m/z	int.	m/z	int.	m/z	int.	m/z	int.	m/z	int.
42	61	43	1000	44	24	55	12	57	130	58	382
59	21	71	36	85	37	100	56				
546 iodomethane											
m/z	int.	m/z	int.	m/z	int.	m/z	int.	m/z	int.	m/z	int.
44	57	127	328	128	17	139	39	140	34	141	120
142	1000	143	12								
547 isobutyl alcohol											
m/z	int.	m/z	int.	m/z	int.	m/z	int.	m/z	int.	m/z	int.
34	21	35	13	36	13	37	11	39	10	42	575
43	1000	44	42	45	21	55	40	56	37	57	21
59	25	73	12	74	63						
548 methacrylonitrile											
m/z	int.	m/z	int.	m/z	int.	m/z	int.	m/z	int.	m/z	int.
38	24	39	21	41	26	42	100	49	19	50	60
51	214	52	446	53	19	62	24	63	59	64	136
65	55	66	400	67	1000	68	51				
549 methyl methacrylate											
m/z	int.	m/z	int.	m/z	int.	m/z	int.	m/z	int.	m/z	int.
42	127	43	52	45	48	53	30	55	100	56	49
59	124	68	28	69	1000	70	51	82	26	85	45
98	20	99	89	100	442	101	22				
550 4-methyl-2-pentanone (methyl isobutyl ketone; MIBK)											
m/z	int.	m/z	int.	m/z	int.	m/z	int.	m/z	int.	m/z	int.
42	69	43	1000	44	54	53	11	55	15	56	13
57	205	58	346	59	20	67	12	69	10	85	96
100	94										
551 1,1,1,2-tetrachloroethane											
m/z	int.	m/z	int.	m/z	int.	m/z	int.	m/z	int.	m/z	int.
47	144	49	163	60	303	61	330	62	98	82	45
84	31	95	416	96	152	97	270	98	84	117	804
121	236	131	1000	133	955	135	301				

Appendix A Mass Spectra in the Form of Mass/Intensity Lists (continued)

552 trichlorofluoromethane											
m/z	int.	m/z	int.	m/z	int.	m/z	int.	m/z	int.	m/z	int.
44	95	47	153	49	43	51	21	52	14	66	162
68	53	82	40	84	28	101	1000	102	10	103	671
105	102	117	16	119	14						
553 1,2,3-trichloropropane											
m/z	int.	m/z	int.	m/z	int.	m/z	int.	m/z	int.	m/z	int.
49	285	51	87	61	300	62	107	63	98	75	1000
76	38	77	302	83	23	96	29	97	166	98	20
99	103	110	265	111	28	112	164	114	25		
554 vinyl acetate											
m/z	int.	m/z	int.	m/z	int.	m/z	int.	m/z	int.	m/z	int.
36	5	42	103	43	1000	44	70	45	8	86	57
951 m-xylene											
m/z	int.	m/z	int.	m/z	int.	m/z	int.	m/z	int.	m/z	int.
65	62	77	124	91	1000	105	245	106	580		
951 0- + p-xylene											
m/z	int.	m/z	int.	m/z	int.	m/z	int.	m/z	int.	m/z	int.
51	88	77	131	91	1000	105	229	106	515		




FIGURE 1 Purging Device for Waters




FIGURE 2 Purging Device for Soils or Waters




FIGURE 3 Trap Construction and Packings




FIGURE 4 Schematic of Purge and Trap Device--Purge Mode




FIGURE 5 Schematic of Purge and Trap Device--Desorb Mode




FIGURE 6 Relative Response Calibration Curve for Toluene. The Dotted Lines Enclose a +/- 10 Percent Error Window




FIGURE 7 Extracted Ion Current Profiles for (A) Toluene, (B) Toluene-d₈, and (C) a Mixture of Toluene and Toluene-d₈




FIGURE 8 Quality Control Charts Showing Area (top graph) and Relative Response of Toluene to Toluene-d₈ (lower graph) Plotted as Function of Time or Analysis Number

MASS CHROMATOGRAM DATA: UOAIID1945 #1 SCANS 1 TO 1200
09/01/84 23:05:00 CALI: UOAIID1945 #1
SAMPLE: UO, S, OPR, 00020, 00, U, NA: NA, NA\$
CONDS.: 16248, 3, 0M, 2MM, 3@45, 45-24008, 15@240, 20ML/MINS
RANGE: G 1, 1200 LABEL: N 0, 4.0 QUAN: A 0, 1.0 J 0 BASE: U 20, 3




FIGURE 9 Chromatogram of Aqueous Performance Standard