



We are in the process of ensuring this document is accessible to all audiences. If you need assistance accessing this document, or any material on the EPA East Palestine, Ohio emergency response web pages, please contact the Region 5 Public Information Officer on-call at: R5\_EastPalestine@epa.gov

September 21, 2023

Mr. Josh Peters  
On-Scene Coordinator  
U.S. Environmental Protection Agency, Region 5  
Superfund and Emergency Management Division  
2565 Plymouth Road  
Ann Arbor, MI 48105

**Subject: Data Validation Report  
E Palestine Site - ER  
EPA Contract No.: 68HE0519D0005  
Task Order/Task Order Line Item No.: 68HE0520F0032/0001EB201  
Document Tracking No. 2016**

Dear Mr. Peters:

Tetra Tech, Inc. (Tetra Tech) is submitting these data validation reports for 78 air samples (including 5 field duplicate samples, 3 media blank samples and 5 field blank samples) collected at the E Palestine site. The samples were collected on June 22 and 23, 2023, and were analyzed for acrylates by Eurofins Analytics, LLC. The final laboratory data package was received on June 28, 2023.

Analytical data were evaluated in general accordance with the Tetra Tech *Quality Assurance Project Plan, East Palestine Train Derailment Site, East Palestine, Columbiana County, Ohio, Revision 3* (April 2023), the Tetra Tech *Quality Assurance Project Plan, Superfund Technical Assessment and Response Team (START V), EPA Region 5, Revision 4* (August 2022), and the *National Functional Guidelines (NFG) for Organic Superfund Methods Data Review* (November 2020).

No qualification or rejection of results was required for these data packages.

If you have any questions regarding this data validation report, please contact me via the project manager.

Sincerely,

**Diane  
MacMillan**  
Digitally signed by  
Diane MacMillan  
Date: 2023.09.22  
16:41:17 -06'00'

Chemical Engineer, PE

Enclosure

cc: Karl Schultz, Tetra Tech Program Manager  
Dustin Grams, Tetra Tech Project Manager  
Mayra Arroyo Ortiz, Tetra Tech Project Document Control Coordinator  
TO-TOLIN File

**ATTACHMENT**

**DATA VALIDATION REPORTS  
EUROFINS ANALYTICS, LLC REPORT NOS.  
B177-153, B177-154, B177-155, AND B177-156**

**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

Site Name	E Palestine Site – ER	TO/TOLIN No.	68HE0520F0032/0001EB201
Document Tracking No.	2016a		
Laboratory Report No.	B177-153	Laboratory	Eurofins Analytics, LLC – Ashland, VA
Analyses	n-Butyl acrylate by NIOSH Method 1450M		
Samples and Matrix	29 air samples including 1 media blank sample, 1 field blank sample, and 2 field duplicate pairs		
Collection Date(s)	06/22/2023		
Field Duplicate Pairs	EPD-PB-OD-04-062223-1/EPD-PB-OD-044-062223-1 EPD-PB-BKBA-01-062223-1/EPD-PB-BKBA-011-062223-1		
Field QC Blanks	EPD-PB-FB-01-062223-1 and EPD-PB-MB-01-062223-1		

**INTRODUCTION**

This checklist summarizes the Stage 2A validation performed on the subject laboratory report, in accordance with the U.S. Environmental Protection Agency (EPA) *Guidance for Labeling Externally Validated Laboratory Analytical Data for Superfund Use* (January 2009). Analytical data were evaluated in general accordance with the Tetra Tech *Quality Assurance Project Plan, East Palestine Train Derailment Site, East Palestine, Columbiana County, Ohio, Revision 3* (April 2023), the Tetra Tech *Quality Assurance Project Plan, Superfund Technical Assessment and Response Team (START V), EPA Region 5, Revision 4* (August 2022), and the EPA *National Functional Guidelines for Organic Superfund Methods Data Review* (November 2020).

**OVERALL EVALUATION**

No rejection or qualification of results was required for this data package. The results may be used as reported by the laboratory.

**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

**Data completeness:**

Within Criteria	Exceedance/Notes
Y	<p>The results for the field blank sample and media blank sample were reported in units of micrograms (µg) while the other field sample results were reported in units of µg, milligrams per cubic meter (mg/m3), and parts per million (ppm) (volume) in the laboratory report and only in units of ppm in the laboratory electronic data deliverable (EDD).</p> <p>The site-specific QAPP specifies analysis of acrylates in air by Eurofins Analytics, LLC standard operating procedure (SOP) IHGC-001-v.22-3. The laboratory confirmed that NIOSH Method 1450M, which is mentioned in the laboratory deliverables, is equivalent to SOP IHGC-001-v.22-3; therefore, these method references may be used interchangeably.</p> <p>The sample analysis time is reported as a default value of 00:00 hours for the laboratory control spike duplicate (LCS D) in the analysis date field of the laboratory EDD. Since the sample analysis time for the LCS D is not required for the validated EDD, this value was not manually revised.</p>

**Sample preservation, receipt, and holding times:**

Within Criteria	Exceedance/Notes
Y	

**Method blanks:**

Within Criteria	Exceedance/Notes
N	<p>Laboratory method blank (LMB IHG230626E) and laboratory reagent blank (LRB IHG230626E) were reported as nondetect (flagged U) with a result of 0 µg in the laboratory EDD rather than at the reporting limit (RL). No qualifications were applied because all n-butyl acrylate sample results were nondetect.</p>

**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

**Field blanks:**

Within Criteria	Exceedance/Notes
N	Per the site-specific QAPP, 1 field blank sample is required for every 10 field samples collected; therefore, 2 field blank samples should have been collected with this sample group. While there was only one field blank sample with this sample group, no qualifications were applied since all samples were nondetect.

**Surrogates and labeled compounds:**

Within Criteria	Exceedance/Notes
NA	

**MS/MSDs:**

Within Criteria	Exceedance/Notes
NA	

**Laboratory duplicates:**

Within Criteria	Exceedance/Notes
NA	

**Field duplicates:**

Within Criteria	Exceedance/Notes
Y	

**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

**LCSs/LCSDs:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
N	The site-specific QAPP requires a LRB, laboratory media blank (LMB), laboratory control sample (LCS), and LCSD to be analyzed per batch of 20 samples. However, the laboratory analyzed 29 field samples in 1 sample preparation batch that included 1 LRB, LMB, LCS, and LCSD, when the batch should have included 2 LRBs, LMBs, LCSs, and LCSDs. The laboratory was contacted about this deviation from the site-specific QAPP and agreed that moving forward they would follow the quality control (QC) sample frequency requirements in the site-specific QAPP. No qualifications were applied based on professional judgment because the QC sample results met the QAPP acceptance criteria, and the QC sample results from previous datasets for this project have met the QAPP acceptance criteria.

**Sample dilutions:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
NA	

**Re-extraction and reanalysis:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
NA	

**MDLs/RLs:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
Y	Method detection limits (MDL) were not reported. Nondetect sample results are reported as less than the RL in the laboratory report and at the RL (flagged U) in the validated EDD and attached analytical results summary.

**Tentatively identified compounds:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
NA	

**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

**Other [None]:**

Within Criteria	Exceedance/Notes
NA	

**Overall Qualifications:**

See results summary pages attached for changes to the laboratory qualifiers based upon this validation. The following is a list of qualifiers and definitions that may be used for the validation of this data package:

J	The analyte was positively identified; the associated value is the approximate concentration of the analyte in the sample.
J+	The analyte was positively identified; the associated value is the approximate concentration of the analyte in the sample and may be biased high.
J-	The analyte was positively identified; the associated value is the approximate concentration of the analyte in the sample and may be biased low.
NJ	The analysis indicates the presence of an analyte that has been “tentatively identified” and the associated value is the approximate concentration of the analyte in the sample.
R	The sample result is rejected as unusable due to serious deficiencies in one or more quality control criteria. The analyte may or may not be present in the sample.
U	The analyte was analyzed for, but was not detected at or above the associated value (reporting limit).
UJ	The analyte was analyzed for, but was not detected at or above the associated value (reporting limit), which is considered approximate due to deficiencies in one or more quality control criteria.

E PALESTINE SITE - ER AIR ANALYTICAL RESULTS SUMMARY  
EUROFINS ANALYTICS, LLC REPORT NO. B177-153

Sample_ID	Method	CAS#	Analyte	Lab_Result	Lab_Qual	RL	Units	VAL_Result	VAL_Qual
EPD-PB-BKBA-01-062223-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-BKBA-011-062223-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-BKBA-02-062223-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-CM-06-062223-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-CM-07-062223-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-CM-08-062223-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-CM-09-062223-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-CM-10-062223-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-CM-11-062223-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-CM-12-062223-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-CM-14-062223-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-DW-G-062223-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-OD-01-062223-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-OD-02-062223-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-OD-03-062223-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-OD-04-062223-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-OD-044-062223-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-OD-05-062223-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-OD-06-062223-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-OD-07-062223-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-UW-C-062223-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-WA-01-062223-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-WA-02-062223-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-WA-03-062223-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-WA-04-062223-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-WA-05-062223-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-WA-06-062223-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-FB-01-062223-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	2	U	2	ug	2	U
EPD-PB-MB-01-062223-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	2	U	2	ug	2	U



**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

Site Name	E Palestine Site – ER	TO/TOLIN No.	68HE0520F0032/0001EB201
Document Tracking No.	2016b		
Laboratory Report No.	B177-154	Laboratory	Eurofins Analytics, LLC – Ashland, VA
Analyses	n-Butyl acrylate by NIOSH Method 1450M		
Samples and Matrix	32 air samples including 2 field blank samples, 2 media blank samples, and 3 field duplicate pairs		
Collection Date(s)	06/22/2023		
Field Duplicate Pairs	EPD-PB-CM-14-062223-2/EPD-PB-CM-144-062223-2 EPD-PB-WA-05-062223-2/EPD-PB-WA-055-062223-2 EPD-PB-OD-04-062223-2/EPD-PB-OD-044-062223-2		
Field QC Blanks	EPD-PB-FB-02-062223-2, EPD-PB-FB-03-062223-2, EPD-PB-MB-02-062223-2, and EPD-PB-MB-03-062223-2		

**INTRODUCTION**

This checklist summarizes the Stage 2A validation performed on the subject laboratory report, in accordance with the U.S. Environmental Protection Agency (EPA) *Guidance for Labeling Externally Validated Laboratory Analytical Data for Superfund Use* (January 2009). Analytical data were evaluated in general accordance with the Tetra Tech *Quality Assurance Project Plan, East Palestine Train Derailment Site, East Palestine, Columbiana County, Ohio, Revision 3* (April 2023), the Tetra Tech *Quality Assurance Project Plan, Superfund Technical Assessment and Response Team (START V), EPA Region 5, Revision 4* (August 2022), and the EPA *National Functional Guidelines for Organic Superfund Methods Data Review* (November 2020).

**OVERALL EVALUATION**

No rejection or qualification of results was required for this data package. The results may be used as reported by the laboratory.

**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

**Data completeness:**

Within Criteria	Exceedance/Notes
Y	<p>The results for the field blank samples and media blank samples were reported in units of micrograms (<math>\mu\text{g}</math>) while the other field sample results were reported in units of <math>\mu\text{g}</math>, milligrams per cubic meter (<math>\text{mg}/\text{m}^3</math>), and parts per million (ppm) (volume) in the laboratory report and only in units of ppm in the laboratory electronic data deliverable (EDD).</p> <p>The sample analysis time is reported as a default value of 00:00 hours for the laboratory control spike duplicate (LCSD) in the analysis date field of the laboratory EDD. Since the sample analysis time for the LCSD is not required for the validated EDD, this value was not manually revised.</p> <p>The site-specific QAPP specifies analysis of acrylates in air by Eurofins Analytics, LLC standard operating procedure (SOP) IHGC-001-v.22-3. The laboratory confirmed that NIOSH Method 1450M, which is mentioned in the laboratory deliverables, is equivalent to SOP IHGC-001-v.22-3; therefore, these method references may be used interchangeably.</p>

**Sample preservation, receipt, and holding times:**

Within Criteria	Exceedance/Notes
Y	

**Method blanks:**

Within Criteria	Exceedance/Notes
N	Laboratory method blank (LMB IHG230626F) and laboratory reagent blank (LRB IHG230626F) were reported as nondetect (flagged U) with a result of 0 $\mu\text{g}$ in the laboratory EDD rather than at the reporting limit (RL). No qualifications were applied because all n-butyl acrylate sample results were nondetect.

**Field blanks:**

Within Criteria	Exceedance/Notes
Y	

**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

**Surrogates and labeled compounds:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
NA	

**MS/MSDs:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
NA	

**Laboratory duplicates:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
NA	

**Field duplicates:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
Y	

**LCSs/LCSDs:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
N	The site-specific QAPP requires an LRB, laboratory media blank, laboratory control sample (LCS), and LCSD to be analyzed per batch of 20 samples. However, the laboratory analyzed 32 field samples total in 1 sample preparation batch that included 1 LRB, LMB, LCS, and LCSD, when the batch should have included 2 LRBs, LMBs, LCSs, and LCSDs. The laboratory was contacted about this deviation from the site-specific QAPP and agreed that moving forward they would follow the quality control (QC) sample frequency requirements in the site-specific QAPP. No qualifications were applied based on professional judgment because the QC sample results met the QAPP acceptance criteria, and the QC sample results from previous datasets for this project have met the QAPP acceptance criteria.

**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

**Sample dilutions:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
Y	

**Re-extraction and reanalysis:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
NA	

**MDLs/RLs:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
Y	Method detection limits (MDL) were not reported. Nondetect sample results are reported as less than the RL in the laboratory report and at the RL (flagged U) in the validated EDD and attached analytical results summary.

**Tentatively identified compounds:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
NA	

**Other [None]:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
NA	

## DATA VALIDATION CHECKLIST – STAGE 2A EPA REGION 5 START CONTRACT

### Overall Qualifications:

See results summary pages attached for changes to the laboratory qualifiers based upon this validation. The following is a list of qualifiers and definitions that may be used for the validation of this data package:

J	The analyte was positively identified; the associated value is the approximate concentration of the analyte in the sample.
J+	The analyte was positively identified; the associated value is the approximate concentration of the analyte in the sample and may be biased high.
J-	The analyte was positively identified; the associated value is the approximate concentration of the analyte in the sample and may be biased low.
NJ	The analysis indicates the presence of an analyte that has been “tentatively identified” and the associated value is the approximate concentration of the analyte in the sample.
R	The sample result is rejected as unusable due to serious deficiencies in one or more quality control criteria. The analyte may or may not be present in the sample.
U	The analyte was analyzed for, but was not detected at or above the associated value (reporting limit).
UJ	The analyte was analyzed for, but was not detected at or above the associated value (reporting limit), which is considered approximate due to deficiencies in one or more quality control criteria.

E PALESTINE SITE - ER AIR ANALYTICAL RESULTS SUMMARY  
EUROFINS ANALYTICS, LLC REPORT NO. B177-154

Sample_ID	Method	CAS#	Analyte	Lab_Result	Lab_Qual	RL	Units	VAL_Result	VAL_Qual
EPD-PB-BKBA-01-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-BKBA-02-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-CM-06-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-CM-07-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-CM-08-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-CM-09-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-CM-10-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-CM-11-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-CM-12-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-CM-14-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-CM-144-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-DW-G-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-OD-01-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-OD-02-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-OD-03-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-OD-04-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-OD-044-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-OD-05-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-OD-06-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-OD-07-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-UW-C-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-WA-01-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-WA-02-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-WA-03-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-WA-04-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-WA-05-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-WA-055-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-WA-06-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0091	U	0.0091	ppm	0.0091	U
EPD-PB-FB-02-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	2	U	2	ug	2	U
EPD-PB-FB-03-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	2	U	2	ug	2	U
EPD-PB-MB-02-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	2	U	2	ug	2	U
EPD-PB-MB-03-062223-2	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	2	U	2	ug	2	U

**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

Site Name	E Palestine Site – ER	TO/TOLIN No.	68HE0520F0032/0001EB201
Document Tracking No.	2016c	Laboratory	Eurofins Analytics, LLC – Ashland, VA
Laboratory Report No.	B177-155	Analyses	
		2-Ethylhexyl acrylate and n-butyl acrylate by laboratory standard operating procedure (SOP) IHGC-P029	
Samples and Matrix	Nine air samples including one field blank sample		
Collection Date(s)	06/23/2023		
Field Duplicate Pairs	None		
Field QC Blanks	EPD-ST-FB-062323-1		

**INTRODUCTION**

This checklist summarizes the Stage 2A validation performed on the subject laboratory report, in accordance with the U.S. Environmental Protection Agency (EPA) *Guidance for Labeling Externally Validated Laboratory Analytical Data for Superfund Use* (January 2009). Analytical data were evaluated in general accordance with the Tetra Tech *Quality Assurance Project Plan, East Palestine Train Derailment Site, East Palestine, Columbiana County, Ohio, Revision 3* (April 2023), the Tetra Tech *Quality Assurance Project Plan, Superfund Technical Assessment and Response Team (START V), EPA Region 5, Revision 4* (August 2022), and the EPA *National Functional Guidelines for Organic Superfund Methods Data Review* (November 2020).

**OVERALL EVALUATION**

No rejection or qualification of results was required for this data package. The results may be used as reported by the laboratory.

**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

**Data completeness:**

Within Criteria	Exceedance/Notes
Y	<p>The results for the field blank sample were reported in units of micrograms (<math>\mu\text{g}</math>) while the other sample results were reported in units of <math>\mu\text{g}</math>, milligrams per cubic meter (<math>\text{mg}/\text{m}^3</math>), and parts per million (ppm) (volume) in the laboratory report and only in units of ppm in the laboratory electronic data deliverable (EDD).</p> <p>Rohm &amp; Haas IH9805 is cited in the AIHA certification as “IHGC-P029” and may be cited by the abbreviation “Rohm &amp; Haas IH9805” or “IHGC-P029” interchangeably throughout the laboratory report.</p> <p>The sample analysis time is reported as a default value of 00:00 hours for the laboratory control spike duplicate (LCSD) in the analysis date field of the laboratory EDD. Since the sample analysis time for the LCSD is not required for the validated EDD, this value was not manually revised.</p>

**Sample preservation, receipt, and holding times:**

Within Criteria	Exceedance/Notes
Y	

**Method blanks:**

Within Criteria	Exceedance/Notes
N	Laboratory method blank (LMB IHG230626A) and laboratory reagent blank (LRB IHG230626A) were reported as nondetect (flagged U) with a result of 0 $\mu\text{g}$ in the laboratory EDD rather than at the reporting limit (RL). No qualifications were applied because all sample results were nondetect.

**Field blanks:**

Within Criteria	Exceedance/Notes
Y	



**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

**Surrogates and labeled compounds:**

Within Criteria	Exceedance/Notes
NA	

**MS/MSDs:**

Within Criteria	Exceedance/Notes
NA	

**Laboratory duplicates:**

Within Criteria	Exceedance/Notes
NA	

**Field duplicates:**

Within Criteria	Exceedance/Notes
NA	

**LCSs/LCSDs:**

Within Criteria	Exceedance/Notes
N	<p>The average laboratory control spike (LCS) and LCSD results associated with all sample results were outside of the acceptance criteria for 2-ethylhexyl acrylate. All sample results were nondetect; therefore, no qualifications were applied.</p> <p>The laboratory report and the laboratory EDD have one or more minor discrepancies in the LCS/LCSD results (+/- 1 ug) and/or percent recoveries (+/- 1%) that were verified with the laboratory to be a significant figures issue. No qualifications were applied.</p>

**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

**Sample dilutions:**

Within Criteria	Exceedance/Notes
NA	

**Re-extraction and reanalysis:**

Within Criteria	Exceedance/Notes
NA	

**MDLs/RLs:**

Within Criteria	Exceedance/Notes
Y	Method detection limits (MDL) were not reported. Nondetect sample results are reported as less than the RL in the laboratory report and at the RL (flagged U) in the validated EDD and attached analytical results summary.

**Tentatively identified compounds:**

Within Criteria	Exceedance/Notes
NA	

**Other [None]:**

Within Criteria	Exceedance/Notes
NA	

## DATA VALIDATION CHECKLIST – STAGE 2A EPA REGION 5 START CONTRACT

### Overall Qualifications:

See results summary pages attached for changes to the laboratory qualifiers based upon this validation. The following is a list of qualifiers and definitions that may be used for the validation of this data package:

J	The analyte was positively identified; the associated value is the approximate concentration of the analyte in the sample.
J+	The analyte was positively identified; the associated value is the approximate concentration of the analyte in the sample and may be biased high.
J-	The analyte was positively identified; the associated value is the approximate concentration of the analyte in the sample and may be biased low.
NJ	The analysis indicates the presence of an analyte that has been “tentatively identified” and the associated value is the approximate concentration of the analyte in the sample.
R	The sample result is rejected as unusable due to serious deficiencies in one or more quality control criteria. The analyte may or may not be present in the sample.
U	The analyte was analyzed for, but was not detected at or above the associated value (reporting limit).
UJ	The analyte was analyzed for, but was not detected at or above the associated value (reporting limit), which is considered approximate due to deficiencies in one or more quality control criteria.

E PALESTINE SITE - ER AIR ANALYTICAL RESULTS SUMMARY  
EUROFINS ANALYTICS, LLC REPORT NO. B177-155

Sample_ID	Method	CAS#	Analyte	Lab_Result	Lab_Qual	RL	Units	VAL_Result	VAL_Qual
EPD-ST-8H-DW-G-062323-1	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-DW-G-062323-1	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-UW-C-062323-1	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-UW-C-062323-1	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-01-062323-1	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-WA-01-062323-1	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-02-062323-1	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-WA-02-062323-1	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-03-062323-1	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-WA-03-062323-1	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-04-062323-1	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-WA-04-062323-1	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-05-062323-1	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-WA-05-062323-1	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-06-062323-1	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-WA-06-062323-1	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-FB-062323-1	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	2.8	U	2.8	ug	2.8	U
EPD-ST-FB-062323-1	IHGC-P029	141-32-2	n-Butyl acrylate	1.3	U	1.3	ug	1.3	U

**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

Site Name	E Palestine Site – ER	TO/TOLIN No.	68HE0520F0032/0001EB201
Document Tracking No.	2016d		
Laboratory Report No.	B177-156	Laboratory	Eurofins Analytics, LLC – Ashland, VA
Analyses	2-Ethylhexyl acrylate and n-butyl acrylate by laboratory standard operating procedure (SOP) IHGC-P029		
Samples and Matrix	Eight air samples including one field blank sample		
Collection Date(s)	06/23/2023		
Field Duplicate Pairs	None		
Field QC Blanks	EPD-ST-FB-062323-2		

**INTRODUCTION**

This checklist summarizes the Stage 2A validation performed on the subject laboratory report, in accordance with the U.S. Environmental Protection Agency (EPA) *Guidance for Labeling Externally Validated Laboratory Analytical Data for Superfund Use* (January 2009). Analytical data were evaluated in general accordance with the Tetra Tech *Quality Assurance Project Plan, East Palestine Train Derailment Site, East Palestine, Columbiana County, Ohio, Revision 3* (April 2023), the Tetra Tech *Quality Assurance Project Plan, Superfund Technical Assessment and Response Team (START V), EPA Region 5, Revision 4* (August 2022), and the EPA *National Functional Guidelines for Organic Superfund Methods Data Review* (November 2020).

**OVERALL EVALUATION**

No rejection or qualification of results was required for this data package. The results may be used as reported by the laboratory.

**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

**Data completeness:**

Within Criteria	Exceedance/Notes
N	<p>Samples EPD-ST-8H-WA-01-062323-2 and EPD-ST-8H-WA-66-062323-2 were voided for all analyses due to the samples being found on the ground during collection. No qualifications were applied.</p> <p>The results for the field blank sample were reported in units of micrograms (µg) while the other sample results were reported in units of µg, milligrams per cubic meter (mg/m3), and parts per million (ppm) (volume) in the laboratory report and only in units of ppm in the laboratory electronic data deliverable (EDD).</p> <p>Rohm &amp; Haas IH9805 is cited in the AIHA certification as “IHGC-P029” and may be cited by the abbreviation “Rohm &amp; Haas IH9805” or “IHGC-P029” interchangeably throughout the laboratory report.</p> <p>The sample analysis time is reported as a default value of 00:00 hours for the laboratory control spike duplicate (LCSD) in the analysis date field of the laboratory EDD. Since the sample analysis time for the LCSD is not required for the validated EDD, this value was not manually revised.</p>

**Sample preservation, receipt, and holding times:**

Within Criteria	Exceedance/Notes
Y	

**Method blanks:**

Within Criteria	Exceedance/Notes
N	<p>Laboratory method blank (LMB IHG230626A) and laboratory reagent blank (LRB IHG230626A) were reported as nondetect (flagged U) with a result of 0 µg in the laboratory EDD rather than at the reporting limit (RL). No qualifications were applied because all sample results were nondetect.</p>

**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

**Field blanks:**

Within Criteria	Exceedance/Notes
Y	

**Surrogates and labeled compounds:**

Within Criteria	Exceedance/Notes
NA	

**MS/MSDs:**

Within Criteria	Exceedance/Notes
NA	

**Laboratory duplicates:**

Within Criteria	Exceedance/Notes
NA	

**Field duplicates:**

Within Criteria	Exceedance/Notes
NA	

**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

**LCSs/LCSDs:**

Within Criteria	Exceedance/Notes
N	<p>The average LCS/LCSD results associated with all sample results were outside of the acceptance criteria for 2-ethylhexyl acrylate. All sample results were nondetect; therefore, no qualifications were applied.</p> <p>The laboratory report and the laboratory EDD have one or more minor discrepancies in the LCS/LCSD results (+/- 1 ug) and/or percent recoveries (+/- 1%) that were verified with the laboratory to be a significant figures issue. No qualifications were applied.</p>

**Sample dilutions:**

Within Criteria	Exceedance/Notes
NA	

**Re-extraction and reanalysis:**

Within Criteria	Exceedance/Notes
NA	

**MDLs/RLs:**

Within Criteria	Exceedance/Notes
Y	Method detection limits (MDL) were not reported. Nondetect sample results are reported as less than the RL in the laboratory report and at the RL (flagged U) in the validated EDD and attached analytical results summary.

**Tentatively identified compounds:**

Within Criteria	Exceedance/Notes
NA	



**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

**Other [None]:**

Within Criteria	Exceedance/Notes
NA	

**Overall Qualifications:**

See results summary pages attached for changes to the laboratory qualifiers based upon this validation. The following is a list of qualifiers and definitions that may be used for the validation of this data package:

J	The analyte was positively identified; the associated value is the approximate concentration of the analyte in the sample.
J+	The analyte was positively identified; the associated value is the approximate concentration of the analyte in the sample and may be biased high.
J-	The analyte was positively identified; the associated value is the approximate concentration of the analyte in the sample and may be biased low.
NJ	The analysis indicates the presence of an analyte that has been “tentatively identified” and the associated value is the approximate concentration of the analyte in the sample.
R	The sample result is rejected as unusable due to serious deficiencies in one or more quality control criteria. The analyte may or may not be present in the sample.
U	The analyte was analyzed for, but was not detected at or above the associated value (reporting limit).
UJ	The analyte was analyzed for, but was not detected at or above the associated value (reporting limit), which is considered approximate due to deficiencies in one or more quality control criteria.

E PALESTINE SITE - ER AIR ANALYTICAL RESULTS SUMMARY  
EUROFINS ANALYTICS, LLC REPORT NO. B177-156

Sample_ID	Method	CAS#	Analyte	Lab_Result	Lab_Qual	RL	Units	VAL_Result	VAL_Qual
EPD-ST-8H-DW-A-062323-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-DW-A-062323-2	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-UW-E-062323-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-UW-E-062323-2	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-02-062323-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-WA-02-062323-2	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-03-062323-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-WA-03-062323-2	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-04-062323-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-WA-04-062323-2	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-05-062323-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-WA-05-062323-2	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-06-062323-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.016	U	0.016	ppm	0.016	U
EPD-ST-8H-WA-06-062323-2	IHGC-P029	141-32-2	n-Butyl acrylate	0.011	U	0.011	ppm	0.011	U
EPD-ST-FB-062323-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	2.8	U	2.8	ug	2.8	U
EPD-ST-FB-062323-2	IHGC-P029	141-32-2	n-Butyl acrylate	1.3	U	1.3	ug	1.3	U