



December 12, 2023

Mr. Josh Peters  
On-Scene Coordinator  
U.S. Environmental Protection Agency, Region 5  
Superfund and Emergency Management Division  
2565 Plymouth Road  
Ann Arbor, MI 48105

We are in the process of ensuring this document is accessible to all audiences. If you need assistance accessing this document, or any material on the EPA East Palestine, Ohio emergency response web pages, please contact the Region 5 Public Information Officer on-call at: R5\_EastPalestine@epa.gov

**Subject: Data Validation Report  
E Palestine Site - ER  
EPA Contract No.: 68HE0519D0005  
Task Order/Task Order Line Item No.: 68HE0520F0032/0001EB201  
Document Tracking No. 2178**

Dear Mr. Peters:

Tetra Tech, Inc. (Tetra Tech) is submitting this data validation report for 60 air samples (including 4 field duplicate samples, 4 field blank samples, and 3 media blanks) collected at the E Palestine site. The samples were collected on August 10 and 11, 2023, and were analyzed for acrylates by Eurofins Analytics, LLC. The final laboratory data package was received on August 16, 2023.

Analytical data were evaluated in general accordance with the Tetra Tech *Quality Assurance Project Plan, East Palestine Train Derailment Site, East Palestine, Columbiana County, Ohio*, Revision 3 (April 2023), the Tetra Tech *Quality Assurance Project Plan, Superfund Technical Assessment and Response Team (START V), EPA Region 5*, Revision 4 (August 2022), and the *National Functional Guidelines (NFG) for Organic Superfund Methods Data Review* (November 2020).

No rejection or qualification of results was required for these data packages. The results may be used as received from the laboratory.

If you have any questions regarding this data validation report, please contact me via the project manager.

Sincerely,

Celina Barnett-Cashman  
Digitally signed by Celina Barnett-Cashman  
Date: 2023.12.12 11:12:59 -06'00'

Environmental Chemist

Enclosure

cc: Karl Schultz, Tetra Tech Program Manager  
Dustin Grams, Tetra Tech Project Manager  
Mayra Arroyo Ortiz, Tetra Tech Project Document Control Coordinator  
TO-TOLIN File

**ATTACHMENT**

**DATA VALIDATION REPORT  
EUROFINS ANALYTICS, LLC REPORT  
NOS. B226-110, B226-112, B226-113, AND B226-114**

**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

Site Name	E Palestine Site - ER	TO/TOLIN No.	68HE0520F0032/0001EB201
Document Tracking No.	2178a		
Laboratory Report No.	B226-110	Laboratory	Eurofins Analytics, LLC – Ashland, VA
Analyses	2-Ethylhexyl acrylate and n-butyl acrylate by laboratory standard operating procedure (SOP) IHGC-P029		
Samples and Matrix	Nine air samples including one field blank		
Collection Date(s)	08/10/2023		
Field Duplicate Pairs	None		
Field QC Blanks	EPD-ST-FB-081023-1		

**INTRODUCTION**

This checklist summarizes the Stage 2A validation performed on the subject laboratory report, in accordance with the U.S. Environmental Protection Agency (EPA) *Guidance for Labeling Externally Validated Laboratory Analytical Data for Superfund Use* (January 2009). Analytical data were evaluated in general accordance with the Tetra Tech *Quality Assurance Project Plan, East Palestine Train Derailment Site, East Palestine, Columbiana County, Ohio*, Revision 3 (April 2023), the Tetra Tech *Quality Assurance Project Plan, Superfund Technical Assessment and Response Team (START V), EPA Region 5*, Revision 4 (August 2022), and the EPA *National Functional Guidelines (NFG) for Organic Superfund Methods Data Review* (November 2020).

**OVERALL EVALUATION**

No rejection or qualification of results was required for this data package. The results may be used as reported by the laboratory.

**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

**Data completeness:**

Within Criteria	Exceedance/Notes
N	<p>The results for the field blank were reported in units of micrograms (µg) while the other sample results were reported in units of µg, milligrams per cubic meter (mg/m3), and parts per million (ppm) (volume) in the laboratory report and only in units of ppm in the laboratory electronic data deliverable (EDD).</p> <p>Rohm &amp; Haas IH9805 was cited in the AIHA certification as “IHGC-P029” and may be cited by the abbreviation “Rohm &amp; Haas IH9805” or “IHGC-P029” interchangeably throughout the laboratory report.</p> <p>Note, the following fields in the laboratory EDD may be formatted as date only or as date/time: Date_Collected, Date_Received, Date_Extracted, and Date_Analyzed. The time value was not required to be provided in the EDD. If no time value was provided, then the entered value may appear as date only or with a default time value of 0:00, 00:00, or similar.</p>

**Sample preservation, receipt, and holding times:**

Within Criteria	Exceedance/Notes
Y	

**Method blanks:**

Within Criteria	Exceedance/Notes
N	<p>Nondetect results for laboratory method blank LMB IHG230814G and laboratory reagent blank LRB IHG230814G were reported as “0” in the laboratory EDD rather than at the reporting limit (RL). The laboratory was contacted on August 28, 2023, and agreed to report nondetect laboratory method blank and LRB results at the RL in future laboratory EDDs.</p>

**Field blanks:**

Within Criteria	Exceedance/Notes
Y	

**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

**Surrogates and labeled compounds:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
NA	

**MS/MSDs:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
NA	

**Laboratory duplicates:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
NA	

**Field duplicates:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
N	Per the site-specific QAPP, 1 field duplicate sample is required per 20 samples collected. However, fewer than 1 field duplicate sample per 20 samples were collected with this sample group. Based on professional judgement, no qualifications were applied.

**LCSs/LCSDs:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
Y	

**Sample dilutions:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
NA	

**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

**Re-extraction and reanalysis:**

Within Criteria	Exceedance/Notes
NA	

**MDLs/RLs:**

Within Criteria	Exceedance/Notes
Y	Method detection limits were not reported. Nondetect sample results were reported as less than the reporting limit in the laboratory report and at the reporting limit (flagged U) in the validated EDD and attached analytical results summary.

**Tentatively identified compounds:**

Within Criteria	Exceedance/Notes
NA	

**Other [None]:**

Within Criteria	Exceedance/Notes
NA	

## DATA VALIDATION CHECKLIST – STAGE 2A EPA REGION 5 START CONTRACT

### Overall Qualifications:

See results summary pages attached for changes to the laboratory qualifiers based upon this validation. The following is a list of qualifiers and definitions that may be used for the validation of this data package:

J	The analyte was positively identified; the associated value is the approximate concentration of the analyte in the sample.
J+	The analyte was positively identified; the associated value is the approximate concentration of the analyte in the sample and may be biased high.
J-	The analyte was positively identified; the associated value is the approximate concentration of the analyte in the sample and may be biased low.
NJ	The analysis indicates the presence of an analyte that has been “tentatively identified” and the associated value is the approximate concentration of the analyte in the sample.
R	The sample result is rejected as unusable due to serious deficiencies in one or more quality control criteria. The analyte may or may not be present in the sample.
U	The analyte was analyzed for, but was not detected at or above the associated value (reporting limit).
UJ	The analyte was analyzed for, but was not detected at or above the associated value (reporting limit), which is considered approximate due to deficiencies in one or more quality control criteria.

E PALESTINE SITE - ER AIR ANALYTICAL RESULTS SUMMARY  
EUROFINS ANALYTICS, LLC REPORT NO. B226-110

Samp_ID	Method	CAS#	Analyte	Lab_Result	Lab_Qual	RL	Units	VAL_Result	VAL_Qual
EPD-ST-8H-DW-C-081023-1	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-DW-C-081023-1	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-UW-G-081023-1	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.014	U	0.014	ppm	0.014	U
EPD-ST-8H-UW-G-081023-1	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-01-081023-1	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-WA-01-081023-1	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-02-081023-1	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-WA-02-081023-1	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-03-081023-1	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-WA-03-081023-1	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-04-081023-1	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.014	U	0.014	ppm	0.014	U
EPD-ST-8H-WA-04-081023-1	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-05-081023-1	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-WA-05-081023-1	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-06-081023-1	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-WA-06-081023-1	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-FB-081023-1	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	2.8	U	2.8	ug	2.8	U
EPD-ST-FB-081023-1	IHGC-P029	141-32-2	n-Butyl acrylate	1.3	U	1.3	ug	1.3	U



**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

Site Name	E Palestine Site - ER	TO/TOLIN No.	68HE0520F0032/0001EB201
Document Tracking No.	2178b		
Laboratory Report No.	B226-112	Laboratory	Eurofins Analytics, LLC – Ashland, VA
Analyses	2-Ethylhexyl acrylate and n-butyl acrylate by laboratory standard operating procedure (SOP) IHGC-P029		
Samples and Matrix	10 air samples including 1 field blank and 1 field duplicate pair		
Collection Date(s)	08/11/2023		
Field Duplicate Pairs	EPD-ST-8H-WA-05-081123-2/EPD-PB-8H-WA-55-081123-2		
Field QC Blanks	EPD-ST-FB-081123-2		

**INTRODUCTION**

This checklist summarizes the Stage 2A validation performed on the subject laboratory report, in accordance with the U.S. Environmental Protection Agency (EPA) *Guidance for Labeling Externally Validated Laboratory Analytical Data for Superfund Use* (January 2009). Analytical data were evaluated in general accordance with the Tetra Tech *Quality Assurance Project Plan, East Palestine Train Derailment Site, East Palestine, Columbiana County, Ohio*, Revision 3 (April 2023), the Tetra Tech *Quality Assurance Project Plan, Superfund Technical Assessment and Response Team (START V), EPA Region 5*, Revision 4 (August 2022), and the EPA *National Functional Guidelines (NFG) for Organic Superfund Methods Data Review* (November 2020).

**OVERALL EVALUATION**

No rejection or qualification of results was required for this data package. The results may be used as reported by the laboratory.

**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

**Data completeness:**

Within Criteria	Exceedance/Notes
N	<p>The results for the field blank were reported in units of micrograms (µg) while the other sample results were reported in units of µg, milligrams per cubic meter (mg/m3), and parts per million (ppm) (volume) in the laboratory report and only in units of ppm in the laboratory electronic data deliverable (EDD).</p> <p>Rohm &amp; Haas IH9805 was cited in the AIHA certification as “IHGC-P029” and may be cited by the abbreviation “Rohm &amp; Haas IH9805” or “IHGC-P029” interchangeably throughout the laboratory report.</p> <p>Note, the following fields in the laboratory EDD may be formatted as date only or as date/time: Date_Collected, Date_Received, Date_Extracted, and Date_Analyzed. The time value was not required to be provided in the EDD. If no time value was provided, then the entered value may appear as date only or with a default time value of 0:00, 00:00, or similar.</p>

**Sample preservation, receipt, and holding times:**

Within Criteria	Exceedance/Notes
Y	

**Method blanks:**

Within Criteria	Exceedance/Notes
N	<p>Nondetect results for laboratory method blank LMB IHG230814H and laboratory reagent blank LRB IHG230814H were reported as “0” in the laboratory EDD rather than at the reporting limit (RL). The laboratory was contacted on August 28, 2023, and agreed to report nondetect laboratory method blank and LRB results at the RL in future laboratory EDDs.</p>

**Field blanks:**

Within Criteria	Exceedance/Notes
Y	

**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

**Surrogates and labeled compounds:**

Within Criteria	Exceedance/Notes
NA	

**MS/MSDs:**

Within Criteria	Exceedance/Notes
NA	

**Laboratory duplicates:**

Within Criteria	Exceedance/Notes
NA	

**Field duplicates:**

Within Criteria	Exceedance/Notes
Y	

**LCSs/LCSDs:**

Within Criteria	Exceedance/Notes
Y	

**Sample dilutions:**

Within Criteria	Exceedance/Notes
NA	

**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

**Re-extraction and reanalysis:**

Within Criteria	Exceedance/Notes
NA	

**MDLs/RLs:**

Within Criteria	Exceedance/Notes
Y	Method detection limits were not reported. Nondetect sample results were reported as less than the reporting limit in the laboratory report and at the reporting limit (flagged U) in the validated EDD and attached analytical results summary.

**Tentatively identified compounds:**

Within Criteria	Exceedance/Notes
NA	

**Other [None]:**

Within Criteria	Exceedance/Notes
NA	

## DATA VALIDATION CHECKLIST – STAGE 2A EPA REGION 5 START CONTRACT

### Overall Qualifications:

See results summary pages attached for changes to the laboratory qualifiers based upon this validation. The following is a list of qualifiers and definitions that may be used for the validation of this data package:

J	The analyte was positively identified; the associated value is the approximate concentration of the analyte in the sample.
J+	The analyte was positively identified; the associated value is the approximate concentration of the analyte in the sample and may be biased high.
J-	The analyte was positively identified; the associated value is the approximate concentration of the analyte in the sample and may be biased low.
NJ	The analysis indicates the presence of an analyte that has been “tentatively identified” and the associated value is the approximate concentration of the analyte in the sample.
R	The sample result is rejected as unusable due to serious deficiencies in one or more quality control criteria. The analyte may or may not be present in the sample.
U	The analyte was analyzed for, but was not detected at or above the associated value (reporting limit).
UJ	The analyte was analyzed for, but was not detected at or above the associated value (reporting limit), which is considered approximate due to deficiencies in one or more quality control criteria.

## E PALESTINE SITE - ER AIR ANALYTICAL RESULTS SUMMARY

EUROFINS ANALYTICS, LLC REPORT NO. B226-112

Samp_ID	Method	CAS#	Analyte	Lab_Result	Lab_Qual	RL	Units	VAL_Result	VAL_Qual
EPD-ST-8H-DW-D-081123-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-DW-D-081123-2	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-UW-H-081123-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-UW-H-081123-2	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-01-081123-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-WA-01-081123-2	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-02-081123-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-WA-02-081123-2	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-03-081123-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-WA-03-081123-2	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-04-081123-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-WA-04-081123-2	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-05-081123-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-WA-05-081123-2	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-06-081123-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-WA-06-081123-2	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-55-081123-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-WA-55-081123-2	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-FB-081123-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	2.8	U	2.8	ug	2.8	U
EPD-ST-FB-081123-2	IHGC-P029	141-32-2	n-Butyl acrylate	1.3	U	1.3	ug	1.3	U

**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

Site Name	E Palestine Site - ER	TO/TOLIN No.	68HE0520F0032/0001EB201
Document Tracking No.	2178c	Laboratory	Eurofins Analytics, LLC – Ashland, VA
Laboratory Report No.	B226-113	Analyses	
		2-Ethylhexyl acrylate and n-butyl acrylate by laboratory standard operating procedure (SOP) IHGC-P029	
		Samples and Matrix	
		12 air samples including 1 field blank, 2 media blanks and 1 field duplicate pair	
		Collection Date(s)	
		08/10/2023	
		Field Duplicate Pairs	
		EPD-ST-8H-WA-06-081023-2/EPD- ST-8H-WA-66-081023-2	
		Field QC Blanks	
		EPD-ST-FB-081023-2, EPD-ST-MB-01-081023-2, and EPD-ST-MB-02-081023-2	

**INTRODUCTION**

This checklist summarizes the Stage 2A validation performed on the subject laboratory report, in accordance with the U.S. Environmental Protection Agency (EPA) *Guidance for Labeling Externally Validated Laboratory Analytical Data for Superfund Use* (January 2009). Analytical data were evaluated in general accordance with the Tetra Tech *Quality Assurance Project Plan, East Palestine Train Derailment Site, East Palestine, Columbiana County, Ohio*, Revision 3 (April 2023), the Tetra Tech *Quality Assurance Project Plan, Superfund Technical Assessment and Response Team (START V), EPA Region 5*, Revision 4 (August 2022), and the EPA *National Functional Guidelines (NFG) for Organic Superfund Methods Data Review* (November 2020).

**OVERALL EVALUATION**

No rejection or qualification of results was required for this data package. The results may be used as reported by the laboratory.

**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

**Data completeness:**

Within Criteria	Exceedance/Notes
N	<p>The results for the field blank were reported in units of micrograms (µg) while the other sample results were reported in units of µg, milligrams per cubic meter (mg/m3), and parts per million (ppm) (volume) in the laboratory report and only in units of ppm in the laboratory electronic data deliverable (EDD).</p> <p>Rohm &amp; Haas IH9805 was cited in the AIHA certification as “IHGC-P029” and may be cited by the abbreviation “Rohm &amp; Haas IH9805” or “IHGC-P029” interchangeably throughout the laboratory report.</p> <p>Note, the following fields in the laboratory EDD may be formatted as date only or as date/time: Date_Collected, Date_Received, Date_Extracted, and Date_Analyzed. The time value was not required to be provided in the EDD. If no time value was provided, then the entered value may appear as date only or with a default time value of 0:00, 00:00, or similar.</p>

**Sample preservation, receipt, and holding times:**

Within Criteria	Exceedance/Notes
Y	

**Method blanks:**

Within Criteria	Exceedance/Notes
N	<p>Nondetect results for laboratory method blank LMB IHG230814J and laboratory reagent blank LRB IHG230814J were reported as “0” in the laboratory EDD rather than at the reporting limit (RL). The laboratory was contacted on August 28, 2023, and agreed to report nondetect laboratory method blank and LRB results at the RL in future laboratory EDDs.</p>

**Field blanks:**

Within Criteria	Exceedance/Notes
Y	



**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

**Surrogates and labeled compounds:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
NA	

**MS/MSDs:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
NA	

**Laboratory duplicates:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
NA	

**Field duplicates:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
Y	

**LCSs/LCSDs:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
N	The recoveries of 2-ethylhexyl acrylate in the LCS and the average percent recovery of the LCS/LCSD were above the laboratory control limits. The sample results are nondetect and were therefore not qualified.

**Sample dilutions:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
NA	

**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

**Re-extraction and reanalysis:**

Within Criteria	Exceedance/Notes
NA	

**MDLs/RLs:**

Within Criteria	Exceedance/Notes
Y	Method detection limits were not reported. Nondetect sample results were reported as less than the reporting limit in the laboratory report and at the reporting limit (flagged U) in the validated EDD and attached analytical results summary.

**Tentatively identified compounds:**

Within Criteria	Exceedance/Notes
NA	

**Other [None]:**

Within Criteria	Exceedance/Notes
NA	

## DATA VALIDATION CHECKLIST – STAGE 2A EPA REGION 5 START CONTRACT

### Overall Qualifications:

See results summary pages attached for changes to the laboratory qualifiers based upon this validation. The following is a list of qualifiers and definitions that may be used for the validation of this data package:

J	The analyte was positively identified; the associated value is the approximate concentration of the analyte in the sample.
J+	The analyte was positively identified; the associated value is the approximate concentration of the analyte in the sample and may be biased high.
J-	The analyte was positively identified; the associated value is the approximate concentration of the analyte in the sample and may be biased low.
NJ	The analysis indicates the presence of an analyte that has been “tentatively identified” and the associated value is the approximate concentration of the analyte in the sample.
R	The sample result is rejected as unusable due to serious deficiencies in one or more quality control criteria. The analyte may or may not be present in the sample.
U	The analyte was analyzed for, but was not detected at or above the associated value (reporting limit).
UJ	The analyte was analyzed for, but was not detected at or above the associated value (reporting limit), which is considered approximate due to deficiencies in one or more quality control criteria.

## E PALESTINE SITE - ER AIR ANALYTICAL RESULTS SUMMARY

EUROFINS ANALYTICS, LLC REPORT NO. B226-113

Samp_ID	Method	CAS#	Analyte	Lab_Result	Lab_Qual	RL	Units	VAL_Result	VAL_Qual
EPD-ST-8H-DW-D-081023-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-DW-D-081023-2	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-UW-H-081023-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-UW-H-081023-2	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-01-081023-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.016	U	0.016	ppm	0.016	U
EPD-ST-8H-WA-01-081023-2	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-02-081023-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-WA-02-081023-2	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-03-081023-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-WA-03-081023-2	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-04-081023-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-WA-04-081023-2	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-05-081023-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-WA-05-081023-2	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-06-081023-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-WA-06-081023-2	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-8H-WA-66-081023-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	0.015	U	0.015	ppm	0.015	U
EPD-ST-8H-WA-66-081023-2	IHGC-P029	141-32-2	n-Butyl acrylate	0.01	U	0.01	ppm	0.01	U
EPD-ST-FB-081023-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	2.8	U	2.8	ug	2.8	U
EPD-ST-FB-081023-2	IHGC-P029	141-32-2	n-Butyl acrylate	1.3	U	1.3	ug	1.3	U
EPD-ST-MB-01-081023-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	2.8	U	2.8	ug	2.8	U
EPD-ST-MB-01-081023-2	IHGC-P029	141-32-2	n-Butyl acrylate	1.3	U	1.3	ug	1.3	U
EPD-ST-MB-02-081023-2	IHGC-P029	103-11-7	2-Ethylhexyl acrylate	2.8	U	2.8	ug	2.8	U
EPD-ST-MB-02-081023-2	IHGC-P029	141-32-2	n-Butyl acrylate	1.3	U	1.3	ug	1.3	U

**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

Site Name	E Palestine Site - ER	TO/TOLIN No.	68HE0520F0032/0001EB201
Document Tracking No.	2178d		
Laboratory Report No.	B226-114	Laboratory	Eurofins Analytics, LLC – Ashland, VA
Analyses	n-Butyl acrylate by NIOSH Method 1450M		
Samples and Matrix	29 air samples including 1 field blank, 1 media blank, and 2 field duplicate pairs		
Collection Date(s)	08/11/2023		
Field Duplicate Pairs	EPD-PB-WA-04-081123-1/EPD- PB-WA-044-081123-1 EPD-PB-CM-06-081123-1/EPD- PB-CM-066-081123-1		
Field QC Blanks	EPD-ST-FB-01-081123-1 and EPD-MB-01-081123-1		

**INTRODUCTION**

This checklist summarizes the Stage 2A validation performed on the subject laboratory report, in accordance with the U.S. Environmental Protection Agency (EPA) *Guidance for Labeling Externally Validated Laboratory Analytical Data for Superfund Use* (January 2009). Analytical data were evaluated in general accordance with the Tetra Tech *Quality Assurance Project Plan, East Palestine Train Derailment Site, East Palestine, Columbiana County, Ohio*, Revision 3 (April 2023), the Tetra Tech *Quality Assurance Project Plan, Superfund Technical Assessment and Response Team (START V), EPA Region 5*, Revision 4 (August 2022), and the EPA *National Functional Guidelines (NFG) for Organic Superfund Methods Data Review* (November 2020).

**OVERALL EVALUATION**

No rejection or qualification of results was required for this data package. The results may be used as reported by the laboratory.

**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

**Data completeness:**

Within Criteria	Exceedance/Notes
N	<p>The results for the field blank and media blank were reported in units of micrograms (µg) while the other field sample results were reported in units of µg, milligrams per cubic meter (mg/m<sup>3</sup>), and parts per million (ppm) (volume) in the laboratory report and only in units of ppm in the laboratory electronic data deliverable (EDD).</p> <p>The site-specific QAPP specifies analysis of acrylates in air by Eurofins Analytics, LLC standard operating procedure (SOP) IHGC-001-v.22-3. The laboratory confirmed that NIOSH Method 1450M, which is mentioned in the laboratory deliverables, is equivalent to SOP IHGC-001-v.22-3; therefore, these method references may be used interchangeably.</p> <p>Note, the following fields in the laboratory EDD may be formatted as date only or as date/time: Date_Collected, Date_Received, Date_Extracted, and Date_Analyzed. The time value was not required to be provided in the EDD. If no time value was provided, then the entered value may appear as date only or with a default time value of 0:00, 00:00, or similar.</p>

**Sample preservation, receipt, and holding times:**

Within Criteria	Exceedance/Notes
Y	

**Method blanks:**

Within Criteria	Exceedance/Notes
N	<p>Nondetect results for laboratory method blanks LMB IHG230814E, LMBIHG230814F and laboratory reagent blanks LRB IHG230814E and LRB IHG230814F were reported as “0” in the laboratory EDD rather than at the reporting limit (RL). The laboratory was contacted on August 28, 2023, and agreed to report nondetect laboratory method blank and LRB results at the RL in future laboratory EDDs. No qualifications were applied.</p>

**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

**Field blanks:**

Within Criteria	Exceedance/Notes
N	Only 1 field blank sample was included in this data package although the site-specific QAPP specifies the collection of 1 field blank per 20 field samples. No qualifications were applied because all sample results were nondetect.

**Surrogates and labeled compounds:**

Within Criteria	Exceedance/Notes
NA	

**MS/MSDs:**

Within Criteria	Exceedance/Notes
NA	

**Laboratory duplicates:**

Within Criteria	Exceedance/Notes
NA	

**Field duplicates:**

Within Criteria	Exceedance/Notes
N	Two field duplicate samples were included in this data package although the site-specific QAPP specifies the collection of 1 field duplicate sample per 10 field samples. Based on professional judgement, no qualifications were applied.

**DATA VALIDATION CHECKLIST – STAGE 2A  
EPA REGION 5 START CONTRACT**

**LCSs/LCSDs:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
Y	

**Sample dilutions:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
NA	

**Re-extraction and reanalysis:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
NA	

**MDLs/RLs:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
Y	Method detection limits were not reported. Nondetect sample results were reported as less than the reporting limit in the laboratory report and at the reporting limit (flagged U) in the validated EDD and attached analytical results summary.

**Tentatively identified compounds:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
NA	

**Other [None]:**

<b>Within Criteria</b>	<b>Exceedance/Notes</b>
NA	



## DATA VALIDATION CHECKLIST – STAGE 2A EPA REGION 5 START CONTRACT

### Overall Qualifications:

See results summary pages attached for changes to the laboratory qualifiers based upon this validation. The following is a list of qualifiers and definitions that may be used for the validation of this data package:

J	The analyte was positively identified; the associated value is the approximate concentration of the analyte in the sample.
J+	The analyte was positively identified; the associated value is the approximate concentration of the analyte in the sample and may be biased high.
J-	The analyte was positively identified; the associated value is the approximate concentration of the analyte in the sample and may be biased low.
NJ	The analysis indicates the presence of an analyte that has been “tentatively identified” and the associated value is the approximate concentration of the analyte in the sample.
R	The sample result is rejected as unusable due to serious deficiencies in one or more quality control criteria. The analyte may or may not be present in the sample.
U	The analyte was analyzed for, but was not detected at or above the associated value (reporting limit).
UJ	The analyte was analyzed for, but was not detected at or above the associated value (reporting limit), which is considered approximate due to deficiencies in one or more quality control criteria.

E PALESTINE SITE - ER AIR ANALYTICAL RESULTS SUMMARY  
EUROFINS ANALYTICS, LLC REPORT NO. B226-114

Samp_ID	Method	CAS#	Analyte	Lab_Result	Lab_Qual	RL	Units	VAL_Result	VAL_Qual
EPD-PB-BKBA-01-081123-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0081	U	0.0081	ppm	0.0081	U
EPD-PB-BKBA-02-081123-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0089	U	0.0089	ppm	0.0089	U
EPD-PB-CM-06-081123-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.009	U	0.009	ppm	0.009	U
EPD-PB-CM-066-081123-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.009	U	0.009	ppm	0.009	U
EPD-PB-CM-07-081123-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0089	U	0.0089	ppm	0.0089	U
EPD-PB-CM-08-081123-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0086	U	0.0086	ppm	0.0086	U
EPD-PB-CM-09-081123-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0079	U	0.0079	ppm	0.0079	U
EPD-PB-CM-10-081123-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.008	U	0.008	ppm	0.008	U
EPD-PB-CM-11-081123-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.008	U	0.008	ppm	0.008	U
EPD-PB-CM-12-081123-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.008	U	0.008	ppm	0.008	U
EPD-PB-CM-14-081123-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.008	U	0.008	ppm	0.008	U
EPD-PB-DW-D-081123-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0072	U	0.0072	ppm	0.0072	U
EPD-PB-FB-01-081123-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	2	U	2	ug	2	U
EPD-PB-MB-01-081123-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	2	U	2	ug	2	U
EPD-PB-OD-01-081123-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0069	U	0.0069	ppm	0.0069	U
EPD-PB-OD-02-081123-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0066	U	0.0066	ppm	0.0066	U
EPD-PB-OD-03-081123-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0066	U	0.0066	ppm	0.0066	U
EPD-PB-OD-04-081123-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0077	U	0.0077	ppm	0.0077	U
EPD-PB-OD-05-081123-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0075	U	0.0075	ppm	0.0075	U
EPD-PB-OD-06-081123-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0081	U	0.0081	ppm	0.0081	U
EPD-PB-OD-07-081123-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0065	U	0.0065	ppm	0.0065	U
EPD-PB-UW-H-081123-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0074	U	0.0074	ppm	0.0074	U
EPD-PB-WA-01-081123-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0068	U	0.0068	ppm	0.0068	U
EPD-PB-WA-02-081123-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0069	U	0.0069	ppm	0.0069	U
EPD-PB-WA-03-081123-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0066	U	0.0066	ppm	0.0066	U
EPD-PB-WA-04-081123-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.007	U	0.007	ppm	0.007	U
EPD-PB-WA-044-081123-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.007	U	0.007	ppm	0.007	U
EPD-PB-WA-05-081123-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.007	U	0.007	ppm	0.007	U
EPD-PB-WA-06-081123-1	NIOSH Method 1450M	141-32-2	n-Butyl acrylate	0.0069	U	0.0069	ppm	0.0069	U